

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121918\
 Data File : BF111627.D
 Acq On : 20 Dec 2018 8:43
 Operator : JU/SJ
 Sample : J6465-04DL 2X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EB03_1.5-3.5DL

Manual Integrations
 APPROVED

Sohil
 12/20/2018 12:42:51 PM

Quant Time: Dec 20 10:25:58 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 19 17:39:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	172728	20.00	ng	0.00
21) Naphthalene-d8	8.18	136	613591	20.00	ng	0.00
39) Acenaphthene-d10	9.93	164	285498	20.00	ng	0.00
64) Phenanthrene-d10	11.42	188	506650	20.00	ng	0.00
76) Chrysene-d12	14.05	240	451177	20.00	ng	0.00
87) Perylene-d12	15.52	264	321075	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	483176	47.68	ng	0.00
7) Phenol-d6	6.52	99	606082	47.00	ng	0.00
23) Nitrobenzene-d5	7.45	82	340576	32.31	ng	0.00
42) 2,4,6-Tribromophenol	10.72	330	147917	43.85	ng	0.00
45) 2-Fluorobiphenyl	9.25	172	688250	35.14	ng	0.00
79) Terphenyl-d14	12.99	244	712062	29.93	ng	0.00
Target Compounds						
10) Phenol	6.53	94	46100	3.168	ng	93
50) Dimethylphthalate	9.64	163	59988	2.873	ng	96
52) Acenaphthene	9.96	154	58540	3.405	ng	94
58) Fluorene	10.47	166	77781	4.155	ng	98
71) Phenanthrene	11.44	178	1147010	42.688	ng	95
72) Anthracene	11.49	178	299790	11.116	ng	95
73) Carbazole	11.64	167	88811	3.392	ng	95
75) Fluoranthene	12.63	202	1756674	64.562	ng	98
78) Pyrene	12.86	202	1523471	47.816	ng	97
81) Benzo(a)anthracene	14.04	228	843682	32.766	ng	100
83) Chrysene	14.07	228	703958	27.817	ng	98
86) Indeno(1,2,3-cd)pyrene	17.00	276	298275	13.865	ng	# 84
88) Benzo(b)fluoranthene	15.09	252	752331m	40.092	ng	
89) Benzo(k)fluoranthene	15.12	252	259391m	14.520	ng	
90) Benzo(a)pyrene	15.46	252	506290m	29.698	ng	
91) Dibenzo(a,h)anthracene	17.01	278	72250	4.840	ng	94
92) Benzo(a,h,i)perylene	17.45	276	266837	18.305	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121918\
Data File : BF111627.D
Acq On : 20 Dec 2018 8:43
Operator : JU/SJ
Sample : J6465-04DL 2X
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :
EB03_1.5-3.5DL

Manual Integrations
APPROVED

Sohil
12/20/2018 12:42:51 PM

Quant Time: Dec 20 10:25:58 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 19 17:39:39 2018
Response via : Initial Calibration

