

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122121\
 Data File : BF126316.D
 Acq On : 22 Dec 2021 1:27
 Operator : JU/CG
 Sample : M5184-08
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-4

Quant Time: Dec 22 05:24:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF121521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 16 12:22:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.816	152	192267	20.000	ng	0.00
21) Naphthalene-d8	8.098	136	839378	20.000	ng	# 0.00
39) Acenaphthene-d10	9.851	164	416074	20.000	ng	0.00
64) Phenanthrene-d10	11.339	188	682179	20.000	ng	# 0.00
76) Chrysene-d12	13.974	240	474427	20.000	ng	# 0.00
86) Perylene-d12	15.421	264	354837	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	818666	62.705	ng	0.00
7) Phenol-d6	6.434	99	674686	40.699	ng	-0.01
23) Nitrobenzene-d5	7.381	82	1463909	94.456	ng	0.00
42) 2,4,6-Tribromophenol	10.645	330	499847	149.487	ng	0.00
45) 2-Fluorobiphenyl	9.180	172	2048003	86.330	ng	0.00
79) Terphenyl-d14	12.927	244	1620210	59.358	ng	0.00


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122121\
Data File : BF126316.D
Acq On : 22 Dec 2021 1:27
Operator : JU/CG
Sample : M5184-08
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-4

Quant Time: Dec 22 05:24:19 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF121521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 16 12:22:48 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.816 min Scan# 804
Delta R.T. -0.000 min
Lab File: BF126316.D
Acq: 22 Dec 2021 1:27




Instrument :
BNA_F
ClientSampleId :
SB-4


Tgt Ion:152 Resp: 192267
Ion Ratio Lower Upper
152 100
150 153.5 117.6 176.4
115 64.2 47.1 70.7



#5
2-Fluorophenol
Concen: 62.705 ng
RT: 5.428 min Scan# 568
Delta R.T. 0.006 min
Lab File: BF126316.D
Acq: 22 Dec 2021 1:27

Tgt Ion:112 Resp: 818666
Ion Ratio Lower Upper
112 100
64 62.4 44.2 66.2
63 29.7 26.2 39.2









#7
Phenol-d6
Concen: 40.699 ng
RT: 6.434 min Scan# 71
Delta R.T. -0.012 min
Lab File: BF126316.D
Acq: 22 Dec 2021 1:27




Instrument :
BNA_F
ClientSampleId :
SB-4


Tgt Ion: 99 Resp: 674686
Ion Ratio Lower Upper
99 100
42 17.3 23.5 35.3#
71 28.8 25.6 38.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.098 min Scan# 1022
Delta R.T. 0.000 min
Lab File: BF126316.D
Acq: 22 Dec 2021 1:27

Tgt Ion: 136 Resp: 839378
Ion Ratio Lower Upper
136 100
137 10.8 9.8 14.6
54 9.3 6.4 9.6
68 7.3 3.3 4.9#





#23

Nitrobenzene-d5

Concen: 94.456 ng

RT: 7.381 min Scan# 90

Delta R.T. 0.000 min

Lab File: BF126316.D

Acq: 22 Dec 2021 1:27

Instrument :





BNA_F

ClientSampleId :

SB-4

Tgt Ion: 82 Resp: 1463909

Ion	Ratio	Lower	Upper
82	100		
128	43.1	40.6	60.8
54	51.6	49.6	74.4



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.851 min Scan# 1320




Delta R.T. -0.006 min


Lab File: BF126316.D

Acq: 22 Dec 2021 1:27

Tgt Ion:164 Resp: 416074

Ion	Ratio	Lower	Upper
164	100		
162	99.5	81.3	121.9
160	45.6	38.0	57.0









#42
2,4,6-Tribromophenol
Concen: 149.487 ng
RT: 10.645 min Scan# 1454
Delta R.T. 0.006 min
Lab File: BF126316.D
Acq: 22 Dec 2021 1:27




Instrument :
BNA_F
ClientSampleId :
SB-4


Tgt Ion:330 Resp: 499847
Ion Ratio Lower Upper
330 100
332 94.4 76.2 114.4
141 32.5 44.6 66.8#



#45
2-Fluorobiphenyl
Concen: 86.330 ng
RT: 9.180 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BF126316.D
Acq: 22 Dec 2021 1:27

Tgt Ion:172 Resp: 2048003
Ion Ratio Lower Upper
172 100
171 35.8 26.5 39.7
170 24.2 18.8 28.2









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.339 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF126316.D
Acq: 22 Dec 2021 1:27




Instrument :
BNA_F
ClientSampleId :
SB-4


Tgt Ion:188 Resp: 682179
Ion Ratio Lower Upper
188 100
94 13.4 6.7 10.1#
80 13.8 7.5 11.3#



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.974 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BF126316.D
Acq: 22 Dec 2021 1:27

Tgt Ion:240 Resp: 474427
Ion Ratio Lower Upper
240 100
120 16.2 10.8 16.2#
236 25.1 21.6 32.4









#79
 Terphenyl-d14
 Concen: 59.358 ng
 RT: 12.927 min Scan# 1842
 Delta R.T. 0.006 min
 Lab File: BF126316.D
 Acq: 22 Dec 2021 1:27

Instrument :
 BNA_F
 ClientSampleId :
 SB-4

Tgt Ion:244 Resp: 1620210
 Ion Ratio Lower Upper
 244 100
 212 6.6 5.9 8.9
 122 15.0 9.0 13.4#



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.421 min Scan# 2267
 Delta R.T. 0.012 min
 Lab File: BF126316.D
 Acq: 22 Dec 2021 1:27

Tgt Ion:264 Resp: 354837
 Ion Ratio Lower Upper
 264 100
 260 23.4 19.6 29.4
 265 21.8 17.2 25.8

