

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF122618\
 Data File : BF111728.D
 Acq On : 26 Dec 2018 21:57
 Operator : JU/SJ
 Sample : J6386-20
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Dec 27 02:16:03 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 19 17:39:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	222394	20.00	ng	0.00
21) Naphthalene-d8	8.18	136	757019	20.00	ng	0.00
39) Acenaphthene-d10	9.93	164	313019	20.00	ng	0.00
64) Phenanthrene-d10	11.42	188	527932	20.00	ng	0.00
76) Chrysene-d12	14.07	240	448251	20.00	ng	0.02
87) Perylene-d12	15.56	264	327573	20.00	ng	0.04

System Monitoring Compounds

5) 2-Fluorophenol	5.54	112	1238974	94.96	ng	0.02
7) Phenol-d6	6.54	99	1557226	93.78	ng	0.02
23) Nitrobenzene-d5	7.46	82	984708	75.71	ng	0.00
42) 2,4,6-Tribromophenol	10.73	330	342896	92.71	ng	0.00
45) 2-Fluorobiphenyl	9.26	172	1602436	74.62	ng	0.00
79) Terphenyl-d14	13.01	244	1359582	57.52	ng	0.01

Target Compounds


						Qvalue
10) Phenol	6.55	94	115456	6.163	ng	99
19) n-Nitroso-di-n-propylamine	7.46	70	135734	12.309	ng	# 79
20) 3+4-Methylphenols	7.30	107	36808	2.517	ng	# 63
25) Isophorone	7.46	82	982631	42.560	ng	# 64
49) Acenaphthylene	9.80	152	112531	3.682	ng	97
50) Dimethylphthalate	9.65	163	228024	9.962	ng	98
65) 4,6-Dinitro-2-methylphenol	10.72	198	780	7.000	ng	# 1
71) Phenanthrene	11.45	178	436179	15.579	ng	94
72) Anthracene	11.50	178	88015	3.132	ng	97
75) Fluoranthene	12.64	202	356325	12.568	ng	95
78) Pyrene	12.87	202	591185	18.676	ng	96
81) Benzo(a)anthracene	14.09	228	202715	7.924	ng	95
83) Chrysene	14.09	228	202715	8.063	ng	98
86) Indeno(1,2,3-cd)pyrene	17.05	276	66685	3.120	ng	# 83
88) Benzo(b)fluoranthene	15.12	252	179482	9.375	ng	# 86
89) Benzo(k)fluoranthene	15.12	252	179482	9.847	ng	# 85
90) Benzo(a)pyrene	15.49	252	130907	7.526	ng	# 92
92) Benzo(a,h,i)perylene	17.50	276	94507	6.355	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF122618\
Data File : BF111728.D
Acq On : 26 Dec 2018 21:57
Operator : JU/SJ
Sample : J6386-20
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Dec 27 02:16:03 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 19 17:39:39 2018
Response via : Initial Calibration






#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.90 min Scan# 819
Delta R.T. 0.01 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
152	100		
150	157.3	126.6	189.8
115	56.8	50.7	76.1






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol
Concen: 94.961 ng
RT: 5.54 min Scan# 587
Delta R.T. 0.02 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57

Tgt Ion	Ratio	Lower	Upper
112	100		
64	61.8	51.8	77.6
63	34.5	26.8	40.2






Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 93.783 ng

RT: 6.54 min Scan# 757

Delta R.T. 0.02 min

Lab File: BF111728.D

Acq: 26 Dec 2018 21:57

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 99 Resp: 1557226

Ion Ratio Lower Upper




99 100

42 22.6 17.4 26.0

71 37.2 26.1 39.1



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#10

Phenol

Concen: 6.163 ng

RT: 6.55 min Scan# 759

Delta R.T. 0.02 min

Lab File: BF111728.D

Acq: 26 Dec 2018 21:57


Tgt Ion: 94 Resp: 115456

Ion Ratio Lower Upper



94 100




65 31.4 11.7 51.7

66 41.9 21.0 61.0



Abundance Ion 94.00 (93.70 to 94.70): BF1
Ion 65.00 (64.70 to 65.70): BF1
Ion 66.00 (65.70 to 66.70): BF1





#19

n-Nitroso-di-n-propylamine

Concen: 12.309 ng

RT: 7.46 min Scan# 913

Delta R.T. 0.15 min

Lab File: BF111728.D





Acq: 26 Dec 2018 21:57

Instrument :

BNA_F

ClientSampleId :

Tot Ion:	70	Resp:	135734
Ion	Ratio	Lower	Upper
70	100		
42	56.5	53.2	79.8
101	0.0	8.2	12.4#
130	2.2	18.1	27.1#



#20

3+4-Methylphenols

Concen: 2.517 ng


RT: 7.30 min Scan# 886


Delta R.T. -0.00 min

Lab File: BF111728.D

Acq: 26 Dec 2018 21:57

Tgt Ion:	107	Resp:	36808
Ion	Ratio	Lower	Upper
107	100		
108	84.7	72.1	112.1
77	40.3	94.1	134.1#
79	25.7	9.8	49.8





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.18 min Scan# 1036

Delta R.T. -0.00 min

Lab File: BF111728.D

Acq: 26 Dec 2018 21:57

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136 Resp: 757019



Ion Ratio Lower Upper

136 100

137 11.2 9.1 13.7

54 9.0 7.7 11.5

68 5.4 4.9 7.3





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#23

Nitrobenzene-d5

Concen: 75.715 ng

RT: 7.46 min Scan# 913

Delta R.T. -0.00 min

Lab File: BF111728.D

Acq: 26 Dec 2018 21:57



Tgt Ion: 82 Resp: 984708

Ion Ratio Lower Upper

82 100

128 42.6 37.4 56.2


54 57.8 47.6 71.4




Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF1





#25

Isophorone

Concen: 42.560 ng

RT: 7.46 min Scan# 913

Delta R.T. -0.26 min

Lab File: BF111728.D

Acq: 26 Dec 2018 21:57

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 82 Resp: 982631

Ion Ratio Lower Upper

82 100

95 0.1 5.4 8.2#

138 0.0 15.1 22.7#



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.93 min Scan# 1334

Delta R.T. -0.00 min

Lab File: BF111728.D

Acq: 26 Dec 2018 21:57




Tgt Ion:164 Resp: 313019


Ion Ratio Lower Upper

164 100

162 102.7 81.4 122.0

160 43.9 35.7 53.5






#42
 2,4,6-Tribromophenol
 Concen: 92.711 ng
 RT: 10.73 min Scan# 1469
 Delta R.T. 0.01 min
 Lab File: BF111728.D
 Acq: 26 Dec 2018 21:57

Instrument :
 BNA_F
 ClientSampled :

Tot Ion	Ratio	Lower	Upper
330	100		
332	97.4	76.0	114.0
141	30.9	31.0	46.6#






Abundance

Ion 329.80 (329.50 to 330.50): E


Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45
 2-Fluorobiphenyl
 Concen: 74.618 ng
 RT: 9.26 min Scan# 1219
 Delta R.T. -0.00 min
 Lab File: BF111728.D
 Acq: 26 Dec 2018 21:57

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.4	33.0	49.4
170	24.3	22.3	33.5






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#49
Acenaphthylene
Concen: 3.682 ng
RT: 9.80 min Scan# 1311
Delta R.T. -0.00 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
152	100		
151	20.9	18.0	27.0
153	14.8	12.3	18.5



#50
Dimethylphthalate
Concen: 9.962 ng
RT: 9.65 min Scan# 1285
Delta R.T. -0.01 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57

Tgt Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.0	4.6
164	10.0	8.9	13.3









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.42 min Scan# 1587
Delta R.T. 0.01 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57

Instrument :
BNA_F
ClientSampleId :

Tot Ion:188 Resp: 527932
Ion Ratio Lower Upper
188 100
94 8.6 9.6 14.4#
80 9.5 9.8 14.6#




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF1
Ion 80.00 (79.70 to 80.70): BF1






#65
4,6-Dinitro-2-methylphenol
Concen: 7.000 ng
RT: 10.72 min Scan# 1468
Delta R.T. 0.21 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57

Tgt Ion:198 Resp: 780
Ion Ratio Lower Upper
198 100
51 322.5 48.0 88.0#
105 395.7 34.0 74.0#



Abundance Ion 198.00 (197.70 to 198.70): E
Ion 51.00 (50.70 to 51.70): BF1
Ion 105.00 (104.70 to 105.70): E









#71
Phenanthrene
Concen: 15.579 ng
RT: 11.45 min Scan# 1591
Delta R.T. 0.01 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57

Instrument :
BNA_F
ClientSampleId :

Tot Ion:178 Resp: 436179
Ion Ratio Lower Upper
178 100
176 18.9 18.1 27.1
179 16.1 14.0 21.0




Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E






#72
Anthracene
Concen: 3.132 ng
RT: 11.50 min Scan# 1600
Delta R.T. 0.01 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57

Tgt Ion:178 Resp: 88015
Ion Ratio Lower Upper
178 100
176 19.7 17.2 25.8
179 16.0 13.8 20.8



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E





#75

Fluoranthene

Concen: 12.568 ng

RT: 12.64 min Scan# 1794

Delta R.T. 0.01 min

Lab File: BF111728.D

Acq: 26 Dec 2018 21:57

Instrument :

BNA_F

ClientSampled :


Tot Ion:202 Resp: 356325

Ion Ratio Lower Upper

202 100

101 9.4 0.0 33.8




203 18.9 0.0 38.6



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.07 min Scan# 2037

Delta R.T. 0.02 min

Lab File: BF111728.D

Acq: 26 Dec 2018 21:57


Tgt Ion:240 Resp: 448251

Ion Ratio Lower Upper

240 100

120 10.5 12.2 18.2#



236 26.1 20.2 30.2




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#78
Pvrene
Concen: 18.676 ng
RT: 12.87 min Scan# 1833
Delta R.T. 0.01 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
202	100		
200	21.1	19.0	28.4
203	17.6	15.2	22.8



#79
Terphenyl-d14
Concen: 57.520 ng
RT: 13.01 min Scan# 1857
Delta R.T. 0.01 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57

Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.7	8.5
122	10.8	13.1	19.7









#81
Benzo(a)anthracene
Concen: 7.924 ng
RT: 14.09 min Scan# 2041
Delta R.T. 0.05 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
228	100		
226	31.1	22.6	34.0
229	22.9	16.3	24.5



#83
Chrysene
Concen: 8.063 ng
RT: 14.09 min Scan# 2041
Delta R.T. 0.02 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.1	25.3	37.9
229	22.9	17.2	25.8









#86
 Indeno(1,2,3-cd)pyrene
 Concen: 3.120 ng
 RT: 17.05 min Scan# 2544
 Delta R.T. 0.05 min
 Lab File: BF111728.D
 Acq: 26 Dec 2018 21:57

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:276 Resp: 66685
 Ion Ratio Lower Upper
 276 100
 138 21.0 27.8 41.6#
 277 28.5 20.1 30.1




Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E






#87
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.56 min Scan# 2290
 Delta R.T. 0.04 min
 Lab File: BF111728.D
 Acq: 26 Dec 2018 21:57

Tgt Ion:264 Resp: 327573
 Ion Ratio Lower Upper
 264 100
 260 23.6 18.3 27.5
 265 23.2 17.7 26.5



Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E









#88
Benzo(b)fluoranthene
Concen: 9.375 ng
RT: 15.12 min Scan# 2216
Delta R.T. 0.04 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
252	100		
253	24.2	18.2	27.2
125	26.2	10.4	15.6



#89
Benzo(k)fluoranthene
Concen: 9.847 ng
RT: 15.12 min Scan# 2216
Delta R.T. -0.00 min
Lab File: BF111728.D
Acq: 26 Dec 2018 21:57

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.2	18.1	27.1
125	26.2	9.8	14.8





#90

Benzo(a)pyrene

Concen: 7.526 ng

RT: 15.49 min Scan# 2279

Delta R.T. 0.04 min

Lab File: BF111728.D


Acq: 26 Dec 2018 21:57

Instrument :

BNA_F

ClientSampled :

Tot Ion:252	Resp:	130907
Ion Ratio	Lower	Upper
252	100	
253	23.8	17.8 26.8
125	22.1	12.3 18.5#






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#92

Benzo(g,h,i)perylene

Concen: 6.355 ng


RT: 17.50 min Scan# 2621

Delta R.T. 0.06 min

Lab File: BF111728.D

Acq: 26 Dec 2018 21:57

Tgt Ion:276	Resp:	94507
Ion Ratio	Lower	Upper
276	100	
277	25.4	19.2 28.8
138	23.3	24.9 37.3#



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

