

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF122618\
 Data File : BF111712.D
 Acq On : 26 Dec 2018 13:29
 Operator : JU/SJ
 Sample : J6498-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSSI-1218-S-DH-9-T

Manual Integrations
 APPROVED

Sohil
 12/27/2018 8:22:15 AM

Quant Time: Dec 26 15:29:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 26 13:36:07 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	176990	20.00	ng	0.00
21) Naphthalene-d8	8.18	136	666056	20.00	ng	0.00
39) Acenaphthene-d10	9.94	164	333013	20.00	ng	0.00
64) Phenanthrene-d10	11.42	188	609972	20.00	ng	0.00
76) Chrysene-d12	14.06	240	468856	20.00	ng	-0.01
87) Perylene-d12	15.54	264	446700	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.53	112	1023507	98.57	ng	0.00
7) Phenol-d6	6.54	99	1327933	100.49	ng	0.00
23) Nitrobenzene-d5	7.46	82	842212	73.60	ng	-0.01
42) 2,4,6-Tribromophenol	10.73	330	396639	100.80	ng	0.00
45) 2-Fluorobiphenyl	9.26	172	1550385	67.86	ng	0.00
79) Terphenyl-d14	13.01	244	1518302	61.41	ng	0.00
Target Compounds						
10) Phenol	6.55	94	98510	6.607	ng	97
50) Dimethylphthalate	9.65	163	225649	9.266	ng	97
75) Fluoranthene	12.63	202	81209	2.479	ng	94
78) Pyrene	12.86	202	80134	2.420	ng	93
88) Benzo(b)fluoranthene	15.10	252	62889m	2.409	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF122618\
Data File : BF111712.D
Acq On : 26 Dec 2018 13:29
Operator : JU/SJ
Sample : J6498-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSSI-1218-S-DH-9-T

Manual Integrations
APPROVED

Sohil
12/27/2018 8:22:15 AM

Quant Time: Dec 26 15:29:16 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 26 13:36:07 2018
Response via : Initial Calibration

