

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122623\
 Data File : BF136739.D
 Acq On : 27 Dec 2023 01:41
 Operator : CG\JU
 Sample : 05446-03RE
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 JC-03-111523B

Quant Time: Dec 27 04:02:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.787	152	63647	20.000	ng	0.00
21) Naphthalene-d8	8.063	136	221898	20.000	ng	0.00
39) Acenaphthene-d10	9.816	164	93512	20.000	ng	-0.01
64) Phenanthrene-d10	11.310	188	153392	20.000	ng	0.00
76) Chrysene-d12	13.968	240	132779	20.000	ng	# 0.00
86) Perylene-d12	15.445	264	93703	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	319522	78.330	ng	0.00
7) Phenol-d6	6.434	99	388084	73.422	ng	0.00
23) Nitrobenzene-d5	7.351	82	232369	53.586	ng	0.00
42) 2,4,6-Tribromophenol	10.610	330	68931	62.590	ng	-0.01
45) 2-Fluorobiphenyl	9.139	172	403660	58.059	ng	-0.01
79) Terphenyl-d14	12.904	244	372456	39.200	ng	-0.01
Target Compounds						Qvalue
9) Benzaldehyde	6.345	77	33910	11.279	ng	# 1
15) Benzyl Alcohol	6.840	79	19732	5.534	ng	# 4
18) Hexachloroethane	7.269	117	7452	4.124	ng	# 1
22) Acetophenone	7.175	105	62431	11.225	ng	# 50
31) Naphthalene	8.081	128	271262	22.252	ng	98
33) 4-Chloroaniline	8.081	127	35955	7.988	ng	# 35
37) 2-Methylnaphthalene	8.775	142	136690	17.423	ng	98
38) 1-Methylnaphthalene	8.875	142	66961	8.700	ng	99
49) Acenaphthylene	9.675	152	25392	2.785	ng	# 90
50) Dimethylphthalate	9.816	163	20393	2.986	ng	# 1
52) Acenaphthene	9.845	154	21863	3.868	ng	92
54) 2,4-Dinitrophenol	10.104	184	1089	4.943	ng	89
58) Fluorene	10.363	166	21732	3.197	ng	# 89
67) 4-Bromophenyl-phenylether	10.610	248	4258	2.500	ng	# 1
70) Pentachlorophenol	11.116	266	13267	14.993	ng	# 88
71) Phenanthrene	11.333	178	122879	15.612	ng	99
72) Anthracene	11.380	178	40910	5.132	ng	96
75) Fluoranthene	12.527	202	221277	26.283	ng	99
78) Pyrene	12.757	202	238396	18.290	ng	98
81) Benzo(a)anthracene	13.957	228	144821	15.707	ng	91
83) Chrysene	13.992	228	147525	16.362	ng	97
87) Indeno(1,2,3-cd)pyrene	16.939	276	45300	6.696	ng	97
88) Benzo(b)fluoranthene	15.010	252	162676	25.993	ng	# 95
89) Benzo(k)fluoranthene	15.010	252	162676	27.508	ng	96
90) Benzo(a)pyrene	15.380	252	79477	15.044	ng	96
91) Dibenzo(a,h)anthracene	16.956	278	13038	2.349	ng	# 73
92) Benzo(g,h,i)perylene	17.398	276	46318	8.148	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122623\
Data File : BF136739.D
Acq On : 27 Dec 2023 01:41
Operator : CG\JU
Sample : 05446-03RE
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
JC-03-111523B

Quant Time: Dec 27 04:02:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 21 01:54:25 2023
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.787 min Scan# 79
Delta R.T. -0.005 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41




Instrument :
BNA_F
ClientSampleId :
JC-03-111523B


Tgt Ion:152 Resp: 63647
Ion Ratio Lower Upper
152 100
150 160.9 127.0 190.6
115 57.0 49.0 73.4



#5
2-Fluorophenol
Concen: 78.330 ng
RT: 5.422 min Scan# 567
Delta R.T. -0.006 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41

Tgt Ion:112 Resp: 319522
Ion Ratio Lower Upper
112 100
64 63.5 51.1 76.7
63 32.2 26.1 39.1









#7
Phenol-d6
Concen: 73.422 ng
RT: 6.434 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41




Instrument :
BNA_F
ClientSampleId :
JC-03-111523B


Tgt Ion: 99 Resp: 388084
Ion Ratio Lower Upper
99 100
42 18.9 16.1 24.1
71 32.3 27.4 41.2



#9
Benzaldehyde
Concen: 11.279 ng
RT: 6.345 min Scan# 724
Delta R.T. -0.006 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41

Tgt Ion: 77 Resp: 33910
Ion Ratio Lower Upper
77 100
105 900.8 84.8 124.8#
106 77.5 77.8 117.8#









#15
Benzyl Alcohol
Concen: 5.534 ng
RT: 6.840 min Scan# 808
Delta R.T. -0.100 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41




Instrument :
BNA_F
ClientSampleId :
JC-03-111523B


Tgt Ion: 79 Resp: 19732
Ion Ratio Lower Upper
79 100
108 0.0 61.0 91.6#
77 143.8 49.8 74.8#



#18
Hexachloroethane
Concen: 4.124 ng
RT: 7.269 min Scan# 881
Delta R.T. -0.029 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41

Tgt Ion: 117 Resp: 7452
Ion Ratio Lower Upper
117 100
119 1320.8 75.8 113.6#
201 0.0 69.1 103.7#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.063 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41

Instrument :
BNA_F
ClientSampleId :
JC-03-111523B

Tgt Ion:136 Resp: 221898




Ion	Ratio	Lower	Upper
136	100		
137	10.3	9.0	13.4
54	7.7	7.2	10.8
68	5.3	4.9	7.3




#22
Acetophenone
Concen: 11.225 ng
RT: 7.175 min Scan# 865
Delta R.T. -0.029 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41

Tgt Ion:105 Resp: 62431

Ion	Ratio	Lower	Upper
105	100		
71	3.8	4.8	7.2#
51	4.3	32.8	49.2#
120	0.4	17.0	25.6#





#23

Nitrobenzene-d5

Concen: 53.586 ng

RT: 7.351 min Scan# 89

Delta R.T. -0.006 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41

Instrument :

BNA_F

ClientSampleId :

JC-03-111523B





Tgt Ion: 82 Resp: 232369

Ion Ratio Lower Upper

82 100

128 48.0 35.3 52.9

54 51.9 44.5 66.7



#31

Naphthalene

Concen: 22.252 ng

RT: 8.081 min Scan# 1019

Delta R.T. -0.012 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41




Tgt Ion: 128 Resp: 271262


Ion Ratio Lower Upper

128 100

129 11.9 8.7 13.1

127 13.1 10.8 16.2









#33
4-Chloroaniline
Concen: 7.988 ng
RT: 8.081 min Scan# 110
Delta R.T. -0.065 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41




Instrument :
BNA_F
ClientSampleId :
JC-03-111523B


Tgt Ion:127	Resp:	35955
Ion Ratio	Lower	Upper
127	100	
129	90.8	26.3 39.5#
65	8.6	25.8 38.8#
92	2.2	15.8 23.6#



#37
2-Methylnaphthalene
Concen: 17.423 ng
RT: 8.775 min Scan# 1137
Delta R.T. -0.012 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41

Tgt Ion:142	Resp:	136690
Ion Ratio	Lower	Upper
142	100	
141	90.9	71.1 106.7





#38

1-Methylnaphthalene

Concen: 8.700 ng

RT: 8.875 min Scan# 1156

Delta R.T. -0.012 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41

Instrument :

BNA_F

ClientSampleId :





JC-03-111523B

Tgt Ion:142 Resp: 66961

Ion Ratio Lower Upper

142 100

141 91.4 73.6 110.4



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.816 min Scan# 1314

Delta R.T. -0.012 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41




Tgt Ion:164 Resp: 93512


Ion Ratio Lower Upper

164 100

162 101.7 81.6 122.4

160 45.3 34.9 52.3





#42

2,4,6-Tribromophenol

Concen: 62.590 ng

RT: 10.610 min Scan# 1451

Delta R.T. -0.012 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41

Instrument :

BNA_F

ClientSampleId :

JC-03-111523B





Tgt Ion:330 Resp: 68931

Ion Ratio Lower Upper

330 100

332 97.7 77.4 116.0

141 30.3 23.4 35.2



#45

2-Fluorobiphenyl

Concen: 58.059 ng

RT: 9.139 min Scan# 1199

Delta R.T. -0.012 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41




Tgt Ion:172 Resp: 403660


Ion Ratio Lower Upper

172 100

171 35.3 29.0 43.6

170 23.8 19.0 28.4





#49

Acenaphthylene

Concen: 2.785 ng

RT: 9.675 min Scan# 1314

Delta R.T. -0.018 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41

Instrument :

BNA_F

ClientSampleId :

JC-03-111523B





Tgt Ion:152 Resp: 25392

Ion Ratio Lower Upper

152 100

151 21.5 16.2 24.4

153 21.7 10.2 15.2#



#50

Dimethylphthalate

Concen: 2.986 ng

RT: 9.816 min Scan# 1314

Delta R.T. 0.259 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41




Tgt Ion:163 Resp: 20393


Ion Ratio Lower Upper

163 100

194 0.7 3.3 4.9#

164 469.8 7.8 11.6#









#52
Acenaphthene
Concen: 3.868 ng
RT: 9.845 min Scan# 11
Delta R.T. -0.018 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41




Instrument :
BNA_F
ClientSampleId :
JC-03-111523B


Tgt Ion:154 Resp: 21863
Ion Ratio Lower Upper
154 100
153 121.8 90.7 136.1
152 60.0 43.7 65.5



#54
2,4-Dinitrophenol
Concen: 4.943 ng
RT: 10.104 min Scan# 1363
Delta R.T. 0.206 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41

Tgt Ion:184 Resp: 1089
Ion Ratio Lower Upper
184 100
63 44.1 40.9 61.3
154 45.3 42.7 64.1





#58

Fluorene

Concen: 3.197 ng

RT: 10.363 min Scan# 1410

Delta R.T. -0.018 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41

Instrument :

BNA_F

ClientSampleId :

JC-03-111523B





Tgt Ion:166 Resp: 21732

Ion Ratio Lower Upper

166 100

165 100.2 73.7 110.5

167 23.1 10.4 15.6#



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.310 min Scan# 1568

Delta R.T. -0.006 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41




Tgt Ion:188 Resp: 153392


Ion Ratio Lower Upper

188 100

94 8.0 7.5 11.3

80 8.8 8.1 12.1









#67
4-Bromophenyl-phenylether
Concen: 2.500 ng
RT: 10.610 min Scan# 1492
Delta R.T. -0.253 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41




Instrument :
BNA_F
ClientSampleId :
JC-03-111523B


Tgt Ion:248 Resp: 4258
Ion Ratio Lower Upper
248 100
250 221.7 77.8 116.6#
141 461.7 59.2 88.8#



#70
Pentachlorophenol
Concen: 14.993 ng
RT: 11.116 min Scan# 1535
Delta R.T. -0.006 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41

Tgt Ion:266 Resp: 13267
Ion Ratio Lower Upper
266 100
268 67.1 52.6 78.8
264 78.2 49.0 73.4#





#71

Phenanthrene

Concen: 15.612 ng

RT: 11.333 min Scan# 1

Delta R.T. -0.012 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41

Instrument :

BNA_F

ClientSampleId :

JC-03-111523B





Tgt Ion:178 Resp: 122879

Ion Ratio Lower Upper

178 100

176 18.4 15.3 22.9

179 15.0 11.9 17.9



#72

Anthracene

Concen: 5.132 ng

RT: 11.380 min Scan# 1580

Delta R.T. -0.018 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41




Tgt Ion:178 Resp: 40910


Ion Ratio Lower Upper

178 100

176 19.7 14.9 22.3

179 16.8 11.8 17.6





#75

Fluoranthene

Concen: 26.283 ng

RT: 12.527 min Scan# 1

Delta R.T. -0.012 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41

Instrument :

BNA_F

ClientSampleId :

JC-03-111523B





Tgt Ion:202 Resp: 221277

Ion Ratio Lower Upper

202 100

101 10.5 0.0 29.3

203 17.3 0.0 37.1



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.968 min Scan# 2020

Delta R.T. 0.000 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41




Tgt Ion:240 Resp: 132779


Ion Ratio Lower Upper

240 100

120 9.9 10.3 15.5#

236 26.2 19.4 29.2









#78
Pyrene
Concen: 18.290 ng
RT: 12.757 min Scan# 1814
Delta R.T. -0.012 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41




Instrument :
BNA_F
ClientSampleId :
JC-03-111523B


Tgt Ion:202 Resp: 238396
Ion Ratio Lower Upper
202 100
200 19.6 15.9 23.9
203 18.3 13.8 20.6



#79
Terphenyl-d14
Concen: 39.200 ng
RT: 12.904 min Scan# 1839
Delta R.T. -0.012 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41

Tgt Ion:244 Resp: 372456
Ion Ratio Lower Upper
244 100
212 6.4 5.0 7.4
122 11.7 9.0 13.4









#81
Benzo(a)anthracene
Concen: 15.707 ng
RT: 13.957 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41




Instrument :
BNA_F
ClientSampleId :
JC-03-111523B


Tgt Ion:228 Resp: 144821
Ion Ratio Lower Upper
228 100
226 32.0 21.4 32.0
229 22.2 14.9 22.3



#83
Chrysene
Concen: 16.362 ng
RT: 13.992 min Scan# 2024
Delta R.T. -0.006 min
Lab File: BF136739.D
Acq: 27 Dec 2023 01:41

Tgt Ion:228 Resp: 147525
Ion Ratio Lower Upper
228 100
226 29.9 24.7 37.1
229 21.9 15.4 23.0





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.445 min Scan# 21

Delta R.T. 0.000 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41

Instrument :

BNA_F

ClientSampleId :

JC-03-111523B





Tgt Ion:264 Resp: 93703

Ion Ratio Lower Upper

264 100

260 21.6 18.3 27.5

265 22.0 17.0 25.6



#87

Indeno(1,2,3-cd)pyrene

Concen: 6.696 ng

RT: 16.939 min Scan# 2525

Delta R.T. -0.012 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41




Tgt Ion:276 Resp: 45300


Ion Ratio Lower Upper

276 100

138 21.5 18.7 28.1

277 26.3 20.2 30.2





#88

Benzo(b)fluoranthene

Concen: 25.993 ng

RT: 15.010 min Scan# 2197

Delta R.T. -0.006 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41

Instrument :

BNA_F

ClientSampleId :

JC-03-111523B





Tgt Ion:252 Resp: 162676

Ion Ratio Lower Upper

252 100

253 23.6 17.4 26.0

125 11.4 7.4 11.2#



#89

Benzo(k)fluoranthene

Concen: 27.508 ng

RT: 15.010 min Scan# 2197

Delta R.T. -0.035 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41




Tgt Ion:252 Resp: 162676


Ion Ratio Lower Upper

252 100

253 23.6 17.0 25.6

125 11.4 8.2 12.4





#90

Benzo(a)pyrene

Concen: 15.044 ng

RT: 15.380 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41

Instrument :

BNA_F

ClientSampleId :

JC-03-111523B





Tgt Ion:252 Resp: 79477

Ion Ratio Lower Upper

252 100

253 23.6 17.0 25.6

125 11.8 8.8 13.2



#91

Dibenzo(a,h)anthracene

Concen: 2.349 ng

RT: 16.956 min Scan# 2528

Delta R.T. -0.012 min

Lab File: BF136739.D

Acq: 27 Dec 2023 01:41




Tgt Ion:278 Resp: 13038


Ion Ratio Lower Upper

278 100

139 22.6 12.2 18.4#

279 38.7 17.8 26.6#





#92
 Benzo(g,h,i)perylene
 Concen: 8.148 ng
 RT: 17.398 min Scan# 20
 Delta R.T. -0.012 min
 Lab File: BF136739.D
 Acq: 27 Dec 2023 01:41

Instrument :
 BNA_F
 ClientSampleId :
 JC-03-111523B

Tgt Ion:276 Resp: 46318
 Ion Ratio Lower Upper
 276 100
 277 27.2 18.3 27.5
 138 20.5 15.6 23.4

