


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF122721\  
 Data File : BF126354.D  
 Acq On : 27 Dec 2021 14:16  
 Operator : JU/CG  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF121521.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Dec 22 13:00:11 2021






AutoFind: Scans 1616, 1617, 1618; Background Corrected with Scan 1611

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 44.0         | 62424      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.3          | 730        | PASS                |
| 69             | 198             | 0.00            | 100             | 38.3         | 54336      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 247        | PASS                |
| 127            | 198             | 10              | 80              | 50.2         | 71261      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 141880     | PASS                |
| 199            | 198             | 5               | 9               | 6.3          | 8970       | PASS                |
| 275            | 198             | 10              | 60              | 24.5         | 34779      | PASS                |
| 365            | 198             | 1               | 100             | 3.3          | 4714       | PASS                |
| 441            | 198             | 0.01            | 100             | 16.5         | 23436      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 149448     | PASS                |
| 443            | 442             | 15              | 24              | 18.5         | 27720      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF122721\  
Data File : BF126354.D  
Acq On : 27 Dec 2021 14:16  
Operator : JU/CG  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Dec 27 15:53:11 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF121521.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Dec 22 13:00:11 2021  
Response via : Initial Calibration



TIC: BF126354.D\data.ms

**(70) Pentachlorophenol (C)**

11.139min (-0.006) 52051.80 ng




response 136213

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 61.10  | 67.09  |
| 264.00 | 69.90  | 62.01  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF122721\  
 Data File : BF126354.D  
 Acq On : 27 Dec 2021 14:16  
 Operator : JU/CG  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Quant Time: Dec 27 15:53:11 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF121521.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Dec 22 13:00:11 2021  
 Response via : Initial Calibration



TIC: BF126354.D\data.ms

**(77) Benzidine**

12.674min ( 0.000) 0.00 ng

response 821669

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 13.50  | 14.22  |
| 183.00 | 10.10  | 11.03  |
| 0.00   | 0.00   | 0.00   |

DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 12/27/2021    | BNA_F            | <u>BF126354.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 311816           | 13.49              |
| DDD           | 754              | 13.19              |
| DDE           | 285              | 12.86              |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 1039          | 312855           | 0.33               |
|               |                  |                    |