

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF123021\
 Data File : BF126415.D
 Acq On : 30 Dec 2021 12:39
 Operator : JU/CG
 Sample : M5241-01 2X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-2S


Quant Time: Dec 30 13:25:56 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 30 04:30:49 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	127799	20.000	ng	0.00
21) Naphthalene-d8	8.081	136	445114	20.000	ng	# 0.00
39) Acenaphthene-d10	9.839	164	145667	20.000	ng	0.00
64) Phenanthrene-d10	11.333	188	212049	20.000	ng	# 0.01
76) Chrysene-d12	13.968	240	192731	20.000	ng	0.01
86) Perylene-d12	15.409	264	219122	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.398	112	457309	51.387	ng	0.00
7) Phenol-d6	6.422	99	552221	47.909	ng	0.00
23) Nitrobenzene-d5	7.357	82	367380	42.379	ng	0.00
42) 2,4,6-Tribromophenol	10.627	330	53972	47.948	ng	0.00
45) 2-Fluorobiphenyl	9.157	172	351677	42.484	ng	0.00
79) Terphenyl-d14	12.921	244	256838	25.892	ng	0.02
Target Compounds						Qvalue
18) Hexachloroethane	7.334	117	21917	6.214	ng	# 1
19) n-Nitroso-di-n-propyla...	7.228	70	22899	3.481	ng	# 65
25) Isophorone	7.610	82	48498	3.125	ng	# 1
27) 2,4-Dimethylphenol	7.739	122	14351	2.387	ng	# 1
32) Benzoic acid	7.875	122	11217	9.268	ng	# 21
35) Caprolactam	8.516	113	70285	50.320	ng	# 16
38) 1-Methylnaphthalene	8.892	142	56804	4.741	ng	96
47) 2-Chloronaphthalene	9.428	162	19068	2.356	ng	92
50) Dimethylphthalate	9.533	163	21544	2.369	ng	# 42
51) 2,6-Dinitrotoluene	9.616	165	14956	7.616	ng	# 69
52) Acenaphthene	9.875	154	35621	4.427	ng	# 86
53) 3-Nitroaniline	9.798	138	11197	4.158	ng	# 22
54) 2,4-Dinitrophenol	9.886	184	723	4.685	ng	# 1
55) Dibenzofuran	10.045	168	26555	2.441	ng	# 26
56) 4-Nitrophenol	9.951	139	6178	3.226	ng	# 1
57) 2,4-Dinitrotoluene	9.998	165	8074	3.187	ng	# 47
58) Fluorene	10.386	166	50406	6.554	ng	# 74
62) 4-Nitroaniline	10.304	138	7611	3.001	ng	86
66) n-Nitrosodiphenylamine	10.504	169	76072	11.563	ng	# 51
71) Phenanthrene	11.351	178	97224	8.968	ng	# 81
72) Anthracene	11.404	178	36863	3.398	ng	# 57
75) Fluoranthene	12.774	202	59172	5.980	ng	75
77) Benzidine	12.645	184	2559	Below Cal		# 1
78) Pyrene	12.774	202	59172	3.633	ng	# 83
81) Benzo(a)anthracene	13.957	228	29168	2.632	ng	# 55
83) Chrysene	13.992	228	33834	2.976	ng	# 55

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 30 13:25:56 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 30 04:30:49 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.798 min Scan# 801
 Delta R.T. 0.000 min
 Lab File: BF126415.D
 Acq: 30 Dec 2021 12:39




Instrument :
 BNA_F
 ClientSampleId :
 TP-2S


Tgt Ion:152 Resp: 127799
 Ion Ratio Lower Upper
 152 100
 150 158.1 128.1 192.1
 115 67.5 56.7 85.1



#5
 2-Fluorophenol
 Concen: 51.387 ng
 RT: 5.398 min Scan# 563
 Delta R.T. -0.006 min
 Lab File: BF126415.D
 Acq: 30 Dec 2021 12:39

Tgt Ion:112 Resp: 457309
 Ion Ratio Lower Upper
 112 100
 64 64.1 49.5 74.3
 63 31.0 23.9 35.9









#7
Phenol-d6
Concen: 47.909 ng
RT: 6.422 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
Client SampleId :
TP-2S


Tgt Ion: 99 Resp: 552221
Ion Ratio Lower Upper
99 100
42 17.8 14.2 21.4
71 29.5 24.3 36.5



#18
Hexachloroethane
Concen: 6.214 ng
RT: 7.334 min Scan# 892
Delta R.T. 0.024 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion: 117 Resp: 21917
Ion Ratio Lower Upper
117 100
119 334.9 75.4 113.0#
201 0.0 56.8 85.2#









#19
n-Nitroso-di-n-propylamine
Concen: 3.481 ng
RT: 7.228 min Scan# 872
Delta R.T. 0.012 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion:	70	Resp:	22899
Ion Ratio	Lower	Upper	
70	100		
42	31.0	45.8	68.8#
101	0.0	6.9	10.3#
130	0.0	14.4	21.6#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.081 min Scan# 1019
Delta R.T. 0.000 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion:	136	Resp:	445114
Ion Ratio	Lower	Upper	
136	100		
137	12.0	8.6	13.0
54	12.6	7.8	11.6#
68	11.2	6.3	9.5#









#23
Nitrobenzene-d5
Concen: 42.379 ng
RT: 7.357 min Scan# 897
Delta R.T. -0.006 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion: 82 Resp: 367380
Ion Ratio Lower Upper
82 100
128 41.0 32.7 49.1
54 49.5 41.7 62.5



#25
Isophorone
Concen: 3.125 ng
RT: 7.610 min Scan# 939
Delta R.T. -0.012 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion: 82 Resp: 48498
Ion Ratio Lower Upper
82 100
95 170.1 5.0 7.4#
138 7.5 12.3 18.5#





#27

2,4-Dimethylphenol

Concen: 2.387 ng

RT: 7.739 min Scan# 90

Delta R.T. 0.006 min

Lab File: BF126415.D

Acq: 30 Dec 2021 12:39

Instrument :

BNA_F

ClientSampleId :

TP-2S





Tgt Ion:122 Resp: 14351

Ion Ratio Lower Upper

122 100

107 336.2 91.2 136.8#

121 100.1 44.4 66.6#



#32

Benzoic acid

Concen: 9.268 ng

RT: 7.875 min Scan# 984

Delta R.T. 0.018 min

Lab File: BF126415.D

Acq: 30 Dec 2021 12:39




Tgt Ion:122 Resp: 11217


Ion Ratio Lower Upper

122 100

105 32.1 101.0 141.0#

77 170.6 74.5 114.5#





#35

Caprolactam

Concen: 50.320 ng

RT: 8.516 min Scan# 1092

Delta R.T. 0.006 min

Lab File: BF126415.D

Acq: 30 Dec 2021 12:39

Instrument :

BNA_F

ClientSampleId :

TP-2S





Tgt Ion:113 Resp: 70285

Ion Ratio Lower Upper

113 100

55 294.1 183.9 223.9#

56 299.9 129.9 169.9#



#38

1-Methylnaphthalene

Concen: 4.741 ng

RT: 8.892 min Scan# 1157

Delta R.T. 0.000 min

Lab File: BF126415.D




Acq: 30 Dec 2021 12:39


Tgt Ion:142 Resp: 56804

Ion Ratio Lower Upper

142 100

141 93.7 72.1 108.1









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.839 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion:164 Resp: 145667
Ion Ratio Lower Upper
164 100
162 101.5 81.4 122.0
160 47.5 36.6 54.8



#42
2,4,6-Tribromophenol
Concen: 47.948 ng
RT: 10.627 min Scan# 1452
Delta R.T. 0.006 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion:330 Resp: 53972
Ion Ratio Lower Upper
330 100
332 96.7 76.0 114.0
141 66.0 30.1 45.1#









#45
2-Fluorobiphenyl
Concen: 42.484 ng
RT: 9.157 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion:172 Resp: 351677
Ion Ratio Lower Upper
172 100
171 35.0 28.3 42.5
170 24.1 19.0 28.6



#47
2-Chloronaphthalene
Concen: 2.356 ng
RT: 9.428 min Scan# 1248
Delta R.T. 0.141 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion:162 Resp: 19068
Ion Ratio Lower Upper
162 100
127 37.1 33.5 50.3
164 27.4 26.2 39.4









#50
Dimethylphthalate
Concen: 2.369 ng
RT: 9.533 min Scan# 1266
Delta R.T. -0.029 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion:163 Resp: 21544
Ion Ratio Lower Upper
163 100
194 24.4 2.6 3.8#
164 30.3 7.8 11.8#



#51
2,6-Dinitrotoluene
Concen: 7.616 ng
RT: 9.616 min Scan# 1280
Delta R.T. -0.006 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion:165 Resp: 14956
Ion Ratio Lower Upper
165 100
63 78.6 48.4 72.6#
89 34.1 50.5 75.7#









#52
Acenaphthene
Concen: 4.427 ng
RT: 9.875 min Scan# 1311
Delta R.T. 0.006 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion:154 Resp: 35621
Ion Ratio Lower Upper
154 100
153 121.1 87.8 131.6
152 62.9 39.4 59.2#



#53
3-Nitroaniline
Concen: 4.158 ng
RT: 9.798 min Scan# 1311
Delta R.T. 0.012 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion:138 Resp: 11197
Ion Ratio Lower Upper
138 100
108 76.0 8.6 13.0#
92 49.5 105.0 157.6#









#54
2,4-Dinitrophenol
Concen: 4.685 ng
RT: 9.886 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion:184 Resp: 723
Ion Ratio Lower Upper
184 100
63 1002.0 70.4 105.6#
154 1108.7 49.7 74.5#



#55
Dibenzofuran
Concen: 2.441 ng
RT: 10.045 min Scan# 1353
Delta R.T. 0.000 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion:168 Resp: 26555
Ion Ratio Lower Upper
168 100
139 70.4 30.6 46.0#
169 66.9 10.0 15.0#









#56
4-Nitrophenol
Concen: 3.226 ng
RT: 9.951 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion:139 Resp: 6178
Ion Ratio Lower Upper
139 100
109 527.5 44.1 84.1#
65 120.8 85.5 125.5



#57
2,4-Dinitrotoluene
Concen: 3.187 ng
RT: 9.998 min Scan# 1345
Delta R.T. -0.023 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion:165 Resp: 8074
Ion Ratio Lower Upper
165 100
63 36.6 43.0 64.6#
89 22.1 71.0 106.4#
182 20.9 1.6 2.4#









#58
Fluorene
Concen: 6.554 ng
RT: 10.386 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion:166 Resp: 50406
Ion Ratio Lower Upper
166 100
165 119.5 77.7 116.5#
167 35.0 10.7 16.1#



#62
4-Nitroaniline
Concen: 3.001 ng
RT: 10.304 min Scan# 1397
Delta R.T. -0.094 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion:138 Resp: 7611
Ion Ratio Lower Upper
138 100
92 52.4 36.2 76.2
108 84.1 46.7 86.7





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.333 min Scan# 11

Delta R.T. 0.012 min

Lab File: BF126415.D

Acq: 30 Dec 2021 12:39

Instrument :

BNA_F

ClientSampleId :

TP-2S





Tgt Ion:188 Resp: 212049

Ion Ratio Lower Upper

188 100

94 16.7 10.1 15.1#

80 16.4 11.8 17.6



#66

n-Nitrosodiphenylamine

Concen: 11.563 ng

RT: 10.504 min Scan# 1431

Delta R.T. 0.006 min

Lab File: BF126415.D

Acq: 30 Dec 2021 12:39




Tgt Ion:169 Resp: 76072


Ion Ratio Lower Upper

169 100

168 23.1 53.4 80.0#

167 13.1 29.4 44.0#





#71

Phenanthrene

Concen: 8.968 ng

RT: 11.351 min Scan# 1575

Delta R.T. 0.006 min

Lab File: BF126415.D

Acq: 30 Dec 2021 12:39

Instrument :

BNA_F

ClientSampleId :

TP-2S





Tgt Ion:178 Resp: 97224

Ion Ratio Lower Upper

178 100

176 20.4 15.7 23.5

179 32.6 12.3 18.5#



#72

Anthracene

Concen: 3.398 ng

RT: 11.404 min Scan# 1584

Delta R.T. 0.006 min

Lab File: BF126415.D

Acq: 30 Dec 2021 12:39




Tgt Ion:178 Resp: 36863


Ion Ratio Lower Upper

178 100

176 25.7 15.0 22.6#

179 47.6 12.5 18.7#









#75
Fluoranthene
Concen: 5.980 ng
RT: 12.774 min Scan# 11
Delta R.T. 0.241 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion:202 Resp: 59172
Ion Ratio Lower Upper
202 100
101 20.4 0.0 35.7
203 33.4 0.0 37.1



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.968 min Scan# 2020
Delta R.T. 0.012 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion:240 Resp: 192731
Ion Ratio Lower Upper
240 100
120 18.2 14.4 21.6
236 25.7 20.5 30.7









#77
Benzidine
Concen: Below Cal
RT: 12.645 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39




Instrument :
BNA_F
ClientSampleId :
TP-2S


Tgt Ion:184 Resp: 2559
Ion Ratio Lower Upper
184 100
185 127.2 11.8 17.8#
183 188.7 8.2 12.2#



#78
Pyrene
Concen: 3.633 ng
RT: 12.774 min Scan# 1817
Delta R.T. 0.012 min
Lab File: BF126415.D
Acq: 30 Dec 2021 12:39

Tgt Ion:202 Resp: 59172
Ion Ratio Lower Upper
202 100
200 21.2 16.6 24.8
203 33.4 13.8 20.8#









#79
 Terphenyl-d14
 Concen: 25.892 ng
 RT: 12.921 min Scan# 11
 Delta R.T. 0.018 min
 Lab File: BF126415.D
 Acq: 30 Dec 2021 12:39




Instrument :
 BNA_F
 ClientSampleId :
 TP-2S


Tgt Ion:244 Resp: 256838
 Ion Ratio Lower Upper
 244 100
 212 6.8 5.0 7.4
 122 13.8 13.3 19.9



#81
 Benzo(a)anthracene
 Concen: 2.632 ng
 RT: 13.957 min Scan# 2018
 Delta R.T. 0.012 min
 Lab File: BF126415.D
 Acq: 30 Dec 2021 12:39

Tgt Ion:228 Resp: 29168
 Ion Ratio Lower Upper
 228 100
 226 38.4 21.4 32.0#
 229 54.2 15.4 23.2#





#83

Chrysene

Concen: 2.976 ng

RT: 13.992 min Scan# 20

Delta R.T. 0.006 min

Lab File: BF126415.D

Acq: 30 Dec 2021 12:39

Instrument :

BNA_F

ClientSampleId :

TP-2S





Tgt Ion:228 Resp: 33834

Ion Ratio Lower Upper

228 100

226 40.7 23.7 35.5#

229 56.9 15.5 23.3#



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.409 min Scan# 2265

Delta R.T. 0.012 min

Lab File: BF126415.D

Acq: 30 Dec 2021 12:39

Tgt Ion:264 Resp: 219122

Ion Ratio Lower Upper

264 100

260 24.1 19.2 28.8

265 22.6 17.4 26.2

