

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF123021\
 Data File : BF126426.D
 Acq On : 30 Dec 2021 18:52
 Operator : JU/CG
 Sample : M5241-05
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-2D

Quant Time: Dec 31 00:37:44 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 30 04:30:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	129244	20.000	ng	0.00
21) Naphthalene-d8	8.080	136	527430	20.000	ng	0.00
39) Acenaphthene-d10	9.833	164	245111	20.000	ng	0.00
64) Phenanthrene-d10	11.316	188	361713	20.000	ng	0.00
76) Chrysene-d12	13.951	240	223378	20.000	ng	0.00
86) Perylene-d12	15.392	264	227640	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	986502	109.613	ng	0.02
7) Phenol-d6	6.428	99	1143894	98.131	ng	0.00
23) Nitrobenzene-d5	7.357	82	880017	85.671	ng	0.00
42) 2,4,6-Tribromophenol	10.621	330	247662	130.756	ng	0.00
45) 2-Fluorobiphenyl	9.157	172	1161553	83.391	ng	0.00
79) Terphenyl-d14	12.904	244	923274	80.306	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF123021\
Data File : BF126426.D
Acq On : 30 Dec 2021 18:52
Operator : JU/CG
Sample : M5241-05
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-2D

Quant Time: Dec 31 00:37:44 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 30 04:30:49 2021
Response via : Initial Calibration

