

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010623\
 Data File : BF131944.D
 Acq On : 06 Jan 2023 16:30
 Operator : CG\JU
 Sample : 01029-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-205

Quant Time: Jan 07 00:22:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 06 23:58:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	55357	20.000	ng	0.00
21) Naphthalene-d8	8.086	136	196829	20.000	ng	0.00
39) Acenaphthene-d10	9.851	164	98876	20.000	ng	0.00
64) Phenanthrene-d10	11.345	188	162592	20.000	ng	0.00
76) Chrysene-d12	13.998	240	124627	20.000	ng	0.00
86) Perylene-d12	15.474	264	99022	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	486285	145.579	ng	0.02
7) Phenol-d6	6.440	99	567405	129.290	ng	0.00
23) Nitrobenzene-d5	7.369	82	469654	95.267	ng	0.00
42) 2,4,6-Tribromophenol	10.645	330	129483	120.529	ng	0.00
45) 2-Fluorobiphenyl	9.169	172	779111	107.930	ng	0.00
79) Terphenyl-d14	12.939	244	741472	100.637	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010623\
Data File : BF131944.D
Acq On : 06 Jan 2023 16:30
Operator : CG\JU
Sample : 01029-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-205

Quant Time: Jan 07 00:22:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 06 23:58:27 2023
Response via : Initial Calibration

