

Data File : BF118564.D
Acq On : 17 Jan 2020 18:07
Operator : JU
Sample : L1008-14
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CL-01-011620

Quant Time: Jan 18 01:02:40 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF011720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 17 15:49:53 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	279027	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	1053770	20.00	ng	0.00
39) Acenaphthene-d10	9.92	164	588850	20.00	ng	-0.01
64) Phenanthrene-d10	11.40	188	1130200	20.00	ng	-0.01
76) Chrysene-d12	14.04	240	1049054	20.00	ng	0.00
87) Perylene-d12	15.52	264	1147679	20.00	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	751270	49.75	ng	-0.01
7) Phenol-d6	6.50	99	919079	46.45	ng	-0.02
23) Nitrobenzene-d5	7.44	82	589323	34.54	ng	-0.02
42) 2,4,6-Tribromophenol	10.70	330	334915	51.45	ng	-0.01
45) 2-Fluorobiphenyl	9.25	172	1258739	38.59	ng	0.00
79) Terphenyl-d14	12.99	244	1844637	34.73	ng	0.00

Target Compounds				Qvalue		
3) Pyridine	3.46	79	270	0.013	ng	# 14
4) n-Nitrosodimethylamine	3.39	42	707	0.077	ng	# 1
6) Aniline	6.54	93	130	0.005	ng	# 41
8) 2-Chlorophenol	6.66	128	283	0.016	ng	# 1
9) Benzaldehyde	6.50	77	1944	0.192	ng	# 12
10) Phenol	6.51	94	3008	0.139	ng	# 47
11) bis(2-Chloroethyl)ether	6.65	93	1370	0.079	ng	# 1
15) Benzyl Alcohol	7.00	79	389	0.024	ng	# 14
16) 2,2'-oxybis(1-Chloropropan	7.17	45	53	0.002	ng	# 65
17) 2-Methylphenol	7.32	107	243	0.017	ng	# 13
19) n-Nitroso-di-n-propylamine	7.44	70	83465	5.882	ng	# 78
20) 3+4-Methylphenols	7.32	107	243	0.014	ng	# 5
22) Acetophenone	7.29	105	508	0.020	ng	# 48
24) Nitrobenzene	7.47	77	252	0.014	ng	# 37
25) Isophorone	7.44	82	586034	15.826	ng	# 63
27) 2,4-Dimethylphenol	7.83	122	849	0.058	ng	# 4
28) bis(2-Chloroethoxy)methane	8.17	93	292	0.012	ng	# 1
31) Naphthalene	8.19	128	2333	0.047	ng	# 82
32) Benzoic acid	7.83	122	849	7.924	ng	# 79
33) 4-Chloroaniline	8.19	127	243	0.011	ng	# 45
37) 2-Methylnaphthalene	8.88	142	1061	0.030	ng	# 84
38) 1-Methylnaphthalene	8.98	142	688	0.021	ng	# 98
46) 1,1'-Biphenyl	9.35	154	2981	0.071	ng	# 89
48) 2-Nitroaniline	9.25	65	2291	0.248	ng	# 1
49) Acenaphthylene	9.79	152	414	0.008	ng	# 57
50) Dimethylphthalate	9.63	163	6160	0.152	ng	# 98
51) 2,6-Dinitrotoluene	9.92	165	77945	10.104	ng	# 27
52) Acenaphthene	9.95	154	412	0.013	ng	# 88
55) Dibenzofuran	10.13	168	789	0.017	ng	# 55
56) 4-Nitrophenol	10.13	139	372	0.058	ng	# 4
57) 2,4-Dinitrotoluene	9.92	165	77945	11.900	ng	# 23
58) Fluorene	10.46	166	946	0.026	ng	# 68
60) Diethylphthalate	10.33	149	3308	0.084	ng	# 91
63) Azobenzene	10.62	77	1492	0.039	ng	# 53

Data File : BF118564.D
Acq On : 17 Jan 2020 18:07
Operator : JU
Sample : L1008-14
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CL-01-011620

Quant Time: Jan 18 01:02:40 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF011720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 17 15:49:53 2020
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
65) 4,6-Dinitro-2-methylphenol	10.70	198	770	9.859	ng	# 1
66) n-Nitrosodiphenylamine	10.70	169	9385	0.258	ng	# 50
67) 4-Bromophenyl-phenylether	10.70	248	20493	1.355	ng	# 1
71) Phenanthrene	11.43	178	2717	0.048	ng	95
72) Anthracene	11.43	178	2717	0.048	ng	96
73) Carbazole	11.63	167	1400	0.027	ng	# 87
74) Di-n-butylphthalate	11.97	149	5091	0.080	ng	99
75) Fluoranthene	12.62	202	1839	0.030	ng	81
77) Benzidine	12.99	184	24159	0.948	ng	# 93
78) Pyrene	12.84	202	1828	0.023	ng	# 74
80) Butylbenzylphthalate	13.47	149	766	0.023	ng	# 52
81) Benzo(a)anthracene	14.04	228	5860	0.088	ng	# 70
83) Chrysene	14.04	228	5860	0.092	ng	# 67
84) Bis(2-ethylhexyl)phthalate	14.03	149	18313	0.440	ng	96
85) Di-n-octyl phthalate	14.64	149	1514	0.024	ng	# 74
86) Indeno(1,2,3-cd)pyrene	16.99	276	815	0.013	ng	# 70
88) Benzo(b)fluoranthene	15.07	252	1525	0.020	ng	# 90
89) Benzo(k)fluoranthene	15.11	252	1562	0.023	ng	# 51
90) Benzo(a)pyrene	15.44	252	884	0.014	ng	# 28
91) Dibenzo(a,h)anthracene	17.00	278	441	0.007	ng	# 55
92) Benzo(g,h,i)perylene	17.42	276	822	0.014	ng	# 55


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BF118564.D
Acq On : 17 Jan 2020 18:07
Operator : JU
Sample : L1008-14
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CL-01-011620

Quant Time: Jan 18 01:02:40 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF011720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 17 15:49:53 2020
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.89 min Scan# 816
 Delta R.T. -0.01 min
 Lab File: BF118564.D
 Acq: 17 Jan 2020 18:07

Instrument :
 BNA_F
 ClientSampleId :
 CL-01-011620

Tgt Ion:152 Resp: 279027
 Ion Ratio Lower Upper
 152 100
 150 152.4 124.9 187.3
 115 58.2 45.8 68.6




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#3


Pyridine
 Concen: 0.013 ng
 RT: 3.46 min Scan# 233
 Delta R.T. 0.01 min
 Lab File: BF118564.D
 Acq: 17 Jan 2020 18:07

Tgt Ion: 79 Resp: 270
 Ion Ratio Lower Upper
 79 100
 52 0.0 51.3 76.9#
 51 87.2 26.7 40.1#



Abundance Ion 79.00 (78.70 to 79.70): BF1
 Ion 52.00 (51.70 to 52.70): BF1
 Ion 51.00 (50.70 to 51.70): BF1





#4

n-Nitrosodimethylamine

Concen: 0.077 ng

RT: 3.39 min Scan# 222

Delta R.T. 0.01 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07

Instrument :

BNA_F

ClientSampleId :

CL-01-011620


Tgt Ion: 42 Resp: 707

Ion Ratio Lower Upper




42 100

74 0.0 103.1 154.7#

44 1275.6 5.4 8.2#



Abundance Ion 42.00 (41.70 to 42.70): BF1
Ion 74.00 (73.70 to 74.70): BF1
Ion 44.00 (43.70 to 44.70): BF1



#5

2-Fluorophenol

Concen: 49.749 ng

RT: 5.49 min Scan# 579

Delta R.T. -0.01 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07


Tgt Ion: 112 Resp: 751270

Ion Ratio Lower Upper



112 100


64 60.7 49.1 73.7

63 32.0 25.8 38.8



Abundance Ion 112.00 (111.70 to 112.70): BF1
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#6

Aniline

Concen: 0.005 ng

RT: 6.54 min Scan# 757

Delta R.T. -0.01 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07

Instrument :

BNA_F

ClientSampleId :

CL-01-011620


Tgt Ion: 93 Resp: 130

Ion Ratio Lower Upper

93 100

66 0.0 32.9 49.3#




65 0.0 16.5 24.7#



Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1



#7

Phenol-d6

Concen: 46.448 ng

RT: 6.50 min Scan# 750

Delta R.T. -0.02 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07


Tgt Ion: 99 Resp: 919079

Ion Ratio Lower Upper

99 100

42 19.1 14.9 22.3



71 35.3 28.3 42.5




Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1





#8

2-Chlorophenol

Concen: 0.016 ng

RT: 6.66 min Scan# 778

Delta R.T. -0.01 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07

Instrument :

BNA_F

ClientSampleId :

CL-01-011620


Tgt Ion:128 Resp: 283

Ion Ratio Lower Upper

128 100

130 190.7 13.2 53.2#




64 507.2 28.9 68.9#



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1



#9

Benzaldehyde

Concen: 0.192 ng

RT: 6.50 min Scan# 750

Delta R.T. 0.06 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07


Tgt Ion: 77 Resp: 1944

Ion Ratio Lower Upper

77 100

105 17.8 75.2 115.2#



106 0.0 71.4 111.4#




Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): E

Ion 106.00 (105.70 to 106.70): E





#10

Phenol

Concen: 0.139 ng

RT: 6.51 min Scan# 752

Delta R.T. -0.02 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07

Instrument :

BNA_F

ClientSampleId :

CL-01-011620





Tgt Ion: 94 Resp: 3008

Ion Ratio Lower Upper

94 100

65 41.3 9.8 49.8

66 86.9 19.5 59.5#



#11

bis(2-Chloroethyl)ether

Concen: 0.079 ng

RT: 6.65 min Scan# 776

Delta R.T. 0.02 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07




Tgt Ion: 93 Resp: 1370


Ion Ratio Lower Upper

93 100

63 728.8 54.0 94.0#

95 469.6 13.5 53.5#





#15

Benzyl Alcohol

Concen: 0.024 ng

RT: 7.00 min Scan# 836

Delta R.T. -0.02 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07

Instrument :

BNA_F

ClientSampleId :

CL-01-011620





Tgt Ion: 79 Resp: 389

Ion Ratio Lower Upper

79 100

108 0.0 61.9 92.9#

77 0.0 51.5 77.3#



#16

2,2'-oxybis(1-chloropropane)

Concen: 0.002 ng

RT: 7.17 min Scan# 865

Delta R.T. 0.01 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07




Tgt Ion: 45 Resp: 53


Ion Ratio Lower Upper

45 100

77 0.0 0.0 35.8

79 0.0 0.0 32.2





#17

2-Methylphenol

Concen: 0.017 ng

RT: 7.32 min Scan# 889

Delta R.T. 0.18 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07





Instrument :

BNA_F

ClientSampleId :

CL-01-011620

Tgt Ion:	107	Resp:	243
Ion	Ratio	Lower	Upper
107	100		
108	0.0	92.3	138.5#
77	0.0	38.6	57.8#
79	83.9	37.0	55.6#



#19

n-Nitroso-di-n-propylamine

Concen: 5.882 ng




RT: 7.44 min Scan# 910


Delta R.T. 0.13 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07

Tgt Ion:	70	Resp:	83465
Ion	Ratio	Lower	Upper
70	100		
42	45.7	43.6	65.4
101	0.0	7.7	11.5#
130	2.1	18.3	27.5#









#20
3+4-Methylphenols
Concen: 0.014 ng
RT: 7.32 min Scan# 889
Delta R.T. 0.02 min
Lab File: BF118564.D
Acq: 17 Jan 2020 18:07




Instrument :
BNA_F
ClientSampleId :
CL-01-011620


Tgt Ion:107 Resp: 243
Ion Ratio Lower Upper
107 100
108 0.0 72.6 112.6#
77 0.0 63.0 103.0#
79 83.9 11.3 51.3#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.17 min Scan# 1034
Delta R.T. -0.01 min
Lab File: BF118564.D
Acq: 17 Jan 2020 18:07

Tgt Ion:136 Resp: 1053770
Ion Ratio Lower Upper
136 100
137 11.3 9.4 14.0
54 7.6 6.2 9.2
68 6.2 5.4 8.0





#22

Acetophenone

Concen: 0.020 ng

RT: 7.29 min Scan# 884

Delta R.T. -0.01 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07

Instrument :

BNA_F

ClientSampleId :

CL-01-011620

Tgt Ion: 105 Resp: 508



Ion Ratio Lower Upper

105 100

71 70.1 3.8 5.6#

51 49.0 24.9 37.3#

120 47.6 19.2 28.8#





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E



#23

Nitrobenzene-d5

Concen: 34.540 ng

RT: 7.44 min Scan# 910

Delta R.T. -0.02 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07



Tgt Ion: 82 Resp: 589323

Ion Ratio Lower Upper

82 100

128 43.0 37.0 55.4


54 46.9 37.4 56.0




Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF1







#24
 Nitrobenzene
 Concen: 0.014 ng
 RT: 7.47 min Scan# 915
 Delta R.T. -0.01 min
 Lab File: BF118564.D
 Acq: 17 Jan 2020 18:07

Instrument :
 BNA_F
 ClientSampleId :
 CL-01-011620



Tgt Ion: 77 Resp: 252
 Ion Ratio Lower Upper
 77 100
 123 0.0 38.2 57.2#
 65 0.0 11.1 16.7#



Abundance Ion 77.00 (76.70 to 77.70): BF1
 Ion 123.00 (122.70 to 123.70): BF1
 Ion 65.00 (64.70 to 65.70): BF1




Time-->




#25
 Isophorone
 Concen: 15.826 ng
 RT: 7.44 min Scan# 910
 Delta R.T. -0.28 min
 Lab File: BF118564.D
 Acq: 17 Jan 2020 18:07


Tgt Ion: 82 Resp: 586034
 Ion Ratio Lower Upper
 82 100
 95 0.1 6.0 9.0#
 138 0.0 15.8 23.8#




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 95.00 (94.70 to 95.70): BF1
 Ion 138.00 (137.70 to 138.70): BF1



Time-->





#27

2,4-Dimethylphenol

Concen: 0.058 ng

RT: 7.83 min Scan# 976

Delta R.T. 0.01 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07

Instrument :

BNA_F

ClientSampleId :

CL-01-011620


Tgt Ion:122 Resp: 849

Ion Ratio Lower Upper

122 100

107 0.0 89.6 134.4#




121 0.0 45.9 68.9#



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 0.012 ng

RT: 8.17 min Scan# 1034

Delta R.T. 0.25 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07


Tgt Ion: 93 Resp: 292

Ion Ratio Lower Upper

93 100

95 331.0 27.3 40.9#



123 0.0 10.2 15.4#




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 123.00 (122.70 to 123.70): E









#31
Naphthalene
Concen: 0.047 ng
RT: 8.19 min Scan# 1038
Delta R.T. -0.01 min
Lab File: BF118564.D
Acq: 17 Jan 2020 18:07




Instrument :
BNA_F
ClientSampleId :
CL-01-011620


Tgt Ion:128 Resp: 2333
Ion Ratio Lower Upper
128 100
129 0.0 10.3 15.5#
127 13.0 12.1 18.1



#32
Benzoic acid
Concen: 7.924 ng
RT: 7.83 min Scan# 976
Delta R.T. -0.11 min
Lab File: BF118564.D
Acq: 17 Jan 2020 18:07

Tgt Ion:122 Resp: 849
Ion Ratio Lower Upper
122 100
105 157.5 105.0 145.0#
77 112.9 80.7 120.7









#33
4-Chloroaniline
Concen: 0.011 ng
RT: 8.19 min Scan# 1038
Delta R.T. -0.05 min
Lab File: BF118564.D
Acq: 17 Jan 2020 18:07




Instrument :
BNA_F
ClientSampleId :
CL-01-011620


Tgt Ion:	127	Resp:	243
Ion	Ratio	Lower	Upper
127	100		
129	0.0	27.0	40.4#
65	0.0	25.8	38.8#
92	0.0	15.8	23.8#



#37
2-Methylnaphthalene
Concen: 0.030 ng
RT: 8.88 min Scan# 1155
Delta R.T. -0.01 min
Lab File: BF118564.D
Acq: 17 Jan 2020 18:07

Tgt Ion:	142	Resp:	1061
Ion	Ratio	Lower	Upper
142	100		
141	105.7	70.6	106.0
115	42.7	29.0	43.4









#38
 1-Methylnaphthalene
 Concen: 0.021 ng
 RT: 8.98 min Scan# 1172
 Delta R.T. -0.01 min
 Lab File: BF118564.D
 Acq: 17 Jan 2020 18:07




Instrument :
 BNA_F
 ClientSampleId :
 CL-01-011620


Tgt Ion:	142	Resp:	688
Ion	Ratio	Lower	Upper
142	100		
141	93.8	73.8	110.6
116	0.0	3.3	4.9#



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.92 min Scan# 1332
 Delta R.T. -0.01 min
 Lab File: BF118564.D
 Acq: 17 Jan 2020 18:07

Tgt Ion:	164	Resp:	588850
Ion	Ratio	Lower	Upper
164	100		
162	102.2	80.6	120.8
160	46.4	36.7	55.1





#42

2,4,6-Tribromophenol

Concen: 51.449 ng

RT: 10.70 min Scan# 1465

Delta R.T. -0.01 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07

Instrument :

BNA_F

ClientSampleId :

CL-01-011620


Tgt Ion:330 Resp: 334915

Ion Ratio Lower Upper

330 100

332 95.4 76.4 114.6




141 32.2 23.4 35.2



Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45

2-Fluorobiphenyl

Concen: 38.591 ng

RT: 9.25 min Scan# 1217

Delta R.T. -0.01 min

Lab File: BF118564.D

Acq: 17 Jan 2020 18:07


Tgt Ion:172 Resp: 1258739

Ion Ratio Lower Upper

172 100

171 36.1 32.4 48.6

170 23.9 22.5 33.7



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E

