

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011720\
 Data File : BF118567.D
 Acq On : 17 Jan 2020 19:31
 Operator : JU
 Sample : L1005-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HR-05-011420

Quant Time: Jan 18 01:32:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 17 15:49:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	305047	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	1096439	20.00	ng	0.00
39) Acenaphthene-d10	9.92	164	613771	20.00	ng	-0.01
64) Phenanthrene-d10	11.41	188	1150592	20.00	ng	0.00
76) Chrysene-d12	14.04	240	982475	20.00	ng	0.00
87) Perylene-d12	15.52	264	1013849	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	2406611	145.77	ng	0.01
7) Phenol-d6	6.52	99	2949706	136.36	ng	0.00
23) Nitrobenzene-d5	7.45	82	1956493	110.21	ng	-0.01
42) 2,4,6-Tribromophenol	10.71	330	1117414	164.69	ng	0.00
45) 2-Fluorobiphenyl	9.25	172	3662763	107.73	ng	0.00
79) Terphenyl-d14	13.00	244	4637709	93.23	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011720\
Data File : BF118567.D
Acq On : 17 Jan 2020 19:31
Operator : JU
Sample : L1005-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HR-05-011420

Quant Time: Jan 18 01:32:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 17 15:49:53 2020
Response via : Initial Calibration

