

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011823\
 Data File : BF132121.D
 Acq On : 18 Jan 2023 16:37
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 19 05:19:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 19 00:42:19 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	40.285	-0.7	102	0.01
3	Pyridine	40.000	40.247	-0.6	102	0.01
4	n-Nitrosodimethylamine	40.000	40.672	-1.7	100	0.01
5 S	2-Fluorophenol	80.000	76.415	4.5	100	0.00
6	Aniline	40.000	39.339	1.7	103	0.00
7 S	Phenol-d6	80.000	76.651	4.2	101	0.00
8	2-Chlorophenol	40.000	39.075	2.3	102	0.00
9	Benzaldehyde	40.000	36.766	8.1	101	0.00
10 C	Phenol	40.000	38.448	3.9	102	0.00
11	bis(2-Chloroethyl)ether	40.000	38.857	2.9	101	0.00
12	1,3-Dichlorobenzene	40.000	38.899	2.8	102	0.00
13 C	1,4-Dichlorobenzene	40.000	38.551	3.6	101	0.00
14	1,2-Dichlorobenzene	40.000	36.476	8.8	101	0.00
15	Benzyl Alcohol	40.000	40.937	-2.3	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.275	1.8	101	0.00
17	2-Methylphenol	40.000	39.670	0.8	103	0.00
18	Hexachloroethane	40.000	39.985	0.0	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.895	5.3	101	0.00
20	3+4-Methylphenols	40.000	39.357	1.6	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	37.277	6.8	103	0.00
23 S	Nitrobenzene-d5	80.000	79.367	0.8	102	0.00
24	Nitrobenzene	40.000	39.992	0.0	103	0.00
25	Isophorone	40.000	39.259	1.9	103	0.00
26 C	2-Nitrophenol	40.000	42.691	-6.7	104	0.00
27	2,4-Dimethylphenol	40.000	38.880	2.8	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.802	5.5	102	0.00
29 C	2,4-Dichlorophenol	40.000	39.205	2.0	102	0.00
30	1,2,4-Trichlorobenzene	40.000	38.679	3.3	102	0.00
31	Naphthalene	40.000	37.520	6.2	101	0.00
32	Benzoic acid	40.000	41.760	-4.4	101	0.00
33	4-Chloroaniline	40.000	39.091	2.3	101	0.00
34 C	Hexachlorobutadiene	40.000	39.081	2.3	102	0.00
35	Caprolactam	40.000	40.374	-0.9	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.648	0.9	102	0.00
37	2-Methylnaphthalene	40.000	37.881	5.3	102	0.00
38	1-Methylnaphthalene	40.000	37.260	6.9	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.037	4.9	101	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.958	-7.4	104	0.00
42 S	2,4,6-Tribromophenol	80.000	78.821	1.5	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.958	-2.4	101	0.00
44	2,4,5-Trichlorophenol	40.000	39.953	0.1	103	0.00
45 S	2-Fluorobiphenyl	80.000	76.796	4.0	102	0.00
46	1,1'-Biphenyl	40.000	38.144	4.6	102	0.00
47	2-Chloronaphthalene	40.000	38.089	4.8	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011823\
 Data File : BF132121.D
 Acq On : 18 Jan 2023 16:37
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 19 05:19:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 19 00:42:19 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.153	-5.4	101	0.00
49	Acenaphthylene	40.000	38.187	4.5	101	0.00
50	Dimethylphthalate	40.000	38.440	3.9	102	0.00
51	2,6-Dinitrotoluene	40.000	40.945	-2.4	101	0.00
52 C	Acenaphthene	40.000	37.872	5.3	102	0.00
53	3-Nitroaniline	40.000	40.456	-1.1	100	0.00
54 P	2,4-Dinitrophenol	40.000	42.079	-5.2	102	0.00
55	Dibenzofuran	40.000	37.376	6.6	100	0.00
56 P	4-Nitrophenol	40.000	42.069	-5.2	102	0.00
57	2,4-Dinitrotoluene	40.000	41.855	-4.6	102	0.00
58	Fluorene	40.000	36.973	7.6	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.292	1.8	99	0.00
60	Diethylphthalate	40.000	39.043	2.4	101	0.00
61	4-Chlorophenyl-phenylether	40.000	37.632	5.9	101	0.00
62	4-Nitroaniline	40.000	41.065	-2.7	103	0.00
63	Azobenzene	40.000	38.552	3.6	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.991	-12.5	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.186	4.5	101	0.00
67	4-Bromophenyl-phenylether	40.000	39.170	2.1	102	0.00
68	Hexachlorobenzene	40.000	38.600	3.5	101	0.00
69	Atrazine	40.000	39.065	2.3	99	0.00
70 C	Pentachlorophenol	40.000	41.775	-4.4	99	0.00
71	Phenanthrene	40.000	37.659	5.9	99	0.00
72	Anthracene	40.000	37.241	6.9	100	0.00
73	Carbazole	40.000	37.940	5.2	100	0.00
74	Di-n-butylphthalate	40.000	38.818	3.0	100	0.00
75 C	Fluoranthene	40.000	37.361	6.6	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzydine	40.000	44.047	-10.1	98	0.00
78	Pyrene	40.000	40.101	-0.3	100	0.00
79 S	Terphenyl-d14	80.000	76.701	4.1	99	0.00
80	Butylbenzylphthalate	40.000	42.678	-6.7	100	0.00
81	Benzo(a)anthracene	40.000	39.106	2.2	97	0.00
82	3,3'-Dichlorobenzidine	40.000	43.885	-9.7	99	0.00
83	Chrysene	40.000	39.234	1.9	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.985	-5.0	100	0.00
85 c	Di-n-octyl phthalate	40.000	43.898	-9.7	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.923	-9.8	100	0.00
88	Benzo(b)fluoranthene	40.000	38.925	2.7	98	0.00
89	Benzo(k)fluoranthene	40.000	39.740	0.6	105	0.00
90 C	Benzo(a)pyrene	40.000	41.692	-4.2	102	0.00
91	Dibenzo(a,h)anthracene	40.000	44.186	-10.5	100	0.00
92	Benzo(g,h,i)perylene	40.000	43.892	-9.7	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011823\
Data File : BF132121.D
Acq On : 18 Jan 2023 16:37
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 19 05:19:29 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 19 00:42:19 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0