

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011923\
 Data File : BF132149.D
 Acq On : 19 Jan 2023 16:23
 Operator : CG\JU
 Sample : PB150317BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB150317BL

Quant Time: Jan 20 01:19:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 20 00:50:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.089	152	168670	20.000	ng	0.00
21) Naphthalene-d8	8.378	136	619300	20.000	ng	0.00
39) Acenaphthene-d10	10.142	164	335591	20.000	ng	0.00
64) Phenanthrene-d10	11.642	188	662166	20.000	ng	0.00
76) Chrysene-d12	14.301	240	448904	20.000	ng	0.00
86) Perylene-d12	15.895	264	328018	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.719	112	1189503	123.195	ng	0.01
7) Phenol-d6	6.701	99	1522943	125.299	ng	0.00
23) Nitrobenzene-d5	7.654	82	1009533	86.597	ng	0.00
42) 2,4,6-Tribromophenol	10.936	330	461341	128.700	ng	0.00
45) 2-Fluorobiphenyl	9.454	172	1715877	74.713	ng	0.00
79) Terphenyl-d14	13.236	244	2110643	78.760	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011923\
Data File : BF132149.D
Acq On : 19 Jan 2023 16:23
Operator : CG\JU
Sample : PB150317BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB150317BL

Quant Time: Jan 20 01:19:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 20 00:50:42 2023
Response via : Initial Calibration

