

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012422\
 Data File : BF126649.D
 Acq On : 24 Jan 2022 19:51
 Operator : CG/JU
 Sample : N1217-08
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-242

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 01/25/2022
 Supervised By : Jagrut Upadhyay 01/25/2022

Quant Time: Jan 25 03:14:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 19 07:04:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.963	152	90765	20.000	ng	0.00
21) Naphthalene-d8	8.245	136	302655	20.000	ng	# 0.00
39) Acenaphthene-d10	10.122	164	142132m	20.000	ng	0.11
64) Phenanthrene-d10	11.686	188	115367m	20.000	ng	0.19
76) Chrysene-d12	14.251	240	155144	20.000	ng	# 0.10
86) Perylene-d12	15.692	264	28148	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.581	112	280644	40.687	ng	0.00
7) Phenol-d6	6.569	99	370860	38.698	ng	-0.01
23) Nitrobenzene-d5	7.522	82	259136	33.513	ng	-0.01
42) 2,4,6-Tribromophenol	10.933	330	36469m	27.418	ng	0.14
45) 2-Fluorobiphenyl	9.375	172	244213m	26.314	ng	0.05
79) Terphenyl-d14	13.245	244	138349	14.957	ng	0.15
Target Compounds						
15) Benzyl Alcohol	7.092	79	47445	5.553	ng	Qvalue 88
22) Acetophenone	7.363	105	48422	5.192	ng	# 89
32) Benzoic acid	8.022	122	92597m	33.033	ng	
84) Bis(2-ethylhexyl)phtha...	14.186	149	364589	49.389	ng	# 85
85) Di-n-octyl phthalate	14.821	149	1711535	151.437	ng	# 58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012422\
Data File : BF126649.D
Acq On : 24 Jan 2022 19:51
Operator : CG/JU
Sample : N1217-08
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
VNJ-242

Quant Time: Jan 25 03:14:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 19 07:04:14 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 01/25/2022
Supervised By : Jagrut Upadhyay 01/25/2022

