

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012422\
 Data File : BF126628.D
 Acq On : 24 Jan 2022 09:55
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 24 13:07:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 19 07:04:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	83	0.00
2	1,4-Dioxane	40.000	39.413	1.5	88	0.00
3	Pyridine	40.000	39.022	2.4	88	0.00
4	n-Nitrosodimethylamine	40.000	37.608	6.0	85	0.00
5 S	2-Fluorophenol	80.000	75.475	5.7	87	0.00
6	Aniline	40.000	37.527	6.2	85	0.00
7 S	Phenol-d6	80.000	75.843	5.2	87	0.00
8	2-Chlorophenol	40.000	37.310	6.7	85	0.01
9	Benzaldehyde	40.000	40.967	-2.4	87	0.01
10 C	Phenol	40.000	37.779	5.6	86	0.00
11	bis(2-Chloroethyl)ether	40.000	38.339	4.2	87	0.00
12	1,3-Dichlorobenzene	40.000	38.115	4.7	87	0.00
13 C	1,4-Dichlorobenzene	40.000	38.183	4.5	87	0.01
14	1,2-Dichlorobenzene	40.000	38.403	4.0	88	0.01
15	Benzyl Alcohol	40.000	37.997	5.0	86	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.772	8.1	83	0.00
17	2-Methylphenol	40.000	37.397	6.5	85	0.00
18	Hexachloroethane	40.000	37.445	6.4	85	0.01
19 P	n-Nitroso-di-n-propylamine	40.000	36.827	7.9	85	0.00
20	3+4-Methylphenols	40.000	38.418	4.0	88	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	82	0.01
22	Acetophenone	40.000	37.085	7.3	86	0.00
23 S	Nitrobenzene-d5	80.000	82.282	-2.9	90	0.00
24	Nitrobenzene	40.000	40.389	-1.0	88	0.00
25	Isophorone	40.000	37.429	6.4	86	0.00
26 C	2-Nitrophenol	40.000	39.095	2.3	81	0.00
27	2,4-Dimethylphenol	40.000	36.486	8.8	84	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.163	7.1	86	0.00
29 C	2,4-Dichlorophenol	40.000	37.798	5.5	86	0.00
30	1,2,4-Trichlorobenzene	40.000	37.909	5.2	87	0.00
31	Naphthalene	40.000	37.515	6.2	86	0.01
32	Benzoic acid	40.000	32.980	17.6	80	0.00
33	4-Chloroaniline	40.000	37.833	5.4	87	0.00
34 C	Hexachlorobutadiene	40.000	38.038	4.9	87	0.01
35	Caprolactam	40.000	40.406	-1.0	92	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.624	3.4	88	0.00
37	2-Methylnaphthalene	40.000	37.582	6.0	88	0.01
38	1-Methylnaphthalene	40.000	37.277	6.8	87	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	82	0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	38.282	4.3	89	0.01
41 P	Hexachlorocyclopentadiene	40.000	36.072	9.8	80	0.00
42 S	2,4,6-Tribromophenol	80.000	80.849	-1.1	94	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.519	6.2	86	0.00
44	2,4,5-Trichlorophenol	40.000	38.279	4.3	89	0.00
45 S	2-Fluorobiphenyl	80.000	73.810	7.7	89	0.01
46	1,1'-Biphenyl	40.000	36.999	7.5	87	0.00
47	2-Chloronaphthalene	40.000	37.534	6.2	89	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012422\
 Data File : BF126628.D
 Acq On : 24 Jan 2022 09:55
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 24 13:07:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 19 07:04:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.493	-16.2	98	0.00
49	Acenaphthylene	40.000	37.472	6.3	89	0.01
50	Dimethylphthalate	40.000	38.364	4.1	91	0.00
51	2,6-Dinitrotoluene	40.000	43.258	-8.1	93	0.00
52 C	Acenaphthene	40.000	38.154	4.6	90	0.01
53	3-Nitroaniline	40.000	45.034	-12.6	97	0.01
54 P	2,4-Dinitrophenol	40.000	40.756	-1.9	90	0.00
55	Dibenzofuran	40.000	38.016	5.0	91	0.00
56 P	4-Nitrophenol	40.000	44.401	-11.0	95	0.00
57	2,4-Dinitrotoluene	40.000	48.258	-20.6	101	0.00
58	Fluorene	40.000	37.526	6.2	90	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.534	1.2	90	0.00
60	Diethylphthalate	40.000	39.051	2.4	92	0.01
61	4-Chlorophenyl-phenylether	40.000	37.608	6.0	90	0.01
62	4-Nitroaniline	40.000	46.804	-17.0	102	0.00
63	Azobenzene	40.000	38.961	2.6	92	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	89	0.01
65	4,6-Dinitro-2-methylphenol	40.000	38.095	4.8	94	0.00
66 c	n-Nitrosodiphenylamine	40.000	35.678	10.8	90	0.00
67	4-Bromophenyl-phenylether	40.000	36.818	8.0	93	0.01
68	Hexachlorobenzene	40.000	37.791	5.5	96	0.00
69	Atrazine	40.000	38.246	4.4	97	0.00
70 C	Pentachlorophenol	40.000	38.138	4.7	92	0.01
71	Phenanthrene	40.000	36.983	7.5	95	0.00
72	Anthracene	40.000	37.011	7.5	95	0.01
73	Carbazole	40.000	37.323	6.7	96	0.01
74	Di-n-butylphthalate	40.000	37.268	6.8	94	0.00
75 C	Fluoranthene	40.000	38.735	3.2	99	0.01
76 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
77	Benzydine	40.000	36.717	8.2	106	0.01
78	Pyrene	40.000	36.442	8.9	99	0.01
79 S	Terphenyl-d14	80.000	74.555	6.8	103	0.01
80	Butylbenzylphthalate	40.000	37.734	5.7	99	0.00
81	Benzo(a)anthracene	40.000	38.473	3.8	105	0.00
82	3,3'-Dichlorobenzidine	40.000	38.117	4.7	104	0.01
83	Chrysene	40.000	38.150	4.6	106	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	38.162	4.6	102	0.01
85 c	Di-n-octyl phthalate	40.000	36.677	8.3	102	0.02
86 I	Perylene-d12	20.000	20.000	0.0	92	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	37.129	7.2	92	0.02
88	Benzo(b)fluoranthene	40.000	39.599	1.0	102	0.02
89	Benzo(k)fluoranthene	40.000	41.825	-4.6	103	0.02
90 C	Benzo(a)pyrene	40.000	39.376	1.6	100	0.02
91	Dibenzo(a,h)anthracene	40.000	37.618	6.0	92	0.02
92	Benzo(g,h,i)perylene	40.000	37.227	6.9	91	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012422\
Data File : BF126628.D
Acq On : 24 Jan 2022 09:55
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 24 13:07:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 19 07:04:14 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0