




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020122\
Data File : BF126766.D
Acq On : 01 Feb 2022 10:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 02 06:18:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 27 03:09:40 2022
Response via : Initial Calibration



TIC: BF126766.D\data.ms

(70) Pentachlorophenol (C)




11.286min (+ 0.000) 35597.31 ng

response 166464

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.40	62.23
264.00	63.00	62.69
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020122\
Data File : BF126766.D
Acq On : 01 Feb 2022 10:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 02 06:18:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 27 03:09:40 2022
Response via : Initial Calibration



TIC: BF126766.D\data.ms

(77) Benzidine**12.833min (+ 0.006) 0.00 ng****response 804503**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.80	14.85
183.00	10.20	11.54
0.00	0.00	0.00