


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020122\
 Data File : BF126766.D
 Acq On : 01 Feb 2022 10:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jan 27 03:09:40 2022






AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1634

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.9	59051	PASS
68	69	0.00	2	1.5	1088	PASS
69	198	0.00	100	46.7	74680	PASS
70	69	0.00	2	0.4	288	PASS
127	198	10	80	48.2	77032	PASS
197	198	0.00	2	0.5	796	PASS
198	198	100	100	100.0	159843	PASS
199	198	5	9	6.6	10569	PASS
275	198	10	60	25.0	40011	PASS
365	198	1	100	3.4	5355	PASS
441	198	0.01	100	13.7	21891	PASS
442	442	50	100	100.0	137387	PASS
443	442	15	24	19.1	26174	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020122\
Data File : BF126766.D
Acq On : 01 Feb 2022 10:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Feb 02 06:18:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 27 03:09:40 2022
Response via : Initial Calibration



TIC: BF126766.D\data.ms

(70) Pentachlorophenol (C)

11.286min (+ 0.000) 35597.31 ng




response 166464

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.40	62.23
264.00	63.00	62.69
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020122\
Data File : BF126766.D
Acq On : 01 Feb 2022 10:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Feb 02 06:18:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 27 03:09:40 2022
Response via : Initial Calibration



TIC: BF126766.D\data.ms

(77) Benzidine

12.833min (+ 0.006) 0.00 ng

response 804503

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.80	14.85
183.00	10.20	11.54
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/1/2022	BNA_F	<u>BF126766.D</u>
Compound Name	Response	Retention Time
DDT	382764	13.657
DDD	5501	13.351
DDE	1672	13.016
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7173	389937	1.84

Instrument :
BNA_F
ClientSampleId :
DFTPP