




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020123\
Data File : BF132261.D
Acq On : 01 Feb 2023 15:38
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 02 05:16:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 02 01:30:34 2023
Response via : Initial Calibration



TIC: BF132261.D\data.ms

(70) Pentachlorophenol (C)




11.413min (-0.006) 569426.91 ng

response 244082

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.20	66.36
264.00	63.10	63.60
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020123\
Data File : BF132261.D
Acq On : 01 Feb 2023 15:38
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 02 05:16:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 02 01:30:34 2023
Response via : Initial Calibration



TIC: BF132261.D\data.ms

(77) Benzidine

12.966min (+ 0.006) 281278.87 ng

response 1248895

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.70	14.65
183.00	11.10	11.32
0.00	0.00	0.00