

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020123\
 Data File : BF132251.D
 Acq On : 01 Feb 2023 10:09
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 02 01:32:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 02 01:30:34 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	39.802	0.5	98	0.00
3	Pyridine	40.000	39.341	1.6	96	0.00
4	n-Nitrosodimethylamine	40.000	37.466	6.3	93	0.00
5 S	2-Fluorophenol	80.000	77.618	3.0	97	0.00
6	Aniline	40.000	38.181	4.5	96	0.00
7 S	Phenol-d6	80.000	76.733	4.1	96	0.00
8	2-Chlorophenol	40.000	39.709	0.7	97	0.00
9	Benzaldehyde	40.000	37.634	5.9	95	0.00
10 C	Phenol	40.000	38.564	3.6	96	0.00
11	bis(2-Chloroethyl)ether	40.000	38.968	2.6	97	0.00
12	1,3-Dichlorobenzene	40.000	39.937	0.2	99	0.00
13 C	1,4-Dichlorobenzene	40.000	39.655	0.9	98	0.00
14	1,2-Dichlorobenzene	40.000	39.624	0.9	98	0.00
15	Benzyl Alcohol	40.000	38.500	3.8	94	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.817	5.5	93	0.00
17	2-Methylphenol	40.000	38.976	2.6	97	0.00
18	Hexachloroethane	40.000	39.728	0.7	98	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.409	9.0	92	0.00
20	3+4-Methylphenols	40.000	39.404	1.5	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	38.714	3.2	96	0.00
23 S	Nitrobenzene-d5	80.000	76.426	4.5	94	0.00
24	Nitrobenzene	40.000	38.840	2.9	94	0.00
25	Isophorone	40.000	37.916	5.2	94	0.00
26 C	2-Nitrophenol	40.000	41.205	-3.0	97	0.00
27	2,4-Dimethylphenol	40.000	39.755	0.6	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.276	1.8	96	0.00
29 C	2,4-Dichlorophenol	40.000	40.056	-0.1	98	0.00
30	1,2,4-Trichlorobenzene	40.000	39.956	0.1	98	0.00
31	Naphthalene	40.000	39.775	0.6	98	0.00
32	Benzoic acid	40.000	40.843	-2.1	97	0.00
33	4-Chloroaniline	40.000	39.739	0.7	96	0.00
34 C	Hexachlorobutadiene	40.000	40.731	-1.8	100	0.00
35	Caprolactam	40.000	39.704	0.7	96	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.647	0.9	96	0.00
37	2-Methylnaphthalene	40.000	39.371	1.6	96	0.00
38	1-Methylnaphthalene	40.000	39.862	0.3	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.309	-0.8	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.665	-4.2	98	0.00
42 S	2,4,6-Tribromophenol	80.000	80.146	-0.2	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.834	-2.1	97	0.00
44	2,4,5-Trichlorophenol	40.000	40.662	-1.7	97	0.00
45 S	2-Fluorobiphenyl	80.000	74.261	7.2	97	0.00
46	1,1'-Biphenyl	40.000	39.928	0.2	97	0.00
47	2-Chloronaphthalene	40.000	39.957	0.1	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020123\
 Data File : BF132251.D
 Acq On : 01 Feb 2023 10:09
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 02 01:32:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 02 01:30:34 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.292	1.8	92	0.00
49	Acenaphthylene	40.000	39.989	0.0	96	0.00
50	Dimethylphthalate	40.000	39.572	1.1	96	0.00
51	2,6-Dinitrotoluene	40.000	41.223	-3.1	98	0.00
52 C	Acenaphthene	40.000	39.565	1.1	97	0.00
53	3-Nitroaniline	40.000	40.530	-1.3	96	0.00
54 P	2,4-Dinitrophenol	40.000	38.028	4.9	93	0.00
55	Dibenzofuran	40.000	39.645	0.9	97	0.00
56 P	4-Nitrophenol	40.000	41.572	-3.9	96	0.00
57	2,4-Dinitrotoluene	40.000	41.491	-3.7	97	0.00
58	Fluorene	40.000	39.459	1.4	98	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.877	-2.2	96	0.00
60	Diethylphthalate	40.000	40.018	-0.0	96	0.00
61	4-Chlorophenyl-phenylether	40.000	39.708	0.7	97	0.00
62	4-Nitroaniline	40.000	39.660	0.9	95	0.00
63	Azobenzene	40.000	38.891	2.8	93	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.735	-6.8	97	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.966	2.6	95	0.00
67	4-Bromophenyl-phenylether	40.000	39.851	0.4	98	0.00
68	Hexachlorobenzene	40.000	39.648	0.9	97	0.00
69	Atrazine	40.000	41.592	-4.0	97	0.00
70 C	Pentachlorophenol	40.000	40.987	-2.5	97	0.00
71	Phenanthrene	40.000	39.069	2.3	97	0.00
72	Anthracene	40.000	39.294	1.8	96	0.00
73	Carbazole	40.000	39.218	2.0	97	0.00
74	Di-n-butylphthalate	40.000	39.652	0.9	96	0.00
75 C	Fluoranthene	40.000	39.295	1.8	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzidine	40.000	37.583	6.0	97	0.00
78	Pyrene	40.000	38.824	2.9	97	0.00
79 S	Terphenyl-d14	80.000	75.674	5.4	97	0.00
80	Butylbenzylphthalate	40.000	39.510	1.2	96	0.00
81	Benzo(a)anthracene	40.000	39.264	1.8	96	0.00
82	3,3'-Dichlorobenzidine	40.000	40.705	-1.8	98	0.00
83	Chrysene	40.000	39.428	1.4	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.300	1.8	96	0.00
85 c	Di-n-octyl phthalate	40.000	39.707	0.7	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.656	3.4	100	0.00
88	Benzo(b)fluoranthene	40.000	39.348	1.6	95	0.00
89	Benzo(k)fluoranthene	40.000	41.858	-4.6	105	0.00
90 C	Benzo(a)pyrene	40.000	40.026	-0.1	99	0.00
91	Dibenzo(a,h)anthracene	40.000	39.570	1.1	101	0.00
92	Benzo(g,h,i)perylene	40.000	37.938	5.2	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020123\
Data File : BF132251.D
Acq On : 01 Feb 2023 10:09
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 02 01:32:09 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 02 01:30:34 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0