

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020123\
 Data File : BF132262.D
 Acq On : 01 Feb 2023 16:11
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 02 00:54:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 05:24:37 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	113	0.00
2	1,4-Dioxane	40.000	41.433	-3.6	115	0.02
3	Pyridine	40.000	39.697	0.8	109	0.02
4	n-Nitrosodimethylamine	40.000	37.902	5.2	106	0.02
5 S	2-Fluorophenol	80.000	77.795	2.8	110	0.00
6	Aniline	40.000	38.429	3.9	109	0.00
7 S	Phenol-d6	80.000	76.493	4.4	108	0.00
8	2-Chlorophenol	40.000	39.563	1.1	110	0.00
9	Benzaldehyde	40.000	37.731	5.7	107	0.00
10 C	Phenol	40.000	38.441	3.9	108	0.00
11	bis(2-Chloroethyl)ether	40.000	38.312	4.2	108	0.00
12	1,3-Dichlorobenzene	40.000	39.577	1.1	111	0.00
13 C	1,4-Dichlorobenzene	40.000	39.828	0.4	112	0.00
14	1,2-Dichlorobenzene	40.000	39.153	2.1	110	0.00
15	Benzyl Alcohol	40.000	38.052	4.9	105	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.601	8.5	102	0.00
17	2-Methylphenol	40.000	38.572	3.6	108	0.00
18	Hexachloroethane	40.000	38.775	3.1	108	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.957	12.6	100	0.00
20	3+4-Methylphenols	40.000	38.603	3.5	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	108	0.00
22	Acetophenone	40.000	38.186	4.5	104	0.00
23 S	Nitrobenzene-d5	80.000	76.886	3.9	104	0.00
24	Nitrobenzene	40.000	38.861	2.8	104	0.00
25	Isophorone	40.000	37.499	6.3	103	0.00
26 C	2-Nitrophenol	40.000	42.416	-6.0	110	0.00
27	2,4-Dimethylphenol	40.000	40.131	-0.3	109	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.427	3.9	104	0.00
29 C	2,4-Dichlorophenol	40.000	40.514	-1.3	109	0.00
30	1,2,4-Trichlorobenzene	40.000	40.256	-0.6	109	0.00
31	Naphthalene	40.000	39.105	2.2	106	0.00
32	Benzoic acid	40.000	39.842	0.4	105	0.00
33	4-Chloroaniline	40.000	39.226	1.9	105	0.00
34 C	Hexachlorobutadiene	40.000	40.607	-1.5	111	0.00
35	Caprolactam	40.000	39.204	2.0	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.009	2.5	105	0.00
37	2-Methylnaphthalene	40.000	38.838	2.9	105	0.00
38	1-Methylnaphthalene	40.000	39.020	2.4	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.859	-2.1	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.873	-4.7	106	0.00
42 S	2,4,6-Tribromophenol	80.000	80.529	-0.7	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.968	-2.4	105	0.00
44	2,4,5-Trichlorophenol	40.000	41.378	-3.4	107	0.00
45 S	2-Fluorobiphenyl	80.000	73.721	7.8	104	0.00
46	1,1'-Biphenyl	40.000	39.971	0.1	104	0.00
47	2-Chloronaphthalene	40.000	40.068	-0.2	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020123\
 Data File : BF132262.D
 Acq On : 01 Feb 2023 16:11
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 02 00:54:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 05:24:37 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.032	2.4	98	0.00
49	Acenaphthylene	40.000	39.748	0.6	103	0.00
50	Dimethylphthalate	40.000	39.433	1.4	102	0.00
51	2,6-Dinitrotoluene	40.000	41.155	-2.9	105	0.00
52 C	Acenaphthene	40.000	39.389	1.5	103	0.00
53	3-Nitroaniline	40.000	40.545	-1.4	103	0.00
54 P	2,4-Dinitrophenol	40.000	39.178	2.1	103	0.00
55	Dibenzofuran	40.000	39.141	2.1	103	0.00
56 P	4-Nitrophenol	40.000	40.918	-2.3	102	0.00
57	2,4-Dinitrotoluene	40.000	41.338	-3.3	104	0.00
58	Fluorene	40.000	38.404	4.0	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.853	-2.1	103	0.00
60	Diethylphthalate	40.000	39.529	1.2	102	0.00
61	4-Chlorophenyl-phenylether	40.000	39.057	2.4	103	0.00
62	4-Nitroaniline	40.000	39.437	1.4	101	0.00
63	Azobenzene	40.000	37.436	6.4	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.610	-11.5	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.745	0.6	102	0.00
67	4-Bromophenyl-phenylether	40.000	40.285	-0.7	104	0.00
68	Hexachlorobenzene	40.000	40.459	-1.1	104	0.00
69	Atrazine	40.000	42.069	-5.2	103	0.00
70 C	Pentachlorophenol	40.000	42.296	-5.7	105	0.00
71	Phenanthrene	40.000	39.294	1.8	102	0.00
72	Anthracene	40.000	39.231	1.9	101	0.00
73	Carbazole	40.000	39.906	0.2	104	0.00
74	Di-n-butylphthalate	40.000	40.124	-0.3	103	0.00
75 C	Fluoranthene	40.000	39.699	0.8	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzidine	40.000	37.817	5.5	98	0.00
78	Pyrene	40.000	41.066	-2.7	102	0.00
79 S	Terphenyl-d14	80.000	80.673	-0.8	104	0.00
80	Butylbenzylphthalate	40.000	42.401	-6.0	103	0.00
81	Benzo(a)anthracene	40.000	39.727	0.7	98	0.00
82	3,3'-Dichlorobenzidine	40.000	39.276	1.8	95	0.00
83	Chrysene	40.000	39.839	0.4	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.414	-3.5	102	0.00
85 c	Di-n-octyl phthalate	40.000	36.660	8.4	89	0.00
86 I	Perylene-d12	20.000	20.000	0.0	92	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.265	-3.2	98	0.00
88	Benzo(b)fluoranthene	40.000	39.456	1.4	88	0.00
89	Benzo(k)fluoranthene	40.000	40.731	-1.8	94	0.00
90 C	Benzo(a)pyrene	40.000	39.968	0.1	91	0.00
91	Dibenzo(a,h)anthracene	40.000	41.680	-4.2	98	-0.01
92	Benzo(g,h,i)perylene	40.000	40.727	-1.8	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020123\
Data File : BF132262.D
Acq On : 01 Feb 2023 16:11
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 02 00:54:30 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 01 05:24:37 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0