

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020323\
 Data File : BF132329.D
 Acq On : 03 Feb 2023 19:07
 Operator : CG\JU
 Sample : 01415-06
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-1-0202

Quant Time: Feb 04 01:40:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 04 01:32:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.054	152	225653	20.000	ng	0.00
21) Naphthalene-d8	8.342	136	850412	20.000	ng	0.00
39) Acenaphthene-d10	10.107	164	426289	20.000	ng	0.00
64) Phenanthrene-d10	11.607	188	726144	20.000	ng	0.00
76) Chrysene-d12	14.266	240	369115	20.000	ng	0.00
86) Perylene-d12	15.848	264	367050	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.678	112	714951	50.928	ng	0.00
7) Phenol-d6	6.666	99	529258	30.580	ng	-0.01
23) Nitrobenzene-d5	7.625	82	1118689	73.202	ng	0.00
42) 2,4,6-Tribromophenol	10.907	330	560979	127.966	ng	0.00
45) 2-Fluorobiphenyl	9.425	172	1957441	70.909	ng	0.00
79) Terphenyl-d14	13.201	244	1885177	98.796	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020323\
Data File : BF132329.D
Acq On : 03 Feb 2023 19:07
Operator : CG\JU
Sample : 01415-06
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-1-0202

Quant Time: Feb 04 01:40:55 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 04 01:32:42 2023
Response via : Initial Calibration

