


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020720\
 Data File : BF118844.D
 Acq On : 7 Feb 2020 9:34
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF020320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Feb 03 16:26:03 2020



AutoFind: Scans 1616, 1617, 1618; Background Corrected with Scan 1611

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.2	56765	PASS
68	69	0.00	2	2.0	1286	PASS
69	198	0.00	100	31.6	63536	PASS
70	69	0.00	2	0.5	287	PASS
127	198	10	80	44.0	88514	PASS
197	198	0.00	2	0.6	1250	PASS
198	198	100	100	100.0	201021	PASS
199	198	5	9	6.8	13752	PASS
275	198	10	60	32.2	64704	PASS
365	198	1	100	3.9	7925	PASS
441	198	0.01	100	15.1	30424	PASS
442	198	50	100	99.0	199037	PASS
443	442	15	24	18.2	36317	PASS