

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF021519\
 Data File : BF112584.D
 Acq On : 15 Feb 2019 17:47
 Operator : JU/SJ
 Sample : K1487-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RHR-1

Quant Time: Feb 15 23:52:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 28 10:38:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	82776	20.00	ng	-0.02
21) Naphthalene-d8	8.11	136	343732	20.00	ng	-0.02
39) Acenaphthene-d10	9.87	164	179050	20.00	ng	-0.02
64) Phenanthrene-d10	11.36	188	367468	20.00	ng	0.00
76) Chrysene-d12	14.01	240	312155	20.00	ng	0.00
87) Perylene-d12	15.49	264	274339	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	361200	92.69	ng	-0.01
7) Phenol-d6	6.49	99	483275	89.24	ng	0.00
23) Nitrobenzene-d5	7.39	82	300389	50.35	ng	-0.02
42) 2,4,6-Tribromophenol	10.67	330	173699	83.35	ng	-0.01
45) 2-Fluorobiphenyl	9.19	172	680793	53.78	ng	-0.02
79) Terphenyl-d14	12.95	244	859770	51.88	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.13	128	33533	2.086	ng	97
50) Dimethylphthalate	9.58	163	28069	2.018	ng	98
55) Dibenzofuran	10.07	168	59096	3.639	ng	95
70) Pentachlorophenol	11.19	266	21303	11.807	ng	99
71) Phenanthrene	11.39	178	958084	50.151	ng	98
73) Carbazole	11.60	167	39010	2.078	ng	97
75) Fluoranthene	12.58	202	286689	13.991	ng	97
78) Pyrene	12.81	202	133017	6.155	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF021519\
Data File : BF112584.D
Acq On : 15 Feb 2019 17:47
Operator : JU/SJ
Sample : K1487-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RHR-1

Quant Time: Feb 15 23:52:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 28 10:38:18 2019
Response via : Initial Calibration

