

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021723\
 Data File : BF132461.D
 Acq On : 17 Feb 2023 21:11
 Operator : CG\JU
 Sample : 01595-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S-322-2

Quant Time: Feb 18 01:02:02 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 18 00:58:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.019	152	243864	20.000	ng	0.00
21) Naphthalene-d8	8.307	136	909621	20.000	ng	0.00
39) Acenaphthene-d10	10.072	164	488396	20.000	ng	0.00
64) Phenanthrene-d10	11.566	188	947047	20.000	ng	0.00
76) Chrysene-d12	14.218	240	705007	20.000	ng	#-0.01
86) Perylene-d12	15.777	264	664781	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.648	112	1413356	92.868	ng	0.00
7) Phenol-d6	6.643	99	1842055	92.694	ng	0.00
23) Nitrobenzene-d5	7.584	82	1184271	62.557	ng	-0.01
42) 2,4,6-Tribromophenol	10.866	330	464581	95.339	ng	0.00
45) 2-Fluorobiphenyl	9.383	172	1812114	57.299	ng	0.00
79) Terphenyl-d14	13.160	244	2184523	56.432	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021723\
Data File : BF132461.D
Acq On : 17 Feb 2023 21:11
Operator : CG\JU
Sample : 01595-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S-322-2

Quant Time: Feb 18 01:02:02 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 18 00:58:25 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.019 min Scan# 713
Delta R.T. -0.006 min
Lab File: BF132461.D
Acq: 17 Feb 2023 21:11




Instrument :
BNA_F
ClientSampleId :
S-322-2


Tgt Ion:152 Resp: 243864
Ion Ratio Lower Upper
152 100
150 154.8 123.8 185.8
115 63.9 50.4 75.6



#5
2-Fluorophenol
Concen: 92.868 ng
RT: 5.648 min Scan# 478
Delta R.T. 0.006 min
Lab File: BF132461.D
Acq: 17 Feb 2023 21:11

Tgt Ion:112 Resp: 1413356
Ion Ratio Lower Upper
112 100
64 64.5 49.7 74.5
63 34.3 25.8 38.8









#7
Phenol-d6
Concen: 92.694 ng
RT: 6.643 min Scan# 64
Delta R.T. -0.006 min
Lab File: BF132461.D
Acq: 17 Feb 2023 21:11




Instrument :
BNA_F
ClientSampleId :
S-322-2


Tgt Ion: 99 Resp: 1842055
Ion Ratio Lower Upper
99 100
42 17.4 14.6 21.8
71 35.9 28.7 43.1



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.307 min Scan# 930
Delta R.T. -0.006 min
Lab File: BF132461.D
Acq: 17 Feb 2023 21:11

Tgt Ion:136 Resp: 909621
Ion Ratio Lower Upper
136 100
137 11.3 8.8 13.2
54 8.7 6.9 10.3
68 7.1 5.4 8.0





#23

Nitrobenzene-d5

Concen: 62.557 ng

RT: 7.584 min Scan# 80

Delta R.T. -0.012 min

Lab File: BF132461.D

Acq: 17 Feb 2023 21:11

Instrument :





BNA_F

ClientSampleId :

S-322-2

Tgt Ion: 82 Resp: 1184271

Ion	Ratio	Lower	Upper
82	100		
128	39.5	30.8	46.2
54	48.9	40.2	60.4



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 10.072 min Scan# 1230




Delta R.T. -0.006 min


Lab File: BF132461.D

Acq: 17 Feb 2023 21:11

Tgt Ion:164 Resp: 488396

Ion	Ratio	Lower	Upper
164	100		
162	100.8	80.8	121.2
160	46.0	37.2	55.8





#42

2,4,6-Tribromophenol

Concen: 95.339 ng

RT: 10.866 min Scan# 1365

Delta R.T. -0.006 min

Lab File: BF132461.D

Acq: 17 Feb 2023 21:11

Instrument :

BNA_F

ClientSampleId :

S-322-2





Tgt Ion:330 Resp: 464581

Ion Ratio Lower Upper

330 100

332 96.6 76.3 114.5

141 36.9 30.3 45.5



#45

2-Fluorobiphenyl

Concen: 57.299 ng

RT: 9.383 min Scan# 1113

Delta R.T. -0.006 min

Lab File: BF132461.D

Acq: 17 Feb 2023 21:11




Tgt Ion:172 Resp: 1812114


Ion Ratio Lower Upper

172 100

171 36.0 28.6 42.8

170 24.6 19.9 29.9









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.566 min Scan# 1486
Delta R.T. -0.006 min
Lab File: BF132461.D
Acq: 17 Feb 2023 21:11




Instrument :
BNA_F
ClientSampleId :
S-322-2


Tgt Ion:188 Resp: 947047
Ion Ratio Lower Upper
188 100
94 10.4 8.0 12.0
80 10.6 8.6 12.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.218 min Scan# 1935
Delta R.T. -0.012 min
Lab File: BF132461.D
Acq: 17 Feb 2023 21:11


Tgt Ion:240 Resp: 705007
Ion Ratio Lower Upper
240 100
120 12.3 7.5 11.3#
236 26.7 21.0 31.6








#79
 Terphenyl-d14
 Concen: 56.432 ng
 RT: 13.160 min Scan# 1755
 Delta R.T. 0.000 min
 Lab File: BF132461.D
 Acq: 17 Feb 2023 21:11

Instrument :
 BNA_F
 ClientSampleId :
 S-322-2



Tgt Ion:244 Resp: 2184523
 Ion Ratio Lower Upper
 244 100
 212 8.0 6.7 10.1
 122 9.3 7.7 11.5



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.777 min Scan# 2200
 Delta R.T. -0.006 min
 Lab File: BF132461.D
 Acq: 17 Feb 2023 21:11

Tgt Ion:264 Resp: 664781
 Ion Ratio Lower Upper
 264 100
 260 24.4 20.0 30.0
 265 22.8 17.2 25.8

