

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022021\
 Data File : BF123251.D
 Acq On : 20 Feb 2021 18:18
 Operator : JU/CG
 Sample : M1425-12
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WATER-COMP

Quant Time: Feb 22 03:17:18 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 16:08:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	124039	20.00	ng	0.00
21) Naphthalene-d8	8.145	136	530356	20.00	ng	0.00
39) Acenaphthene-d10	9.904	164	304563	20.00	ng	0.00
64) Phenanthrene-d10	11.392	188	443546	20.00	ng	# 0.00
76) Chrysene-d12	14.051	240	281414	20.00	ng	0.00
86) Perylene-d12	15.527	264	270265	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	273578	31.75	ng	0.00
7) Phenol-d6	6.510	99	265648	22.69	ng	0.02
23) Nitrobenzene-d5	7.422	82	611360	52.77	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	139836	45.09	ng	0.00
45) 2-Fluorobiphenyl	9.227	172	1187860	56.02	ng	0.00
79) Terphenyl-d14	12.992	244	817059	55.83	ng	0.00
Target Compounds						
10) Phenol	6.528	94	133240	10.44	ng	Qvalue 76
20) 3+4-Methylphenols	7.263	107	24657	2.41	ng	# 84
59) 2,3,4,6-Tetrachlorophenol	10.239	232	16308	2.96	ng	# 87
60) Diethylphthalate	10.322	149	62373	2.77	ng	97
70) Pentachlorophenol	11.210	266	29851	9.28	ng	95
84) Bis(2-ethylhexyl)phtha...	14.033	149	159703	12.49	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022021\
Data File : BF123251.D
Acq On : 20 Feb 2021 18:18
Operator : JU/CG
Sample : M1425-12
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WATER-COMP

Quant Time: Feb 22 03:17:18 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration

