

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022021\
 Data File : BF123240.D
 Acq On : 20 Feb 2021 10:26
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 22 03:14:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 16:08:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	129	0.00
2	1,4-Dioxane	40.000	32.340	19.1	102	0.02
3	Pyridine	40.000	35.619	11.0	101	0.02
4	n-Nitrosodimethylamine	40.000	35.804	10.5	105	0.02
5 S	2-Fluorophenol	80.000	79.268	0.9	132	0.00
6	Aniline	40.000	41.135	-2.8	137	0.00
7 S	Phenol-d6	80.000	79.811	0.2	133	0.00
8	2-Chlorophenol	40.000	39.620	1.0	130	0.00
9	Benzaldehyde	40.000	45.102	-12.8	139	0.00
10 C	Phenol	40.000	40.410	-1.0	135	0.00
11	bis(2-Chloroethyl)ether	40.000	40.668	-1.7	138	0.00
12	1,3-Dichlorobenzene	40.000	39.233	1.9	128	0.00
13 C	1,4-Dichlorobenzene	40.000	39.050	2.4	128	0.00
14	1,2-Dichlorobenzene	40.000	38.989	2.5	128	0.00
15	Benzyl Alcohol	40.000	41.284	-3.2	134	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.106	-5.3	146	0.00
17	2-Methylphenol	40.000	39.937	0.2	133	0.00
18	Hexachloroethane	40.000	41.472	-3.7	133	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.402	-3.5	135	0.00
20	3+4-Methylphenols	40.000	37.522	6.2	130	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	131	0.00
22	Acetophenone	40.000	38.976	2.6	131	0.00
23 S	Nitrobenzene-d5	80.000	80.178	-0.2	134	0.00
24	Nitrobenzene	40.000	40.783	-2.0	137	0.00
25	Isophorone	40.000	40.632	-1.6	138	0.00
26 C	2-Nitrophenol	40.000	41.560	-3.9	134	0.00
27	2,4-Dimethylphenol	40.000	39.929	0.2	134	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.869	-2.2	139	0.00
29 C	2,4-Dichlorophenol	40.000	39.420	1.4	132	0.00
30	1,2,4-Trichlorobenzene	40.000	38.450	3.9	127	0.00
31	Naphthalene	40.000	38.585	3.5	129	0.00
32	Benzoic acid	40.000	44.139	-10.3	138	0.00
33	4-Chloroaniline	40.000	39.047	2.4	132	0.00
34 C	Hexachlorobutadiene	40.000	37.568	6.1	124	0.00
35	Caprolactam	40.000	40.314	-0.8	137	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.792	0.5	133	0.00
37	2-Methylnaphthalene	40.000	38.627	3.4	130	0.00
38	1-Methylnaphthalene	40.000	38.863	2.8	131	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	132	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.455	6.4	124	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.423	1.4	118	0.00
42 S	2,4,6-Tribromophenol	80.000	74.449	6.9	120	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.381	1.5	128	0.00
44	2,4,5-Trichlorophenol	40.000	39.821	0.4	132	0.00
45 S	2-Fluorobiphenyl	80.000	74.294	7.1	126	0.00
46	1,1'-Biphenyl	40.000	38.641	3.4	130	0.00
47	2-Chloronaphthalene	40.000	38.686	3.3	130	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022021\
 Data File : BF123240.D
 Acq On : 20 Feb 2021 10:26
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 22 03:14:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 16:08:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.386	-6.0	138	0.00
49	Acenaphthylene	40.000	38.777	3.1	130	0.00
50	Dimethylphthalate	40.000	38.457	3.9	131	0.00
51	2,6-Dinitrotoluene	40.000	40.361	-0.9	132	0.00
52 C	Acenaphthene	40.000	39.288	1.8	131	0.00
53	3-Nitroaniline	40.000	40.085	-0.2	131	0.00
54 P	2,4-Dinitrophenol	40.000	40.884	-2.2	129	0.00
55	Dibenzofuran	40.000	38.380	4.0	130	0.00
56 P	4-Nitrophenol	40.000	41.137	-2.8	129	0.00
57	2,4-Dinitrotoluene	40.000	39.324	1.7	128	0.00
58	Fluorene	40.000	37.202	7.0	125	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.781	3.0	124	0.00
60	Diethylphthalate	40.000	38.064	4.8	128	0.00
61	4-Chlorophenyl-phenylether	40.000	36.736	8.2	124	0.00
62	4-Nitroaniline	40.000	39.484	1.3	130	0.00
63	Azobenzene	40.000	39.575	1.1	135	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	127	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.183	-8.0	125	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.365	1.6	127	0.00
67	4-Bromophenyl-phenylether	40.000	38.411	4.0	124	0.00
68	Hexachlorobenzene	40.000	37.843	5.4	120	0.00
69	Atrazine	40.000	38.106	4.7	122	0.00
70 C	Pentachlorophenol	40.000	40.501	-1.3	121	0.00
71	Phenanthrene	40.000	38.474	3.8	125	0.00
72	Anthracene	40.000	38.672	3.3	125	0.00
73	Carbazole	40.000	38.807	3.0	126	0.00
74	Di-n-butylphthalate	40.000	38.722	3.2	127	0.00
75 C	Fluoranthene	40.000	37.515	6.2	120	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	114	0.00
77	Benzydine	40.000	41.898	-4.7	129	0.00
78	Pyrene	40.000	41.596	-4.0	122	0.00
79 S	Terphenyl-d14	80.000	78.459	1.9	115	0.00
80	Butylbenzylphthalate	40.000	42.564	-6.4	124	0.00
81	Benzo(a)anthracene	40.000	39.397	1.5	114	0.00
82	3,3'-Dichlorobenzidine	40.000	41.054	-2.6	119	0.00
83	Chrysene	40.000	39.328	1.7	114	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.658	-1.6	122	0.00
85 c	Di-n-octyl phthalate	40.000	40.533	-1.3	120	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.065	9.8	96	0.00
88	Benzo(b)fluoranthene	40.000	41.402	-3.5	110	0.00
89	Benzo(k)fluoranthene	40.000	40.747	-1.9	104	0.00
90 C	Benzo(a)pyrene	40.000	40.079	-0.2	106	0.00
91	Dibenzo(a,h)anthracene	40.000	35.651	10.9	96	-0.01
92	Benzo(g,h,i)perylene	40.000	36.264	9.3	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022021\
Data File : BF123240.D
Acq On : 20 Feb 2021 10:26
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 22 03:14:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0