

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022323\
 Data File : BF132522.D
 Acq On : 23 Feb 2023 16:04
 Operator : CG\JU
 Sample : 01657-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-17-SB02

Quant Time: Feb 24 00:37:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 23 00:59:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.007	152	173425	20.000	ng	0.00
21) Naphthalene-d8	8.295	136	655671	20.000	ng	0.00
39) Acenaphthene-d10	10.054	164	361629	20.000	ng	0.00
64) Phenanthrene-d10	11.548	188	712694	20.000	ng	0.00
76) Chrysene-d12	14.207	240	397081	20.000	ng	-0.01
86) Perylene-d12	15.760	264	349128	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.643	112	1078163	100.898	ng	0.00
7) Phenol-d6	6.637	99	1393979	99.957	ng	0.00
23) Nitrobenzene-d5	7.572	82	926515	67.043	ng	0.00
42) 2,4,6-Tribromophenol	10.854	330	400674	102.594	ng	0.00
45) 2-Fluorobiphenyl	9.372	172	1489282	62.707	ng	0.00
79) Terphenyl-d14	13.142	244	1807292	76.953	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022323\
Data File : BF132522.D
Acq On : 23 Feb 2023 16:04
Operator : CG\JU
Sample : 01657-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-17-SB02

Quant Time: Feb 24 00:37:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 23 00:59:42 2023
Response via : Initial Calibration

