


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF022323\  
Data File : BF132515.D  
Acq On : 23 Feb 2023 12:04  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF022223.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Thu Feb 23 00:59:42 2023






AutoFind: Scans 1522, 1523, 1524; Background Corrected with Scan 1516

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 32.7         | 97955      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.9          | 1888       | PASS                |
| 69             | 198             | 0.00            | 100             | 33.6         | 100592     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.6          | 602        | PASS                |
| 127            | 198             | 10              | 80              | 45.6         | 136557     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.3          | 821        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 299477     | PASS                |
| 199            | 198             | 5               | 9               | 6.6          | 19730      | PASS                |
| 275            | 198             | 10              | 60              | 28.1         | 84200      | PASS                |
| 365            | 198             | 1               | 100             | 4.1          | 12138      | PASS                |
| 441            | 198             | 0.01            | 100             | 16.4         | 49192      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 318549     | PASS                |
| 443            | 442             | 15              | 24              | 20.1         | 64112      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF022323\  
Data File : BF132515.D  
Acq On : 23 Feb 2023 12:04  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Feb 24 03:29:02 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF022223.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Feb 23 00:59:42 2023  
Response via : Initial Calibration



TIC: BF132515.D\data.ms

**(70) Pentachlorophenol (C)**

**11.354min (-0.006) 446437.59 ng**




**response 316145**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 63.66  |
| 264.00 | 63.60  | 64.49  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF022323\  
Data File : BF132515.D  
Acq On : 23 Feb 2023 12:04  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Feb 24 03:29:02 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF022223.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Feb 23 00:59:42 2023  
Response via : Initial Calibration



TIC: BF132515.D\data.ms

(77) Benzidine

12.901min (+ 0.006) 1068509.59 ng

response 1716698

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 14.60  | 14.65  |
| 183.00 | 10.90  | 11.28  |
| 0.00   | 0.00   | 0.00   |

# DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 2/23/2023     | BNA_F            | <u>BF132515.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 849216           | 13.724             |
| DDD           | 19773            | 13.419             |
| DDE           | 1261             | 13.077             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 21034         | 870250           | 2.42               |
|               |                  |                    |

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP