

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030822\
 Data File : BF127253.D
 Acq On : 08 Mar 2022 12:36
 Operator : CG\JU
 Sample : N1763-09
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 3-4-SAMPLE-LOCATION-2

Quant Time: Mar 08 16:04:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 02:16:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	94701	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	388474	20.000	ng	# 0.00
39) Acenaphthene-d10	9.922	164	194500	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	270368	20.000	ng	# 0.00
76) Chrysene-d12	14.057	240	136677	20.000	ng	# 0.00
86) Perylene-d12	15.551	264	142560	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	589942	96.282	ng	0.00
7) Phenol-d6	6.510	99	765874	92.633	ng	0.00
23) Nitrobenzene-d5	7.439	82	532541	61.877	ng	-0.01
42) 2,4,6-Tribromophenol	10.716	330	205316	91.683	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	799190	60.490	ng	0.00
79) Terphenyl-d14	12.998	244	551721	59.341	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030822\
Data File : BF127253.D
Acq On : 08 Mar 2022 12:36
Operator : CG\JU
Sample : N1763-09
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
3-4-SAMPLE-LOCATION-2

Quant Time: Mar 08 16:04:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 05 02:16:58 2022
Response via : Initial Calibration

