

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030822\
 Data File : BF127250.D
 Acq On : 08 Mar 2022 10:52
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 08 16:02:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 02:16:58 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2	1,4-Dioxane	0.592	0.559	5.6	70	-0.02
3	Pyridine	1.553	1.811	-16.6	85	-0.01
4	n-Nitrosodimethylamine	0.761	0.831	-9.2	78	-0.01
5 S	2-Fluorophenol	1.294	1.637	-26.5#	93	0.00
6	Aniline	2.069	2.429	-17.4	84	0.00
7 S	Phenol-d6	1.746	2.081	-19.2	87	0.00
8	2-Chlorophenol	1.388	1.592	-14.7	84	0.00
9	Benzaldehyde	1.063	1.252	-17.8	95	0.00
10 C	Phenol	1.764	2.242	-27.1#	94	0.00
11	bis(2-Chloroethyl)ether	1.384	1.544	-11.6	82	0.00
12	1,3-Dichlorobenzene	1.498	1.603	-7.0	79	0.00
13 C	1,4-Dichlorobenzene	1.518	1.651	-8.8	81	0.00
14	1,2-Dichlorobenzene	1.448	1.544	-6.6	78	0.00
15	Benzyl Alcohol	1.471	1.436	2.4	70	0.00
16	2,2'-oxybis(1-Chloropropane	2.179	1.827	16.2	61	0.00
17	2-Methylphenol	1.201	1.239	-3.2	75	0.00
18	Hexachloroethane	0.582	0.611	-5.0	77	0.00
19 P	n-Nitroso-di-n-propylamine	1.125	1.079	4.1	70	0.00
20	3+4-Methylphenols	1.560	1.577	-1.1	73	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	67	0.00
22	Acetophenone	0.527	0.555	-5.3	74	0.00
23 S	Nitrobenzene-d5	0.443	0.470	-6.1	74	0.00
24	Nitrobenzene	0.419	0.435	-3.8	72	0.00
25	Isophorone	0.733	0.743	-1.4	71	0.00
26 C	2-Nitrophenol	0.178	0.206	-15.7	79	0.00
27	2,4-Dimethylphenol	0.293	0.289	1.4	70	0.00
28	bis(2-Chloroethoxy)methane	0.445	0.466	-4.7	74	0.00
29 C	2,4-Dichlorophenol	0.275	0.314	-14.2	79	0.00
30	1,2,4-Trichlorobenzene	0.308	0.351	-14.0	80	0.00
31	Naphthalene	1.044	1.129	-8.1	77	0.00
32	Benzoic acid	0.208	0.170	18.3	52	0.02
33	4-Chloroaniline	0.411	0.440	-7.1	75	0.00
34 C	Hexachlorobutadiene	0.180	0.215	-19.4	84	0.00
35	Caprolactam	0.096	0.101	-5.2	71	0.01
36 C	4-Chloro-3-methylphenol	0.352	0.387	-9.9	75	0.00
37	2-Methylnaphthalene	0.677	0.745	-10.0	77	0.00
38	1-Methylnaphthalene	0.649	0.736	-13.4	80	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	68	0.00
40	1,2,4,5-Tetrachlorobenzene	0.526	0.634	-20.5	87	0.00
41 P	Hexachlorocyclopentadiene	0.286	0.287	-0.3	68	0.00
42 S	2,4,6-Tribromophenol	0.230	0.289	-25.7#	87	0.00
43 C	2,4,6-Trichlorophenol	0.386	0.441	-14.2	80	0.00
44	2,4,5-Trichlorophenol	0.406	0.466	-14.8	80	0.00
45 S	2-Fluorobiphenyl	1.359	1.589	-16.9	85	0.00
46	1,1'-Biphenyl	1.585	1.760	-11.0	80	0.00
47	2-Chloronaphthalene	1.175	1.308	-11.3	79	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030822\
 Data File : BF127250.D
 Acq On : 08 Mar 2022 10:52
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 08 16:02:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 02:16:58 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.464	0.475	-2.4	72	0.00
49	Acenaphthylene	1.909	2.090	-9.5	78	0.00
50	Dimethylphthalate	1.430	1.593	-11.4	79	0.00
51	2,6-Dinitrotoluene	0.291	0.334	-14.8	80	0.00
52 C	Acenaphthene	1.241	1.227	1.1	70	0.00
53	3-Nitroaniline	0.356	0.387	-8.7	76	0.00
54 P	2,4-Dinitrophenol	0.154	0.172	-11.7	72	0.00
55	Dibenzofuran	1.707	1.889	-10.7	79	0.00
56 P	4-Nitrophenol	0.263	0.240	8.7	62	0.00
57	2,4-Dinitrotoluene	0.393	0.440	-12.0	77	0.00
58	Fluorene	1.330	1.470	-10.5	79	0.00
59	2,3,4,6-Tetrachlorophenol	0.322	0.365	-13.4	79	0.00
60	Diethylphthalate	1.494	1.633	-9.3	77	0.00
61	4-Chlorophenyl-phenylether	0.627	0.729	-16.3	83	0.00
62	4-Nitroaniline	0.351	0.385	-9.7	75	0.00
63	Azobenzene	1.663	1.676	-0.8	71	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	68	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.144	-22.0	79	0.00
66 c	n-Nitrosodiphenylamine	0.656	0.731	-11.4	79	0.00
67	4-Bromophenyl-phenylether	0.223	0.262	-17.5	83	0.00
68	Hexachlorobenzene	0.241	0.281	-16.6	84	0.00
69	Atrazine	0.203	0.231	-13.8	81	0.00
70 C	Pentachlorophenol	0.131	0.120	8.4	60	0.00
71	Phenanthrene	1.119	1.228	-9.7	79	0.00
72	Anthracene	1.114	1.225	-10.0	78	0.00
73	Carbazole	1.022	1.104	-8.0	77	0.00
74	Di-n-butylphthalate	1.314	1.428	-8.7	77	0.00
75 C	Fluoranthene	1.063	1.163	-9.4	78	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	70	0.00
77	Benzidine	0.544	0.579	-6.4	73	0.00
78	Pyrene	1.738	1.911	-10.0	78	0.00
79 S	Terphenyl-d14	1.361	1.591	-16.9	85	0.00
80	Butylbenzylphthalate	0.823	0.897	-9.0	76	0.00
81	Benzo(a)anthracene	1.348	1.471	-9.1	78	0.00
82	3,3'-Dichlorobenzidine	0.425	0.483	-13.6	80	0.00
83	Chrysene	1.268	1.380	-8.8	78	0.00
84	Bis(2-ethylhexyl)phthalate	1.080	1.137	-5.3	75	0.00
85 c	Di-n-octyl phthalate	1.555	1.599	-2.8	75	0.00
86 I	Perylene-d12	1.000	1.000	0.0	76	0.01
87	Indeno(1,2,3-cd)pyrene	1.311	1.464	-11.7	81	0.02
88	Benzo(b)fluoranthene	1.343	1.465	-9.1	85	0.00
89	Benzo(k)fluoranthene	1.296	1.327	-2.4	78	0.01
90 C	Benzo(a)pyrene	1.049	1.136	-8.3	82	0.00
91	Dibenzo(a,h)anthracene	1.085	1.213	-11.8	82	0.02
92	Benzo(g,h,i)perylene	1.069	1.193	-11.6	80	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030822\
Data File : BF127250.D
Acq On : 08 Mar 2022 10:52
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 08 16:02:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 05 02:16:58 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1