



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030822\
 Data File : BF127249.D
 Acq On : 08 Mar 2022 09:54
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Mar 05 02:16:58 2022






AutoFind: Scans 1592, 1593, 1594; Background Corrected with Scan 1587

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.0	55716	PASS
68	69	0.00	2	1.9	1067	PASS
69	198	0.00	100	39.6	56555	PASS
70	69	0.00	2	0.5	280	PASS
127	198	10	80	53.1	75739	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	142696	PASS
199	198	5	9	6.5	9303	PASS
275	198	10	60	31.1	44352	PASS
365	198	1	100	4.7	6644	PASS
441	198	0.01	100	18.0	25622	PASS
442	442	50	100	100.0	160232	PASS
443	442	15	24	19.6	31411	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030822\
 Data File : BF127249.D
 Acq On : 08 Mar 2022 09:54
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Mar 08 15:59:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 02:16:58 2022
 Response via : Initial Calibration



TIC: BF127249.D\data.ms

(70) Pentachlorophenol (C)

11.216min (-0.000) 43579.38 ng




response 101580

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.40	67.60
264.00	63.00	63.15
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030822\
Data File : BF127249.D
Acq On : 08 Mar 2022 09:54
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 08 15:59:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 05 02:16:58 2022
Response via : Initial Calibration



TIC: BF127249.D\data.ms

(77) Benzidine

12.757min (+ 0.006) 40274.76 ng

response 622767

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.80	14.50
183.00	10.20	12.11
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/8/2022	BNA_F	<u>BF127249.D</u>
Compound Name	Response	Retention Time
DDT	295064	13.57
DDD	3133	13.27
DDE	470	12.93
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3603	298667	1.21

Instrument :
BNA_F
ClientSampleId :
DFTPP