

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF030823\  
 Data File : BF132692.D  
 Acq On : 08 Mar 2023 16:11  
 Operator : CG\JU  
 Sample : 01825-02  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 HH-TP1-0306

Quant Time: Mar 09 00:37:42 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF022223.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Mar 07 13:42:31 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min)      |
|-----------------------------|--------|------|----------|---------|-------|---------------|
| -----                       |        |      |          |         |       |               |
| Internal Standards          |        |      |          |         |       |               |
| 1) 1,4-Dichlorobenzene-d4   | 6.987  | 152  | 170393   | 20.000  | ng    | 0.00          |
| 21) Naphthalene-d8          | 8.269  | 136  | 624650   | 20.000  | ng    | -0.01         |
| 39) Acenaphthene-d10        | 10.033 | 164  | 327527   | 20.000  | ng    | 0.00          |
| 64) Phenanthrene-d10        | 11.527 | 188  | 609099   | 20.000  | ng    | 0.00          |
| 76) Chrysene-d12            | 14.186 | 240  | 362810   | 20.000  | ng    | -0.01         |
| 86) Perylene-d12            | 15.733 | 264  | 406842   | 20.000  | ng    | 0.00          |
| System Monitoring Compounds |        |      |          |         |       |               |
| 5) 2-Fluorophenol           | 5.628  | 112  | 1105216  | 105.270 | ng    | 0.01          |
| 7) Phenol-d6                | 6.616  | 99   | 1314078  | 95.905  | ng    | 0.00          |
| 23) Nitrobenzene-d5         | 7.551  | 82   | 982915   | 74.656  | ng    | -0.01         |
| 42) 2,4,6-Tribromophenol    | 10.827 | 330  | 402254   | 113.723 | ng    | 0.00          |
| 45) 2-Fluorobiphenyl        | 9.351  | 172  | 1567607  | 72.878  | ng    | 0.00          |
| 79) Terphenyl-d14           | 13.121 | 244  | 1743212  | 81.236  | ng    | 0.00          |
| Target Compounds            |        |      |          |         |       |               |
| 77) Benzdine                | 12.874 | 184  | 904741   | 102.441 | ng    | Qvalue<br>100 |
| -----                       |        |      |          |         |       |               |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF030823\  
Data File : BF132692.D  
Acq On : 08 Mar 2023 16:11  
Operator : CG\JU  
Sample : 01825-02  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
HH-TP1-0306

Quant Time: Mar 09 00:37:42 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF022223.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Mar 07 13:42:31 2023  
Response via : Initial Calibration

