

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031122\
 Data File : BF127295.D
 Acq On : 11 Mar 2022 10:54
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 11 12:37:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 16:51:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	116	0.00
2	1,4-Dioxane	40.000	41.490	-3.7	127	0.01
3	Pyridine	40.000	40.689	-1.7	117	0.00
4	n-Nitrosodimethylamine	40.000	40.654	-1.6	115	0.01
5 S	2-Fluorophenol	80.000	72.155	9.8	109	0.00
6	Aniline	40.000	39.549	1.1	113	0.00
7 S	Phenol-d6	80.000	76.078	4.9	112	0.00
8	2-Chlorophenol	40.000	37.174	7.1	108	0.00
9	Benzaldehyde	40.000	42.483	-6.2	120	0.00
10 C	Phenol	40.000	38.748	3.1	112	0.00
11	bis(2-Chloroethyl)ether	40.000	38.143	4.6	108	0.00
12	1,3-Dichlorobenzene	40.000	39.863	0.3	117	0.00
13 C	1,4-Dichlorobenzene	40.000	38.631	3.4	112	0.00
14	1,2-Dichlorobenzene	40.000	37.299	6.8	108	0.00
15	Benzyl Alcohol	40.000	39.772	0.6	111	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.789	5.5	107	0.00
17	2-Methylphenol	40.000	38.070	4.8	111	0.00
18	Hexachloroethane	40.000	37.055	7.4	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.460	6.3	106	0.00
20	3+4-Methylphenols	40.000	35.891	10.3	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	114	0.00
22	Acetophenone	40.000	37.583	6.0	107	0.00
23 S	Nitrobenzene-d5	80.000	77.953	2.6	111	0.00
24	Nitrobenzene	40.000	39.305	1.7	111	0.00
25	Isophorone	40.000	40.862	-2.2	118	0.00
26 C	2-Nitrophenol	40.000	40.427	-1.1	116	0.00
27	2,4-Dimethylphenol	40.000	41.627	-4.1	119	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.107	-2.8	114	0.00
29 C	2,4-Dichlorophenol	40.000	40.677	-1.7	113	0.00
30	1,2,4-Trichlorobenzene	40.000	40.409	-1.0	115	0.00
31	Naphthalene	40.000	39.453	1.4	115	0.00
32	Benzoic acid	40.000	38.117	4.7	114	0.00
33	4-Chloroaniline	40.000	40.503	-1.3	117	0.00
34 C	Hexachlorobutadiene	40.000	40.491	-1.2	113	0.00
35	Caprolactam	40.000	40.012	-0.0	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.936	-2.3	114	0.00
37	2-Methylnaphthalene	40.000	40.173	-0.4	115	0.00
38	1-Methylnaphthalene	40.000	39.541	1.1	111	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	113	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.679	0.8	113	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.427	-1.1	111	0.00
42 S	2,4,6-Tribromophenol	80.000	79.501	0.6	111	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.789	-2.0	112	0.00
44	2,4,5-Trichlorophenol	40.000	42.463	-6.2	119	0.00
45 S	2-Fluorobiphenyl	80.000	78.631	1.7	115	0.00
46	1,1'-Biphenyl	40.000	39.053	2.4	113	0.00
47	2-Chloronaphthalene	40.000	39.004	2.5	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031122\
 Data File : BF127295.D
 Acq On : 11 Mar 2022 10:54
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 11 12:37:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 16:51:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.981	0.0	112	0.00
49	Acenaphthylene	40.000	38.974	2.6	109	0.00
50	Dimethylphthalate	40.000	38.341	4.1	112	0.00
51	2,6-Dinitrotoluene	40.000	38.902	2.7	108	0.00
52 C	Acenaphthene	40.000	38.524	3.7	111	0.00
53	3-Nitroaniline	40.000	40.588	-1.5	115	0.00
54 P	2,4-Dinitrophenol	40.000	38.406	4.0	104	0.00
55	Dibenzofuran	40.000	38.650	3.4	110	0.00
56 P	4-Nitrophenol	40.000	40.742	-1.9	105	0.00
57	2,4-Dinitrotoluene	40.000	40.404	-1.0	111	0.00
58	Fluorene	40.000	38.307	4.2	109	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.036	-2.6	114	0.00
60	Diethylphthalate	40.000	39.416	1.5	113	0.00
61	4-Chlorophenyl-phenylether	40.000	38.700	3.2	108	0.00
62	4-Nitroaniline	40.000	38.742	3.1	108	0.00
63	Azobenzene	40.000	39.408	1.5	108	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	109	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.040	-10.1	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.485	-1.2	111	0.00
67	4-Bromophenyl-phenylether	40.000	40.753	-1.9	108	0.00
68	Hexachlorobenzene	40.000	39.951	0.1	104	0.00
69	Atrazine	40.000	43.652	-9.1	112	0.00
70 C	Pentachlorophenol	40.000	45.085	-12.7	111	0.00
71	Phenanthrene	40.000	38.441	3.9	107	0.00
72	Anthracene	40.000	38.648	3.4	107	0.00
73	Carbazole	40.000	38.328	4.2	105	0.00
74	Di-n-butylphthalate	40.000	41.657	-4.1	114	0.00
75 C	Fluoranthene	40.000	40.214	-0.5	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	117	0.00
77	Benzydine	40.000	39.338	1.7	135	0.00
78	Pyrene	40.000	40.958	-2.4	121	0.00
79 S	Terphenyl-d14	80.000	76.459	4.4	114	0.00
80	Butylbenzylphthalate	40.000	40.181	-0.5	115	0.00
81	Benzo(a)anthracene	40.000	39.786	0.5	115	0.00
82	3,3'-Dichlorobenzidine	40.000	40.077	-0.2	121	0.00
83	Chrysene	40.000	39.488	1.3	118	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.789	0.5	113	0.00
85 c	Di-n-octyl phthalate	40.000	41.325	-3.3	121	0.00
86 I	Perylene-d12	20.000	20.000	0.0	125	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.433	6.4	114	0.00
88	Benzo(b)fluoranthene	40.000	39.870	0.3	124	0.00
89	Benzo(k)fluoranthene	40.000	43.274	-8.2	138	0.00
90 C	Benzo(a)pyrene	40.000	40.162	-0.4	122	0.00
91	Dibenzo(a,h)anthracene	40.000	37.679	5.8	114	0.00
92	Benzo(g,h,i)perylene	40.000	36.693	8.3	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031122\
Data File : BF127295.D
Acq On : 11 Mar 2022 10:54
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 11 12:37:16 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 10 16:51:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0