

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031622\
 Data File : BF127359.D
 Acq On : 16 Mar 2022 20:57
 Operator : CG\JU
 Sample : N1951-07
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DB-13-(10-15)-3-12-22

Integration Parameters: rteint.p

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M

Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC: BF127359.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	5.110	477	480	490	rBV	20981	29989	1.06%	0.165%
2	5.492	541	545	548	rBV	2713146	2513622	88.83%	13.834%
3	6.498	712	716	728	rBV	2367422	2672935	94.45%	14.711%
4	6.863	775	778	782	rBV	638323	508208	17.96%	2.797%
5	7.434	870	875	878	rBV	1942904	1751790	61.90%	9.641%
6	8.145	993	996	1000	rBV	728622	653558	23.10%	3.597%
7	9.228	1175	1180	1183	rBV	2952438	2829857	100.00%	15.574%
8	9.904	1292	1295	1304	rVB	981415	763711	26.99%	4.203%
9	10.698	1426	1430	1444	rBV	1973763	1910788	67.52%	10.516%
10	11.398	1545	1549	1552	rBV	959554	800842	28.30%	4.407%
11	12.986	1814	1819	1822	rBV	2945245	2663317	94.11%	14.658%
12	14.045	1995	1999	2005	rVB	586059	560473	19.81%	3.085%
13	15.527	2246	2251	2260	rBV	396369	511147	18.06%	2.813%

Sum of corrected areas: 18170237

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031622\
Data File : BF127359.D
Acq On : 16 Mar 2022 20:57
Operator : CG\JU
Sample : N1951-07
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DB-13-(10-15)-3-12-22

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031622\
Data File : BF127359.D
Acq On : 16 Mar 2022 20:57
Operator : CG\JU
Sample : N1951-07
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DB-13-(10-15)-3-12-22

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc
------------------	----	---------	-------	----------	---	----	------	------