

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031821\
 Data File : BF123379.D
 Acq On : 18 Mar 2021 11:17
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 12:34:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF031721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 18 02:38:45 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	95	0.00
2	1,4-Dioxane	40.000	40.480	-1.2	100	0.00
3	Pyridine	40.000	40.243	-0.6	97	-0.01
4	n-Nitrosodimethylamine	40.000	39.662	0.8	94	-0.01
5 S	2-Fluorophenol	80.000	74.663	6.7	98	0.00
6	Aniline	40.000	35.480	11.3	98	0.00
7 S	Phenol-d6	80.000	69.491	13.1	97	0.00
8	2-Chlorophenol	40.000	36.655	8.4	95	0.00
9	Benzaldehyde	40.000	40.211	-0.5	99	0.00
10 C	Phenol	40.000	40.347	-0.9	98	0.00
11	bis(2-Chloroethyl)ether	40.000	38.682	3.3	95	0.00
12	1,3-Dichlorobenzene	40.000	38.069	4.8	95	0.00
13 C	1,4-Dichlorobenzene	40.000	36.233	9.4	95	0.00
14	1,2-Dichlorobenzene	40.000	34.361	14.1	96	0.00
15	Benzyl Alcohol	40.000	34.408	14.0	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.175	2.1	97	0.00
17	2-Methylphenol	40.000	37.083	7.3	96	0.00
18	Hexachloroethane	40.000	38.784	3.0	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	33.827	15.4	95	0.00
20	3+4-Methylphenols	40.000	39.921	0.2	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	33.979	15.1	94	0.00
23 S	Nitrobenzene-d5	80.000	77.015	3.7	95	0.00
24	Nitrobenzene	40.000	38.432	3.9	95	0.00
25	Isophorone	40.000	38.852	2.9	95	0.00
26 C	2-Nitrophenol	40.000	39.306	1.7	94	0.00
27	2,4-Dimethylphenol	40.000	38.687	3.3	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.898	2.8	96	0.00
29 C	2,4-Dichlorophenol	40.000	38.904	2.7	95	0.00
30	1,2,4-Trichlorobenzene	40.000	37.719	5.7	95	0.00
31	Naphthalene	40.000	36.256	9.4	96	0.00
32	Benzoic acid	40.000	41.835	-4.6	93	0.00
33	4-Chloroaniline	40.000	39.248	1.9	95	0.00
34 C	Hexachlorobutadiene	40.000	37.854	5.4	96	0.00
35	Caprolactam	40.000	40.596	-1.5	94	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.011	2.5	95	0.00
37	2-Methylnaphthalene	40.000	36.070	9.8	95	0.00
38	1-Methylnaphthalene	40.000	35.944	10.1	94	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.529	6.2	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.134	2.2	87	0.00
42 S	2,4,6-Tribromophenol	80.000	76.832	4.0	93	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.819	3.0	94	0.00
44	2,4,5-Trichlorophenol	40.000	38.092	4.8	91	0.00
45 S	2-Fluorobiphenyl	80.000	84.539	-5.7	95	0.00
46	1,1'-Biphenyl	40.000	34.244	14.4	95	0.00
47	2-Chloronaphthalene	40.000	36.382	9.0	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031821\
 Data File : BF123379.D
 Acq On : 18 Mar 2021 11:17
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 12:34:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF031721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 18 02:38:45 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.800	3.0	92	0.00
49	Acenaphthylene	40.000	35.944	10.1	94	0.00
50	Dimethylphthalate	40.000	38.101	4.7	94	0.00
51	2,6-Dinitrotoluene	40.000	39.261	1.8	93	0.00
52 C	Acenaphthene	40.000	35.391	11.5	93	0.00
53	3-Nitroaniline	40.000	39.568	1.1	94	0.00
54 P	2,4-Dinitrophenol	40.000	40.967	-2.4	88	0.00
55	Dibenzofuran	40.000	38.900	2.8	94	0.00
56 P	4-Nitrophenol	40.000	42.521	-6.3	90	0.00
57	2,4-Dinitrotoluene	40.000	39.514	1.2	94	0.00
58	Fluorene	40.000	38.574	3.6	94	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.530	1.2	93	0.00
60	Diethylphthalate	40.000	36.409	9.0	94	0.00
61	4-Chlorophenyl-phenylether	40.000	39.139	2.2	94	0.00
62	4-Nitroaniline	40.000	40.456	-1.1	94	0.00
63	Azobenzene	40.000	36.870	7.8	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.867	-2.2	92	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.510	8.7	94	0.00
67	4-Bromophenyl-phenylether	40.000	37.650	5.9	95	0.00
68	Hexachlorobenzene	40.000	37.486	6.3	95	0.00
69	Atrazine	40.000	38.582	3.5	95	0.00
70 C	Pentachlorophenol	40.000	41.203	-3.0	91	0.00
71	Phenanthrene	40.000	33.600	16.0	96	0.00
72	Anthracene	40.000	33.747	15.6	96	0.00
73	Carbazole	40.000	34.195	14.5	94	0.00
74	Di-n-butylphthalate	40.000	35.054	12.4	98	0.00
75 C	Fluoranthene	40.000	40.371	-0.9	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	41.038	-2.6	110	0.00
78	Pyrene	40.000	36.077	9.8	98	0.00
79 S	Terphenyl-d14	80.000	65.652	17.9	97	0.00
80	Butylbenzylphthalate	40.000	38.602	3.5	100	0.00
81	Benzo(a)anthracene	40.000	36.465	8.8	100	0.00
82	3,3'-Dichlorobenzidine	40.000	38.523	3.7	101	0.00
83	Chrysene	40.000	36.711	8.2	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.574	6.1	100	0.00
85 c	Di-n-octyl phthalate	40.000	39.801	0.5	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	34.602	13.5	101	0.00
88	Benzo(b)fluoranthene	40.000	37.329	6.7	99	0.00
89	Benzo(k)fluoranthene	40.000	39.464	1.3	113	0.00
90 C	Benzo(a)pyrene	40.000	37.945	5.1	106	0.00
91	Dibenzo(a,h)anthracene	40.000	34.654	13.4	102	0.00
92	Benzo(g,h,i)perylene	40.000	34.272	14.3	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031821\
Data File : BF123379.D
Acq On : 18 Mar 2021 11:17
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 18 12:34:35 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF031721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 18 02:38:45 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0