

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040722\
 Data File : BF127687.D
 Acq On : 07 Apr 2022 17:03
 Operator : CG\JU
 Sample : N2258-09
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TW-E5-DUP

Quant Time: Apr 07 17:23:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 06 13:55:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.963	152	115803	20.000	ng	0.00
21) Naphthalene-d8	8.245	136	413896	20.000	ng	# 0.00
39) Acenaphthene-d10	10.010	164	243382	20.000	ng	0.00
64) Phenanthrene-d10	11.498	188	456543	20.000	ng	0.00
76) Chrysene-d12	14.145	240	353242	20.000	ng	-0.01
86) Perylene-d12	15.674	264	306792	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.575	112	464861	65.228	ng	0.00
7) Phenol-d6	6.569	99	385984	41.400	ng	-0.01
23) Nitrobenzene-d5	7.528	82	785963	84.657	ng	0.00
42) 2,4,6-Tribromophenol	10.798	330	359632	138.857	ng	0.00
45) 2-Fluorobiphenyl	9.327	172	1366679	79.439	ng	0.00
79) Terphenyl-d14	13.092	244	1622674	75.823	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040722\
Data File : BF127687.D
Acq On : 07 Apr 2022 17:03
Operator : CG\JU
Sample : N2258-09
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TW-E5-DUP

Quant Time: Apr 07 17:23:04 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 06 13:55:27 2022
Response via : Initial Calibration

