




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040722\
Data File : BF127673.D
Acq On : 07 Apr 2022 09:40
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 07 17:05:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 06 13:55:27 2022
Response via : Initial Calibration






TIC: BF127673.D\data.ms

(70) Pentachlorophenol (C)**11.298min (-0.000) 392502.53 ng****response 261246**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	63.22
264.00	60.60	62.24
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF040722\
Data File : BF127673.D
Acq On : 07 Apr 2022 09:40
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 07 17:05:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF040622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 06 13:55:27 2022
Response via : Initial Calibration



TIC: BF127673.D\data.ms

(77) Benzidine

12.851min (+ 0.012) 5065.52 ng m

response 1175114

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	14.80
183.00	11.70	11.85
0.00	0.00	0.00