

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041621\
 Data File : BF123595.D
 Acq On : 16 Apr 2021 16:48
 Operator : JU/CG
 Sample : PB135774BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB135774BL

Quant Time: Apr 16 17:35:18 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 10:31:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.869	152	247044	20.00	ng	0.00
21) Naphthalene-d8	8.151	136	949824	20.00	ng	0.00
39) Acenaphthene-d10	9.910	164	468164	20.00	ng	0.00
64) Phenanthrene-d10	11.404	188	904239	20.00	ng	0.00
76) Chrysene-d12	14.051	240	673817	20.00	ng	0.00
86) Perylene-d12	15.539	264	524182	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.492	112	1972216	129.03	ng	0.01
7) Phenol-d6	6.498	99	2481118	117.14	ng	0.00
23) Nitrobenzene-d5	7.433	82	1808396	87.53	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	751041	141.67	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	2809718	89.47	ng	0.00
79) Terphenyl-d14	12.998	244	3103307	89.51	ng	0.00

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF041621\
Data File : BF123595.D
Acq On : 16 Apr 2021 16:48
Operator : JU/CG
Sample : PB135774BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB135774BL

Quant Time: Apr 16 17:35:18 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF041521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 16 10:31:33 2021
Response via : Initial Calibration

