

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042623\
 Data File : BF133066.D
 Acq On : 26 Apr 2023 15:54
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 27 01:56:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 27 01:56:03 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.614	0.568	7.5	95	0.00
3	Pyridine	1.631	1.536	5.8	94	0.00
4	n-Nitrosodimethylamine	0.845	0.716	15.3	87	0.00
5 S	2-Fluorophenol	1.324	1.254	5.3	97	0.00
6	Aniline	2.117	1.973	6.8	92	0.00
7 S	Phenol-d6	1.643	1.522	7.4	96	0.00
8	2-Chlorophenol	1.344	1.247	7.2	94	0.00
9	Benzaldehyde	0.791	0.778	1.6	105	0.00
10 C	Phenol	1.813	1.617	10.8	91	0.00
11	bis(2-Chloroethyl)ether	1.360	1.231	9.5	92	0.00
12	1,3-Dichlorobenzene	1.429	1.324	7.3	95	0.00
13 C	1,4-Dichlorobenzene	1.432	1.319	7.9	93	0.00
14	1,2-Dichlorobenzene	1.321	1.240	6.1	96	0.00
15	Benzyl Alcohol	1.135	1.015	10.6	88	0.00
16	2,2'-oxybis(1-Chloropropane	2.363	1.994	15.6	86	0.00
17	2-Methylphenol	1.231	1.153	6.3	95	0.00
18	Hexachloroethane	0.498	0.459	7.8	94	0.00
19 P	n-Nitroso-di-n-propylamine	1.023	0.842	17.7	86	0.00
20	3+4-Methylphenols	1.450	1.324	8.7	94	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.463	0.420	9.3	94	0.00
23 S	Nitrobenzene-d5	0.371	0.336	9.4	90	0.00
24	Nitrobenzene	0.375	0.338	9.9	90	0.00
25	Isophorone	0.704	0.619	12.1	89	0.00
26 C	2-Nitrophenol	0.181	0.181	0.0	96	0.00
27	2,4-Dimethylphenol	0.311	0.291	6.4	93	0.00
28	bis(2-Chloroethoxy)methane	0.416	0.374	10.1	90	0.00
29 C	2,4-Dichlorophenol	0.283	0.274	3.2	96	0.00
30	1,2,4-Trichlorobenzene	0.293	0.280	4.4	97	0.00
31	Naphthalene	1.000	0.933	6.7	94	0.00
32	Benzoic acid	0.192	0.228	-18.8	110	0.00
33	4-Chloroaniline	0.404	0.405	-0.2	102	0.00
34 C	Hexachlorobutadiene	0.162	0.153	5.6	95	0.00
35	Caprolactam	0.095	0.097	-2.1	98	0.00
36 C	4-Chloro-3-methylphenol	0.305	0.292	4.3	95	0.00
37	2-Methylnaphthalene	0.684	0.638	6.7	94	0.00
38	1-Methylnaphthalene	0.640	0.590	7.8	94	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	0.538	0.498	7.4	94	0.00
41 P	Hexachlorocyclopentadiene	0.314	0.297	5.4	91	0.00
42 S	2,4,6-Tribromophenol	0.201	0.200	0.5	98	0.00
43 C	2,4,6-Trichlorophenol	0.372	0.352	5.4	95	0.00
44	2,4,5-Trichlorophenol	0.430	0.416	3.3	96	0.00
45 S	2-Fluorobiphenyl	1.195	1.146	4.1	97	0.00
46	1,1'-Biphenyl	1.586	1.465	7.6	94	0.00
47	2-Chloronaphthalene	1.161	1.077	7.2	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042623\
 Data File : BF133066.D
 Acq On : 26 Apr 2023 15:54
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 27 01:56:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 27 01:56:03 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.384	0.363	5.5	92	0.00
49	Acenaphthylene	1.855	1.755	5.4	95	0.00
50	Dimethylphthalate	1.395	1.317	5.6	96	0.00
51	2,6-Dinitrotoluene	0.314	0.306	2.5	98	0.00
52 C	Acenaphthene	1.242	1.179	5.1	96	0.00
53	3-Nitroaniline	0.373	0.369	1.1	96	0.00
54 P	2,4-Dinitrophenol	0.158	0.173	-9.5	102	0.00
55	Dibenzofuran	1.695	1.583	6.6	95	0.00
56 P	4-Nitrophenol	0.289	0.287	0.7	95	0.00
57	2,4-Dinitrotoluene	0.400	0.414	-3.5	100	0.00
58	Fluorene	1.241	1.169	5.8	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.333	0.324	2.7	98	0.00
60	Diethylphthalate	1.317	1.251	5.0	96	0.00
61	4-Chlorophenyl-phenylether	0.605	0.566	6.4	97	0.00
62	4-Nitroaniline	0.343	0.375	-9.3	111	0.00
63	Azobenzene	1.395	1.250	10.4	91	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	0.121	0.133	-9.9	105	0.00
66 c	n-Nitrosodiphenylamine	0.649	0.600	7.6	98	0.00
67	4-Bromophenyl-phenylether	0.217	0.203	6.5	98	0.00
68	Hexachlorobenzene	0.222	0.210	5.4	98	0.00
69	Atrazine	0.187	0.184	1.6	100	0.00
70 C	Pentachlorophenol	0.158	0.154	2.5	96	0.00
71	Phenanthrene	1.075	0.988	8.1	98	0.00
72	Anthracene	1.100	1.025	6.8	99	0.00
73	Carbazole	0.980	0.933	4.8	100	0.00
74	Di-n-butylphthalate	1.109	1.102	0.6	101	0.00
75 C	Fluoranthene	1.079	1.042	3.4	101	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
77	Benzydine	0.338	0.455	-34.6#	150	0.00
78	Pyrene	1.714	1.580	7.8	101	0.00
79 S	Terphenyl-d14	1.160	1.052	9.3	101	0.00
80	Butylbenzylphthalate	0.607	0.659	-8.6	111	0.00
81	Benzo(a)anthracene	1.375	1.251	9.0	99	0.00
82	3,3'-Dichlorobenzidine	0.380	0.409	-7.6	112	0.00
83	Chrysene	1.375	1.260	8.4	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.762	0.724	5.0	99	0.00
85 c	Di-n-octyl phthalate	1.242	1.175	5.4	98	0.00
86 I	Perylene-d12	1.000	1.000	0.0	109	0.00
87	Indeno(1,2,3-cd)pyrene	1.138	1.208	-6.2	109	0.00
88	Benzo(b)fluoranthene	1.272	1.204	5.3	98	0.00
89	Benzo(k)fluoranthene	1.261	1.187	5.9	100	0.00
90 C	Benzo(a)pyrene	1.158	1.121	3.2	101	0.00
91	Dibenzo(a,h)anthracene	0.949	1.010	-6.4	107	0.00
92	Benzo(g,h,i)perylene	0.937	1.036	-10.6	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042623\
Data File : BF133066.D
Acq On : 26 Apr 2023 15:54
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 27 01:56:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 27 01:56:03 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0