




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051122\
 Data File : BF128190.D
 Acq On : 11 May 2022 12:55
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 12 09:44:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 09 15:30:11 2022
 Response via : Initial Calibration



TIC: BF128190.D\data.ms

(70) Pentachlorophenol (C)




11.227min (-0.006) 38769.75 ng

response 102305

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.30	65.18
264.00	66.40	60.60
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051122\
Data File : BF128190.D
Acq On : 11 May 2022 12:55
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 12 09:44:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 09 15:30:11 2022
Response via : Initial Calibration



TIC: BF128190.D\data.ms

(77) Benzidine

12.769min (-0.000) 0.00 ng

response 659675

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.10	14.37
183.00	11.20	11.69
0.00	0.00	0.00