


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051122\
 Data File : BF128184.D
 Acq On : 11 May 2022 09:40
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon May 09 15:30:11 2022






AutoFind: Scans 1594, 1595, 1596; Background Corrected with Scan 1588

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.0	50142	PASS
68	69	0.00	2	1.6	797	PASS
69	198	0.00	100	38.0	48923	PASS
70	69	0.00	2	0.4	185	PASS
127	198	10	80	46.5	59795	PASS
197	198	0.00	2	0.6	804	PASS
198	198	100	100	100.0	128616	PASS
199	198	5	9	7.1	9134	PASS
275	198	10	60	29.2	37579	PASS
365	198	1	100	4.6	5966	PASS
441	198	0.01	100	14.8	19062	PASS
442	442	50	100	100.0	121683	PASS
443	442	15	24	19.0	23093	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051122\
Data File : BF128184.D
Acq On : 11 May 2022 09:40
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 12 09:42:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 09 15:30:11 2022
Response via : Initial Calibration



TIC: BF128184.D\data.ms

(70) Pentachlorophenol (C)

11.227min (-0.006) 300514.24 ng




response 102986

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.30	66.59
264.00	66.40	62.11
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051122\
Data File : BF128184.D
Acq On : 11 May 2022 09:40
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 12 09:42:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 09 15:30:11 2022
Response via : Initial Calibration



TIC: BF128184.D\data.ms

(77) Benzidine

12.769min (-0.000) 217070.12 ng

response 644088

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.10	14.68
183.00	11.20	11.83
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/11/2022	BNA_F	<u>BF128184.D</u>
Compound Name	Response	Retention Time
DDT	329131	13.586
DDD	5175	13.286
DDE	577	12.951
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5752	334883	1.72

Instrument :
BNA_F
ClientSampleId :
DFTPP