

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051822\
 Data File : BF128307.D
 Acq On : 18 May 2022 12:31
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 18 16:48:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 18 01:31:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.558	0.567	-1.6	100	0.00
3	Pyridine	1.481	1.512	-2.1	102	0.00
4	n-Nitrosodimethylamine	0.738	0.798	-8.1	105	0.00
5 S	2-Fluorophenol	1.255	1.246	0.7	101	0.00
6	Aniline	1.874	1.884	-0.5	100	0.00
7 S	Phenol-d6	1.659	1.640	1.1	100	0.00
8	2-Chlorophenol	1.278	1.273	0.4	100	0.00
9	Benzaldehyde	0.958	0.935	2.4	114	0.00
10 C	Phenol	1.750	1.737	0.7	98	0.00
11	bis(2-Chloroethyl)ether	1.311	1.353	-3.2	103	0.00
12	1,3-Dichlorobenzene	1.438	1.448	-0.7	101	0.00
13 C	1,4-Dichlorobenzene	1.453	1.465	-0.8	101	0.00
14	1,2-Dichlorobenzene	1.370	1.360	0.7	100	0.00
15	Benzyl Alcohol	1.311	1.378	-5.1	104	0.00
16	2,2'-oxybis(1-Chloropropane	1.943	2.064	-6.2	105	0.00
17	2-Methylphenol	1.084	1.099	-1.4	101	0.00
18	Hexachloroethane	0.549	0.570	-3.8	104	0.00
19 P	n-Nitroso-di-n-propylamine	1.022	1.056	-3.3	104	0.00
20	3+4-Methylphenols	1.342	1.340	0.1	101	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.498	0.490	1.6	102	0.00
23 S	Nitrobenzene-d5	0.444	0.455	-2.5	103	0.00
24	Nitrobenzene	0.410	0.428	-4.4	104	0.00
25	Isophorone	0.697	0.731	-4.9	105	0.00
26 C	2-Nitrophenol	0.180	0.187	-3.9	103	0.00
27	2,4-Dimethylphenol	0.278	0.266	4.3	97	0.00
28	bis(2-Chloroethoxy)methane	0.440	0.462	-5.0	106	0.00
29 C	2,4-Dichlorophenol	0.292	0.298	-2.1	103	0.00
30	1,2,4-Trichlorobenzene	0.335	0.339	-1.2	103	0.00
31	Naphthalene	0.988	0.994	-0.6	102	0.00
32	Benzoic acid	0.208	0.167	19.7	75	0.00
33	4-Chloroaniline	0.392	0.400	-2.0	105	0.00
34 C	Hexachlorobutadiene	0.225	0.228	-1.3	102	0.00
35	Caprolactam	0.095	0.094	1.1	98	0.00
36 C	4-Chloro-3-methylphenol	0.326	0.331	-1.5	102	0.00
37	2-Methylnaphthalene	0.666	0.666	0.0	102	0.00
38	1-Methylnaphthalene	0.640	0.641	-0.2	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.590	0.592	-0.3	103	0.00
41 P	Hexachlorocyclopentadiene	0.233	0.191	18.0	77	0.00
42 S	2,4,6-Tribromophenol	0.229	0.216	5.7	96	0.00
43 C	2,4,6-Trichlorophenol	0.380	0.381	-0.3	101	0.00
44	2,4,5-Trichlorophenol	0.405	0.403	0.5	100	0.00
45 S	2-Fluorobiphenyl	1.278	1.234	3.4	101	0.00
46	1,1'-Biphenyl	1.456	1.444	0.8	101	0.00
47	2-Chloronaphthalene	1.079	1.065	1.3	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051822\
 Data File : BF128307.D
 Acq On : 18 May 2022 12:31
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 18 16:48:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 18 01:31:11 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.387	0.408	-5.4	104	0.00
49	Acenaphthylene	1.632	1.598	2.1	99	0.00
50	Dimethylphthalate	1.291	1.295	-0.3	103	0.00
51	2,6-Dinitrotoluene	0.283	0.286	-1.1	102	0.00
52 C	Acenaphthene	1.132	1.099	2.9	99	0.00
53	3-Nitroaniline	0.307	0.300	2.3	99	0.00
54 P	2,4-Dinitrophenol	0.157	0.134	14.6	83	0.00
55	Dibenzofuran	1.607	1.550	3.5	100	0.00
56 P	4-Nitrophenol	0.214	0.186	13.1	85	0.00
57	2,4-Dinitrotoluene	0.379	0.360	5.0	95	0.00
58	Fluorene	1.252	1.205	3.8	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.346	0.326	5.8	94	0.00
60	Diethylphthalate	1.288	1.281	0.5	102	0.00
61	4-Chlorophenyl-phenylether	0.624	0.602	3.5	99	0.00
62	4-Nitroaniline	0.311	0.300	3.5	97	0.00
63	Azobenzene	1.410	1.447	-2.6	104	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	0.125	0.128	-2.4	91	0.00
66 c	n-Nitrosodiphenylamine	0.583	0.606	-3.9	99	0.00
67	4-Bromophenyl-phenylether	0.215	0.226	-5.1	100	0.00
68	Hexachlorobenzene	0.235	0.242	-3.0	99	0.00
69	Atrazine	0.197	0.202	-2.5	99	0.00
70 C	Pentachlorophenol	0.114	0.105	7.9	80	0.00
71	Phenanthrene	0.996	0.996	0.0	97	0.00
72	Anthracene	0.994	0.989	0.5	96	0.00
73	Carbazole	0.945	0.945	0.0	95	0.00
74	Di-n-butylphthalate	1.088	1.099	-1.0	97	0.00
75 C	Fluoranthene	1.080	1.020	5.6	92	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
77	Benzydine	0.433	0.498	-15.0	92	0.00
78	Pyrene	1.450	1.631	-12.5	91	0.00
79 S	Terphenyl-d14	1.066	1.176	-10.3	91	0.00
80	Butylbenzylphthalate	0.603	0.684	-13.4	90	0.00
81	Benzo(a)anthracene	1.277	1.318	-3.2	83	0.00
82	3,3'-Dichlorobenzidine	0.287	0.295	-2.8	85	0.00
83	Chrysene	1.219	1.207	1.0	81	0.00
84	Bis(2-ethylhexyl)phthalate	0.815	0.885	-8.6	86	0.00
85 c	Di-n-octyl phthalate	1.301	1.293	0.6	78	0.00
86 I	Perylene-d12	1.000	1.000	0.0	73	0.00
87	Indeno(1,2,3-cd)pyrene	1.210	1.302	-7.6	81	0.00
88	Benzo(b)fluoranthene	1.240	1.335	-7.7	78	0.00
89	Benzo(k)fluoranthene	1.224	1.185	3.2	69	0.00
90 C	Benzo(a)pyrene	1.001	1.016	-1.5	73	0.00
91	Dibenzo(a,h)anthracene	0.979	1.066	-8.9	81	0.00
92	Benzo(g,h,i)perylene	0.992	1.109	-11.8	85	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051822\
Data File : BF128307.D
Acq On : 18 May 2022 12:31
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 18 16:48:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 18 01:31:11 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0