

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF053023\
 Data File : BF133516.D
 Acq On : 30 May 2023 11:52
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 30 13:00:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 30 12:58:40 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.541	0.535	1.1	87	0.00
3	Pyridine	1.518	1.502	1.1	89	0.00
4	n-Nitrosodimethylamine	0.828	0.815	1.6	88	0.00
5 S	2-Fluorophenol	1.199	1.152	3.9	88	0.00
6	Aniline	2.004	1.975	1.4	90	0.00
7 S	Phenol-d6	1.476	1.432	3.0	89	0.00
8	2-Chlorophenol	1.181	1.150	2.6	87	0.00
9	Benzaldehyde	0.826	0.753	8.8	91	0.00
10 C	Phenol	1.480	1.466	0.9	88	0.00
11	bis(2-Chloroethyl)ether	1.224	1.206	1.5	90	0.00
12	1,3-Dichlorobenzene	1.311	1.285	2.0	90	0.00
13 C	1,4-Dichlorobenzene	1.315	1.295	1.5	90	0.00
14	1,2-Dichlorobenzene	1.213	1.193	1.6	90	0.00
15	Benzyl Alcohol	1.120	1.102	1.6	88	0.00
16	2,2'-oxybis(1-Chloropropane	2.321	2.225	4.1	87	0.00
17	2-Methylphenol	1.119	1.090	2.6	89	0.00
18	Hexachloroethane	0.455	0.441	3.1	85	0.00
19 P	n-Nitroso-di-n-propylamine	0.955	0.876	8.3	86	0.00
20	3+4-Methylphenols	1.265	1.285	-1.6	89	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
22	Acetophenone	0.461	0.439	4.8	89	0.00
23 S	Nitrobenzene-d5	0.300	0.325	-8.3	89	0.00
24	Nitrobenzene	0.319	0.341	-6.9	88	0.00
25	Isophorone	0.698	0.672	3.7	88	0.00
26 C	2-Nitrophenol	0.121	0.118	2.5	77	0.00
27	2,4-Dimethylphenol	0.297	0.292	1.7	89	0.00
28	bis(2-Chloroethoxy)methane	0.405	0.388	4.2	88	0.00
29 C	2,4-Dichlorophenol	0.267	0.269	-0.7	88	0.00
30	1,2,4-Trichlorobenzene	0.289	0.285	1.4	90	0.00
31	Naphthalene	0.970	0.956	1.4	91	0.00
32	Benzoic acid	0.172	0.144	16.3	72	0.00
33	4-Chloroaniline	0.416	0.412	1.0	91	0.00
34 C	Hexachlorobutadiene	0.176	0.173	1.7	89	0.00
35	Caprolactam	0.088	0.091	-3.4	90	0.00
36 C	4-Chloro-3-methylphenol	0.299	0.299	0.0	89	0.00
37	2-Methylnaphthalene	0.662	0.653	1.4	90	0.00
38	1-Methylnaphthalene	0.617	0.605	1.9	90	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
40	1,2,4,5-Tetrachlorobenzene	0.593	0.581	2.0	90	0.00
41 P	Hexachlorocyclopentadiene	0.304	0.291	4.3	81	0.00
42 S	2,4,6-Tribromophenol	0.199	0.204	-2.5	88	0.00
43 C	2,4,6-Trichlorophenol	0.379	0.377	0.5	86	0.00
44	2,4,5-Trichlorophenol	0.440	0.449	-2.0	87	0.00
45 S	2-Fluorobiphenyl	1.296	1.304	-0.6	91	0.00
46	1,1'-Biphenyl	1.625	1.598	1.7	91	0.00
47	2-Chloronaphthalene	1.178	1.152	2.2	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF053023\
 Data File : BF133516.D
 Acq On : 30 May 2023 11:52
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 30 13:00:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 30 12:58:40 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.320	0.350	-9.4	86	0.00
49	Acenaphthylene	2.010	1.991	0.9	90	0.00
50	Dimethylphthalate	1.406	1.379	1.9	89	0.00
51	2,6-Dinitrotoluene	0.227	0.253	-11.5	86	0.00
52 C	Acenaphthene	1.217	1.204	1.1	91	0.00
53	3-Nitroaniline	0.291	0.334	-14.8	92	0.00
54 P	2,4-Dinitrophenol	0.097	0.095	2.1	90	0.00
55	Dibenzofuran	1.811	1.803	0.4	91	0.00
56 P	4-Nitrophenol	0.240	0.250	-4.2	90	0.00
57	2,4-Dinitrotoluene	0.272	0.332	-22.1	95	0.00
58	Fluorene	1.327	1.306	1.6	92	0.00
59	2,3,4,6-Tetrachlorophenol	0.332	0.349	-5.1	89	0.00
60	Diethylphthalate	1.459	1.435	1.6	89	0.00
61	4-Chlorophenyl-phenylether	0.637	0.614	3.6	90	0.00
62	4-Nitroaniline	0.297	0.344	-15.8	93	0.00
63	Azobenzene	1.461	1.432	2.0	89	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	0.070	0.071	-1.4	93	0.00
66 c	n-Nitrosodiphenylamine	0.649	0.625	3.7	92	0.00
67	4-Bromophenyl-phenylether	0.212	0.206	2.8	89	0.00
68	Hexachlorobenzene	0.222	0.214	3.6	90	0.00
69	Atrazine	0.178	0.175	1.7	90	0.00
70 C	Pentachlorophenol	0.133	0.137	-3.0	87	0.00
71	Phenanthrene	1.013	0.984	2.9	92	0.00
72	Anthracene	1.043	1.019	2.3	93	0.00
73	Carbazole	0.986	0.970	1.6	93	0.00
74	Di-n-butylphthalate	1.166	1.123	3.7	88	0.00
75 C	Fluoranthene	1.119	1.115	0.4	94	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
77	Benzydine	0.585	0.468	20.0	84	0.00
78	Pyrene	1.702	1.640	3.6	92	0.00
79 S	Terphenyl-d14	1.125	1.048	6.8	92	0.00
80	Butylbenzylphthalate	0.673	0.651	3.3	85	0.00
81	Benzo(a)anthracene	1.332	1.295	2.8	93	0.00
82	3,3'-Dichlorobenzidine	0.416	0.417	-0.2	93	0.00
83	Chrysene	1.345	1.338	0.5	93	0.00
84	Bis(2-ethylhexyl)phthalate	0.842	0.810	3.8	86	0.00
85 c	Di-n-octyl phthalate	1.402	1.430	-2.0	84	0.00
86 I	Perylene-d12	1.000	1.000	0.0	88	0.00
87	Indeno(1,2,3-cd)pyrene	1.274	1.215	4.6	86	0.00
88	Benzo(b)fluoranthene	1.302	1.329	-2.1	92	0.00
89	Benzo(k)fluoranthene	1.270	1.240	2.4	86	0.00
90 C	Benzo(a)pyrene	1.185	1.159	2.2	87	0.00
91	Dibenzo(a,h)anthracene	1.033	0.985	4.6	86	0.00
92	Benzo(g,h,i)perylene	1.042	1.040	0.2	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF053023\
Data File : BF133516.D
Acq On : 30 May 2023 11:52
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 30 13:00:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 30 12:58:40 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0