

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060322\
 Data File : BF128684.D
 Acq On : 03 Jun 2022 17:39
 Operator : CG\JU
 Sample : N3099-05RE
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 20220426-SPLP-AMENDED-COMPOSITE-FRE

Quant Time: Jun 03 23:42:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 02 17:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.822	152	72625	20.000	ng	0.00
21) Naphthalene-d8	8.104	136	255997	20.000	ng	# 0.00
39) Acenaphthene-d10	9.863	164	123706	20.000	ng	0.00
64) Phenanthrene-d10	11.351	188	204711	20.000	ng	# 0.00
76) Chrysene-d12	13.992	240	198050	20.000	ng	# 0.00
86) Perylene-d12	15.456	264	181137	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	248175	54.430	ng	0.00
7) Phenol-d6	6.486	99	179197	32.197	ng	-0.01
23) Nitrobenzene-d5	7.386	82	462007	84.026	ng	0.00
42) 2,4,6-Tribromophenol	10.657	330	154422	105.649	ng	0.00
45) 2-Fluorobiphenyl	9.186	172	763324	86.436	ng	0.00
79) Terphenyl-d14	12.939	244	822340	72.150	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060322\
Data File : BF128684.D
Acq On : 03 Jun 2022 17:39
Operator : CG\JU
Sample : N3099-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
20220426-SPLP-AMENDED-COMPOSITE-FRE

Quant Time: Jun 03 23:42:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 02 17:07:23 2022
Response via : Initial Calibration

