

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060322\
 Data File : BF128670.D
 Acq On : 03 Jun 2022 10:14
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 03 23:40:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 02 17:07:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	39.908	0.2	96	0.00
3	Pyridine	40.000	39.134	2.2	95	0.00
4	n-Nitrosodimethylamine	40.000	41.191	-3.0	99	0.00
5 S	2-Fluorophenol	80.000	77.868	2.7	97	0.00
6	Aniline	40.000	39.066	2.3	99	0.00
7 S	Phenol-d6	80.000	79.285	0.9	100	0.00
8	2-Chlorophenol	40.000	39.029	2.4	99	0.00
9	Benzaldehyde	40.000	32.053	19.9	100	0.00
10 C	Phenol	40.000	39.689	0.8	99	0.00
11	bis(2-Chloroethyl)ether	40.000	39.754	0.6	100	0.00
12	1,3-Dichlorobenzene	40.000	39.800	0.5	100	0.00
13 C	1,4-Dichlorobenzene	40.000	39.670	0.8	101	0.00
14	1,2-Dichlorobenzene	40.000	40.160	-0.4	103	0.00
15	Benzyl Alcohol	40.000	41.438	-3.6	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.197	-0.5	105	0.00
17	2-Methylphenol	40.000	40.615	-1.5	105	-0.01
18	Hexachloroethane	40.000	41.347	-3.4	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.321	-3.3	104	0.00
20	3+4-Methylphenols	40.000	40.601	-1.5	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	40.245	-0.6	101	0.00
23 S	Nitrobenzene-d5	80.000	82.824	-3.5	105	0.00
24	Nitrobenzene	40.000	41.025	-2.6	103	0.00
25	Isophorone	40.000	40.464	-1.2	103	0.00
26 C	2-Nitrophenol	40.000	42.025	-5.1	102	0.00
27	2,4-Dimethylphenol	40.000	40.132	-0.3	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.572	-1.4	101	0.00
29 C	2,4-Dichlorophenol	40.000	40.437	-1.1	100	0.00
30	1,2,4-Trichlorobenzene	40.000	40.113	-0.3	100	0.00
31	Naphthalene	40.000	40.194	-0.5	105	0.00
32	Benzoic acid	40.000	43.648	-9.1	106	0.00
33	4-Chloroaniline	40.000	39.688	0.8	105	0.00
34 C	Hexachlorobutadiene	40.000	40.853	-2.1	108	0.00
35	Caprolactam	40.000	39.014	2.5	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.570	-1.4	106	0.00
37	2-Methylnaphthalene	40.000	40.274	-0.7	106	0.00
38	1-Methylnaphthalene	40.000	40.197	-0.5	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.938	-2.3	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.003	-2.5	104	0.00
42 S	2,4,6-Tribromophenol	80.000	81.368	-1.7	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.264	-0.7	106	0.00
44	2,4,5-Trichlorophenol	40.000	40.315	-0.8	106	0.00
45 S	2-Fluorobiphenyl	80.000	81.649	-2.1	108	0.00
46	1,1'-Biphenyl	40.000	39.760	0.6	106	0.00
47	2-Chloronaphthalene	40.000	39.902	0.2	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060322\
 Data File : BF128670.D
 Acq On : 03 Jun 2022 10:14
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 03 23:40:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 02 17:07:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.534	-6.3	111	0.00
49	Acenaphthylene	40.000	40.152	-0.4	106	0.00
50	Dimethylphthalate	40.000	40.262	-0.7	106	0.00
51	2,6-Dinitrotoluene	40.000	41.859	-4.6	107	0.00
52 C	Acenaphthene	40.000	40.461	-1.2	105	0.00
53	3-Nitroaniline	40.000	40.510	-1.3	104	0.00
54 P	2,4-Dinitrophenol	40.000	40.285	-0.7	102	0.00
55	Dibenzofuran	40.000	39.824	0.4	106	0.00
56 P	4-Nitrophenol	40.000	40.090	-0.2	108	-0.01
57	2,4-Dinitrotoluene	40.000	41.517	-3.8	105	0.00
58	Fluorene	40.000	40.256	-0.6	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.644	-1.6	105	0.00
60	Diethylphthalate	40.000	40.587	-1.5	107	0.00
61	4-Chlorophenyl-phenylether	40.000	40.005	-0.0	104	0.00
62	4-Nitroaniline	40.000	39.001	2.5	100	0.00
63	Azobenzene	40.000	41.067	-2.7	108	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.864	-7.2	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.854	-4.6	103	0.00
67	4-Bromophenyl-phenylether	40.000	42.377	-5.9	104	0.00
68	Hexachlorobenzene	40.000	42.070	-5.2	102	0.00
69	Atrazine	40.000	41.599	-4.0	101	0.00
70 C	Pentachlorophenol	40.000	42.531	-6.3	102	0.00
71	Phenanthrene	40.000	41.325	-3.3	103	0.00
72	Anthracene	40.000	41.248	-3.1	103	0.00
73	Carbazole	40.000	40.719	-1.8	102	0.00
74	Di-n-butylphthalate	40.000	42.403	-6.0	103	0.00
75 C	Fluoranthene	40.000	40.532	-1.3	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	87	0.00
77	Benzidine	40.000	42.778	-6.9	100	0.00
78	Pyrene	40.000	43.028	-7.6	97	0.00
79 S	Terphenyl-d14	80.000	87.258	-9.1	94	0.00
80	Butylbenzylphthalate	40.000	43.149	-7.9	91	0.00
81	Benzo(a)anthracene	40.000	41.131	-2.8	89	0.00
82	3,3'-Dichlorobenzidine	40.000	41.647	-4.1	91	0.00
83	Chrysene	40.000	40.893	-2.2	89	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.182	-10.5	94	0.00
85 c	Di-n-octyl phthalate	40.000	42.640	-6.6	90	0.00
86 I	Perylene-d12	20.000	20.000	0.0	90	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.339	-5.8	106	0.00
88	Benzo(b)fluoranthene	40.000	38.913	2.7	85	0.00
89	Benzo(k)fluoranthene	40.000	42.684	-6.7	93	0.00
90 C	Benzo(a)pyrene	40.000	40.479	-1.2	90	0.00
91	Dibenzo(a,h)anthracene	40.000	41.991	-5.0	102	0.00
92	Benzo(g,h,i)perylene	40.000	41.813	-4.5	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060322\
Data File : BF128670.D
Acq On : 03 Jun 2022 10:14
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 03 23:40:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 02 17:07:23 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0