

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062122\
 Data File : BF128901.D
 Acq On : 21 Jun 2022 17:02
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 22 01:51:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.466	0.495	-6.2	111	0.00
3	Pyridine	1.238	1.330	-7.4	114	0.00
4	n-Nitrosodimethylamine	0.859	0.827	3.7	101	0.00
5 S	2-Fluorophenol	1.256	1.294	-3.0	112	0.00
6	Aniline	1.653	1.708	-3.3	115	0.00
7 S	Phenol-d6	1.533	1.607	-4.8	115	0.00
8	2-Chlorophenol	1.285	1.374	-6.9	118	0.00
9	Benzaldehyde	0.934	1.052	-12.6	153#	0.00
10 C	Phenol	1.620	1.672	-3.2	112	0.00
11	bis(2-Chloroethyl)ether	1.187	1.270	-7.0	117	0.00
12	1,3-Dichlorobenzene	1.485	1.583	-6.6	117	0.00
13 C	1,4-Dichlorobenzene	1.500	1.593	-6.2	118	0.00
14	1,2-Dichlorobenzene	1.403	1.471	-4.8	118	0.00
15	Benzyl Alcohol	1.282	1.153	10.1	100	0.00
16	2,2'-oxybis(1-Chloropropane	1.956	1.987	-1.6	115	0.00
17	2-Methylphenol	1.015	1.447	-42.6#	160#	0.00
18	Hexachloroethane	0.558	0.595	-6.6	116	0.00
19 P	n-Nitroso-di-n-propylamine	0.953	1.007	-5.7	116	0.00
20	3+4-Methylphenols	1.356	1.447	-6.7	117	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
22	Acetophenone	0.523	0.566	-8.2	114	0.00
23 S	Nitrobenzene-d5	0.430	0.466	-8.4	116	0.00
24	Nitrobenzene	0.394	0.431	-9.4	116	0.00
25	Isophorone	0.652	0.700	-7.4	115	0.00
26 C	2-Nitrophenol	0.174	0.201	-15.5	118	0.00
27	2,4-Dimethylphenol	0.265	0.293	-10.6	114	0.00
28	bis(2-Chloroethoxy)methane	0.412	0.444	-7.8	113	0.00
29 C	2,4-Dichlorophenol	0.306	0.335	-9.5	114	0.00
30	1,2,4-Trichlorobenzene	0.345	0.375	-8.7	114	0.00
31	Naphthalene	1.026	1.110	-8.2	119	0.00
32	Benzoic acid	0.188	0.117	37.8#	63	0.00
33	4-Chloroaniline	0.387	0.406	-4.9	116	0.00
34 C	Hexachlorobutadiene	0.245	0.262	-6.9	118	0.00
35	Caprolactam	0.085	0.090	-5.9	117	0.00
36 C	4-Chloro-3-methylphenol	0.334	0.356	-6.6	117	0.00
37	2-Methylnaphthalene	0.653	0.704	-7.8	119	0.00
38	1-Methylnaphthalene	0.631	0.678	-7.4	121	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	0.621	0.730	-17.6	119	0.00
41 P	Hexachlorocyclopentadiene	0.373	0.412	-10.5	109	0.00
42 S	2,4,6-Tribromophenol	0.236	0.250	-5.9	104	0.00
43 C	2,4,6-Trichlorophenol	0.420	0.418	0.5	102	0.00
44	2,4,5-Trichlorophenol	0.434	0.453	-4.4	107	0.00
45 S	2-Fluorobiphenyl	1.428	1.613	-13.0	117	0.00
46	1,1'-Biphenyl	1.501	1.691	-12.7	117	0.00
47	2-Chloronaphthalene	1.145	1.312	-14.6	119	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062122\
 Data File : BF128901.D
 Acq On : 21 Jun 2022 17:02
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 22 01:51:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.392	0.445	-13.5	116	0.00
49	Acenaphthylene	1.753	1.907	-8.8	112	0.00
50	Dimethylphthalate	1.369	1.512	-10.4	113	0.00
51	2,6-Dinitrotoluene	0.270	0.318	-17.8	117	0.00
52 C	Acenaphthene	1.142	1.128	1.2	100	0.00
53	3-Nitroaniline	0.293	0.329	-12.3	112	0.00
54 P	2,4-Dinitrophenol	0.150	0.151	-0.7	98	0.00
55	Dibenzofuran	1.641	1.784	-8.7	112	0.00
56 P	4-Nitrophenol	0.213	0.199	6.6	98	0.00
57	2,4-Dinitrotoluene	0.361	0.422	-16.9	115	0.00
58	Fluorene	1.272	1.386	-9.0	110	0.00
59	2,3,4,6-Tetrachlorophenol	0.359	0.338	5.8	95	0.00
60	Diethylphthalate	1.376	1.489	-8.2	111	0.00
61	4-Chlorophenyl-phenylether	0.660	0.728	-10.3	112	0.00
62	4-Nitroaniline	0.297	0.328	-10.4	111	0.00
63	Azobenzene	1.355	1.460	-7.7	110	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.131	-11.0	111	0.00
66 c	n-Nitrosodiphenylamine	0.615	0.680	-10.6	112	0.00
67	4-Bromophenyl-phenylether	0.223	0.250	-12.1	114	0.00
68	Hexachlorobenzene	0.248	0.270	-8.9	110	0.00
69	Atrazine	0.198	0.206	-4.0	105	0.00
70 C	Pentachlorophenol	0.124	0.132	-6.5	106	0.00
71	Phenanthrene	1.003	1.091	-8.8	113	0.00
72	Anthracene	0.993	1.105	-11.3	115	0.00
73	Carbazole	0.922	1.002	-8.7	112	0.00
74	Di-n-butylphthalate	1.123	1.219	-8.5	110	0.00
75 C	Fluoranthene	1.049	1.204	-14.8	115	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
77	Benzidine	0.399	0.465	-16.5	123	0.00
78	Pyrene	1.488	1.667	-12.0	114	0.00
79 S	Terphenyl-d14	1.151	1.277	-10.9	107	0.00
80	Butylbenzylphthalate	0.625	0.715	-14.4	108	0.00
81	Benzo(a)anthracene	1.247	1.372	-10.0	106	0.00
82	3,3'-Dichlorobenzidine	0.267	0.318	-19.1	117	0.00
83	Chrysene	1.189	1.290	-8.5	106	0.00
84	Bis(2-ethylhexyl)phthalate	0.812	0.978	-20.4	115	0.00
85 c	Di-n-octyl phthalate	1.224	1.432	-17.0	111	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	1.205	1.324	-9.9	121	0.00
88	Benzo(b)fluoranthene	1.278	1.392	-8.9	105	0.00
89	Benzo(k)fluoranthene	1.202	1.354	-12.6	109	0.00
90 C	Benzo(a)pyrene	1.004	1.080	-7.6	105	0.00
91	Dibenzo(a,h)anthracene	1.000	1.109	-10.9	119	0.00
92	Benzo(g,h,i)perylene	0.991	0.916	7.6	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062122\
Data File : BF128901.D
Acq On : 21 Jun 2022 17:02
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 22 01:51:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 22 01:02:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0