

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062122\
 Data File : BF128901.D
 Acq On : 21 Jun 2022 17:02
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 22 01:51:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	110	0.00
2	1,4-Dioxane	40.000	42.456	-6.1	111	0.00
3	Pyridine	40.000	42.969	-7.4	114	0.00
4	n-Nitrosodimethylamine	40.000	38.537	3.7	101	0.00
5 S	2-Fluorophenol	80.000	82.442	-3.1	112	0.00
6	Aniline	40.000	41.341	-3.4	115	0.00
7 S	Phenol-d6	80.000	83.878	-4.8	115	0.00
8	2-Chlorophenol	40.000	42.789	-7.0	118	0.00
9	Benzaldehyde	40.000	45.048	-12.6	153	0.00
10 C	Phenol	40.000	41.273	-3.2	112	0.00
11	bis(2-Chloroethyl)ether	40.000	42.796	-7.0	117	0.00
12	1,3-Dichlorobenzene	40.000	42.623	-6.6	117	0.00
13 C	1,4-Dichlorobenzene	40.000	42.474	-6.2	118	0.00
14	1,2-Dichlorobenzene	40.000	41.932	-4.8	118	0.00
15	Benzyl Alcohol	40.000	35.967	10.1	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.633	-1.6	115	0.00
17	2-Methylphenol	40.000	56.984	-42.5#	160	0.00
18	Hexachloroethane	40.000	42.651	-6.6	116	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.255	-5.6	116	0.00
20	3+4-Methylphenols	40.000	42.683	-6.7	117	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	43.309	-8.3	114	0.00
23 S	Nitrobenzene-d5	80.000	86.704	-8.4	116	0.00
24	Nitrobenzene	40.000	43.685	-9.2	116	0.00
25	Isophorone	40.000	42.926	-7.3	115	0.00
26 C	2-Nitrophenol	40.000	46.333	-15.8	118	0.00
27	2,4-Dimethylphenol	40.000	44.302	-10.8	114	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.015	-7.5	113	0.00
29 C	2,4-Dichlorophenol	40.000	43.784	-9.5	114	0.00
30	1,2,4-Trichlorobenzene	40.000	43.523	-8.8	114	0.00
31	Naphthalene	40.000	43.270	-8.2	119	0.00
32	Benzoic acid	40.000	24.885	37.8#	63	0.00
33	4-Chloroaniline	40.000	41.944	-4.9	116	0.00
34 C	Hexachlorobutadiene	40.000	42.771	-6.9	118	0.00
35	Caprolactam	40.000	42.415	-6.0	117	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.565	-6.4	117	0.00
37	2-Methylnaphthalene	40.000	43.131	-7.8	119	0.00
38	1-Methylnaphthalene	40.000	42.998	-7.5	121	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	47.035	-17.6	119	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.139	-10.3	109	0.00
42 S	2,4,6-Tribromophenol	80.000	84.727	-5.9	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.820	0.4	102	0.00
44	2,4,5-Trichlorophenol	40.000	41.789	-4.5	107	0.00
45 S	2-Fluorobiphenyl	80.000	90.374	-13.0	117	0.00
46	1,1'-Biphenyl	40.000	45.085	-12.7	117	0.00
47	2-Chloronaphthalene	40.000	45.819	-14.5	119	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062122\
 Data File : BF128901.D
 Acq On : 21 Jun 2022 17:02
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 22 01:51:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.410	-13.5	116	0.00
49	Acenaphthylene	40.000	43.516	-8.8	112	0.00
50	Dimethylphthalate	40.000	44.190	-10.5	113	0.00
51	2,6-Dinitrotoluene	40.000	47.100	-17.8	117	0.00
52 C	Acenaphthene	40.000	39.532	1.2	100	0.00
53	3-Nitroaniline	40.000	44.871	-12.2	112	0.00
54 P	2,4-Dinitrophenol	40.000	40.037	-0.1	98	0.00
55	Dibenzofuran	40.000	43.499	-8.7	112	0.00
56 P	4-Nitrophenol	40.000	37.417	6.5	98	0.00
57	2,4-Dinitrotoluene	40.000	46.786	-17.0	115	0.00
58	Fluorene	40.000	43.563	-8.9	110	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.748	5.6	95	0.00
60	Diethylphthalate	40.000	43.300	-8.2	111	0.00
61	4-Chlorophenyl-phenylether	40.000	44.099	-10.2	112	0.00
62	4-Nitroaniline	40.000	44.234	-10.6	111	0.00
63	Azobenzene	40.000	43.110	-7.8	110	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.503	-11.3	111	0.00
66 c	n-Nitrosodiphenylamine	40.000	44.220	-10.5	112	0.00
67	4-Bromophenyl-phenylether	40.000	44.776	-11.9	114	0.00
68	Hexachlorobenzene	40.000	43.635	-9.1	110	0.00
69	Atrazine	40.000	41.738	-4.3	105	0.00
70 C	Pentachlorophenol	40.000	42.600	-6.5	106	0.00
71	Phenanthrene	40.000	43.517	-8.8	113	0.00
72	Anthracene	40.000	44.495	-11.2	115	0.00
73	Carbazole	40.000	43.468	-8.7	112	0.00
74	Di-n-butylphthalate	40.000	43.413	-8.5	110	0.00
75 C	Fluoranthene	40.000	45.914	-14.8	115	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzidine	40.000	46.658	-16.6	123	0.00
78	Pyrene	40.000	44.805	-12.0	114	0.00
79 S	Terphenyl-d14	80.000	88.740	-10.9	107	0.00
80	Butylbenzylphthalate	40.000	45.738	-14.3	108	0.00
81	Benzo(a)anthracene	40.000	44.017	-10.0	106	0.00
82	3,3'-Dichlorobenzidine	40.000	47.619	-19.0	117	0.00
83	Chrysene	40.000	43.403	-8.5	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	48.138	-20.3	115	0.00
85 c	Di-n-octyl phthalate	40.000	46.800	-17.0	111	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.940	-9.8	121	0.00
88	Benzo(b)fluoranthene	40.000	43.567	-8.9	105	0.00
89	Benzo(k)fluoranthene	40.000	45.037	-12.6	109	0.00
90 C	Benzo(a)pyrene	40.000	43.020	-7.6	105	0.00
91	Dibenzo(a,h)anthracene	40.000	44.373	-10.9	119	0.00
92	Benzo(g,h,i)perylene	40.000	36.957	7.6	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062122\
Data File : BF128901.D
Acq On : 21 Jun 2022 17:02
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 22 01:51:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 22 01:02:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0