

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062322\
 Data File : BF128957.D
 Acq On : 23 Jun 2022 17:40
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 24 04:31:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:50:48 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	40.473	-1.2	99	0.00
3	Pyridine	40.000	45.857	-14.6	114	0.00
4	n-Nitrosodimethylamine	40.000	41.461	-3.7	102	0.00
5 S	2-Fluorophenol	80.000	84.269	-5.3	107	0.00
6	Aniline	40.000	43.545	-8.9	113	0.00
7 S	Phenol-d6	80.000	85.490	-6.9	110	0.00
8	2-Chlorophenol	40.000	44.674	-11.7	116	0.00
9	Benzaldehyde	40.000	47.249	-18.1	151	0.00
10 C	Phenol	40.000	43.776	-9.4	111	0.00
11	bis(2-Chloroethyl)ether	40.000	45.428	-13.6	117	0.00
12	1,3-Dichlorobenzene	40.000	43.841	-9.6	113	0.00
13 C	1,4-Dichlorobenzene	40.000	43.884	-9.7	114	0.00
14	1,2-Dichlorobenzene	40.000	42.768	-6.9	112	0.00
15	Benzyl Alcohol	40.000	41.006	-2.5	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	45.929	-14.8	122	0.00
17	2-Methylphenol	40.000	59.112	-47.8#	155	0.00
18	Hexachloroethane	40.000	45.024	-12.6	115	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.766	-11.9	115	0.00
20	3+4-Methylphenols	40.000	44.277	-10.7	114	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	111	0.00
22	Acetophenone	40.000	41.737	-4.3	112	0.00
23 S	Nitrobenzene-d5	80.000	85.705	-7.1	116	0.00
24	Nitrobenzene	40.000	43.986	-10.0	118	0.00
25	Isophorone	40.000	44.172	-10.4	120	0.00
26 C	2-Nitrophenol	40.000	45.970	-14.9	119	0.00
27	2,4-Dimethylphenol	40.000	44.734	-11.8	117	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.421	-8.6	115	0.00
29 C	2,4-Dichlorophenol	40.000	43.219	-8.0	114	0.00
30	1,2,4-Trichlorobenzene	40.000	42.022	-5.1	112	0.00
31	Naphthalene	40.000	43.390	-8.5	121	0.00
32	Benzoic acid	40.000	26.023	34.9#	67	0.00
33	4-Chloroaniline	40.000	42.293	-5.7	119	0.00
34 C	Hexachlorobutadiene	40.000	40.906	-2.3	115	0.00
35	Caprolactam	40.000	47.140	-17.9	132	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.563	-8.9	122	0.00
37	2-Methylnaphthalene	40.000	43.274	-8.2	121	0.00
38	1-Methylnaphthalene	40.000	43.250	-8.1	124	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	116	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.578	-3.9	117	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.725	0.7	108	0.00
42 S	2,4,6-Tribromophenol	80.000	76.441	4.4	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.763	5.6	107	0.00
44	2,4,5-Trichlorophenol	40.000	38.320	4.2	108	0.00
45 S	2-Fluorobiphenyl	80.000	79.786	0.3	114	0.00
46	1,1'-Biphenyl	40.000	41.065	-2.7	118	0.00
47	2-Chloronaphthalene	40.000	42.434	-6.1	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062322\
 Data File : BF128957.D
 Acq On : 23 Jun 2022 17:40
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 24 04:31:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:50:48 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.816	-14.5	129	0.00
49	Acenaphthylene	40.000	41.475	-3.7	118	0.00
50	Dimethylphthalate	40.000	43.644	-9.1	124	0.00
51	2,6-Dinitrotoluene	40.000	45.611	-14.0	126	0.00
52 C	Acenaphthene	40.000	38.776	3.1	108	0.00
53	3-Nitroaniline	40.000	47.019	-17.5	130	0.00
54 P	2,4-Dinitrophenol	40.000	38.015	5.0	103	0.00
55	Dibenzofuran	40.000	40.791	-2.0	117	0.00
56 P	4-Nitrophenol	40.000	35.737	10.7	103	0.00
57	2,4-Dinitrotoluene	40.000	43.396	-8.5	118	0.00
58	Fluorene	40.000	42.086	-5.2	117	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	35.741	10.6	99	0.00
60	Diethylphthalate	40.000	42.298	-5.7	120	0.00
61	4-Chlorophenyl-phenylether	40.000	41.345	-3.4	116	0.00
62	4-Nitroaniline	40.000	43.349	-8.4	120	0.00
63	Azobenzene	40.000	43.641	-9.1	123	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	115	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.435	-11.1	121	0.00
66 c	n-Nitrosodiphenylamine	40.000	43.298	-8.2	120	0.00
67	4-Bromophenyl-phenylether	40.000	43.063	-7.7	120	0.00
68	Hexachlorobenzene	40.000	42.781	-7.0	118	0.00
69	Atrazine	40.000	42.634	-6.6	117	0.00
70 C	Pentachlorophenol	40.000	38.238	4.4	104	0.00
71	Phenanthrene	40.000	42.215	-5.5	119	0.00
72	Anthracene	40.000	43.013	-7.5	122	0.00
73	Carbazole	40.000	42.373	-5.9	119	0.00
74	Di-n-butylphthalate	40.000	42.669	-6.7	118	0.00
75 C	Fluoranthene	40.000	42.949	-7.4	118	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	49.909	-24.8	135	0.00
78	Pyrene	40.000	44.900	-12.2	117	0.00
79 S	Terphenyl-d14	80.000	84.992	-6.2	105	0.00
80	Butylbenzylphthalate	40.000	46.049	-15.1	112	0.00
81	Benzo(a)anthracene	40.000	44.195	-10.5	110	0.00
82	3,3'-Dichlorobenzidine	40.000	48.720	-21.8	123	0.00
83	Chrysene	40.000	44.086	-10.2	111	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	46.536	-16.3	114	0.00
85 c	Di-n-octyl phthalate	40.000	47.281	-18.2	115	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.347	-8.4	123	0.00
88	Benzo(b)fluoranthene	40.000	43.746	-9.4	108	0.00
89	Benzo(k)fluoranthene	40.000	44.898	-12.2	111	0.00
90 C	Benzo(a)pyrene	40.000	44.363	-10.9	111	0.00
91	Dibenzo(a,h)anthracene	40.000	43.409	-8.5	120	0.00
92	Benzo(g,h,i)perylene	40.000	43.167	-7.9	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062322\
Data File : BF128957.D
Acq On : 23 Jun 2022 17:40
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 24 04:31:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:50:48 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0