

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062521\
 Data File : BF124448.D
 Acq On : 25 Jun 2021 12:50
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 25 13:23:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 25 13:23:23 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	39.920	0.2	98	0.00
3	Pyridine	40.000	39.384	1.5	91	0.00
4	n-Nitrosodimethylamine	40.000	40.877	-2.2	101	0.00
5 S	2-Fluorophenol	80.000	80.210	-0.3	97	0.00
6	Aniline	40.000	39.803	0.5	99	0.00
7 S	Phenol-d6	80.000	81.303	-1.6	98	0.00
8	2-Chlorophenol	40.000	39.281	1.8	97	0.00
9	Benzaldehyde	40.000	41.331	-3.3	96	0.00
10 C	Phenol	40.000	38.896	2.8	100	0.00
11	bis(2-Chloroethyl)ether	40.000	40.515	-1.3	102	0.00
12	1,3-Dichlorobenzene	40.000	38.838	2.9	97	0.00
13 C	1,4-Dichlorobenzene	40.000	38.220	4.5	97	0.00
14	1,2-Dichlorobenzene	40.000	39.487	1.3	99	0.00
15	Benzyl Alcohol	40.000	39.425	1.4	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.778	3.1	97	0.00
17	2-Methylphenol	40.000	40.432	-1.1	102	0.00
18	Hexachloroethane	40.000	39.619	1.0	98	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.655	-1.6	104	0.00
20	3+4-Methylphenols	40.000	40.297	-0.7	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	38.866	2.8	99	0.00
23 S	Nitrobenzene-d5	80.000	77.876	2.7	101	0.00
24	Nitrobenzene	40.000	38.065	4.8	105	0.00
25	Isophorone	40.000	38.317	4.2	103	0.00
26 C	2-Nitrophenol	40.000	38.151	4.6	102	0.00
27	2,4-Dimethylphenol	40.000	38.122	4.7	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.350	4.1	101	0.00
29 C	2,4-Dichlorophenol	40.000	37.890	5.3	101	0.00
30	1,2,4-Trichlorobenzene	40.000	38.387	4.0	100	0.00
31	Naphthalene	40.000	37.252	6.9	99	0.00
32	Benzoic acid	40.000	39.733	0.7	111	0.00
33	4-Chloroaniline	40.000	37.423	6.4	98	0.00
34 C	Hexachlorobutadiene	40.000	36.692	8.3	102	0.00
35	Caprolactam	40.000	38.606	3.5	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.332	1.7	107	0.00
37	2-Methylnaphthalene	40.000	37.920	5.2	100	0.00
38	1-Methylnaphthalene	40.000	38.223	4.4	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.944	2.6	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	46.991	-17.5	211	0.00
42 S	2,4,6-Tribromophenol	80.000	84.609	-5.8	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.401	4.0	102	0.00
44	2,4,5-Trichlorophenol	40.000	38.726	3.2	100	0.00
45 S	2-Fluorobiphenyl	80.000	80.975	-1.2	104	0.00
46	1,1'-Biphenyl	40.000	41.679	-4.2	106	0.00
47	2-Chloronaphthalene	40.000	39.474	1.3	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062521\
 Data File : BF124448.D
 Acq On : 25 Jun 2021 12:50
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 25 13:23:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 25 13:23:23 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.095	-2.7	103	0.00
49	Acenaphthylene	40.000	39.963	0.1	106	0.00
50	Dimethylphthalate	40.000	41.040	-2.6	108	0.00
51	2,6-Dinitrotoluene	40.000	44.021	-10.1	116	0.00
52 C	Acenaphthene	40.000	40.625	-1.6	107	0.00
53	3-Nitroaniline	40.000	45.009	-12.5	118	0.00
54 P	2,4-Dinitrophenol	40.000	39.032	2.4	100	0.00
55	Dibenzofuran	40.000	40.530	-1.3	108	0.00
56 P	4-Nitrophenol	40.000	41.752	-4.4	110	0.00
57	2,4-Dinitrotoluene	40.000	42.029	-5.1	111	0.00
58	Fluorene	40.000	40.666	-1.7	109	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.712	0.7	103	0.00
60	Diethylphthalate	40.000	41.321	-3.3	114	0.00
61	4-Chlorophenyl-phenylether	40.000	40.879	-2.2	109	0.00
62	4-Nitroaniline	40.000	43.615	-9.0	118	0.00
63	Azobenzene	40.000	40.311	-0.8	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.414	-1.0	110	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.701	3.2	110	0.00
67	4-Bromophenyl-phenylether	40.000	37.233	6.9	107	0.00
68	Hexachlorobenzene	40.000	39.744	0.6	117	0.00
69	Atrazine	40.000	44.735	-11.8	116	0.00
70 C	Pentachlorophenol	40.000	37.790	5.5	108	0.00
71	Phenanthrene	40.000	40.082	-0.2	116	0.00
72	Anthracene	40.000	39.643	0.9	113	0.00
73	Carbazole	40.000	39.638	0.9	115	0.00
74	Di-n-butylphthalate	40.000	40.499	-1.2	119	0.00
75 C	Fluoranthene	40.000	40.586	-1.5	120	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	122	0.00
77	Benzydine	40.000	43.948	-9.9	136	0.00
78	Pyrene	40.000	39.088	2.3	127	0.00
79 S	Terphenyl-d14	80.000	79.396	0.8	125	0.00
80	Butylbenzylphthalate	40.000	39.955	0.1	133	0.00
81	Benzo(a)anthracene	40.000	39.760	0.6	128	0.00
82	3,3'-Dichlorobenzidine	40.000	37.348	6.6	124	0.00
83	Chrysene	40.000	38.968	2.6	126	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.753	0.6	129	0.00
85 c	Di-n-octyl phthalate	40.000	37.858	5.4	122	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.503	3.7	100	0.00
88	Benzo(b)fluoranthene	40.000	37.901	5.2	105	0.00
89	Benzo(k)fluoranthene	40.000	40.912	-2.3	116	0.00
90 C	Benzo(a)pyrene	40.000	39.166	2.1	111	0.00
91	Dibenzo(a,h)anthracene	40.000	38.104	4.7	100	0.00
92	Benzo(g,h,i)perylene	40.000	38.622	3.4	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062521\
Data File : BF124448.D
Acq On : 25 Jun 2021 12:50
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 25 13:23:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 25 13:23:23 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0