

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062722\
 Data File : BF129026.D
 Acq On : 27 Jun 2022 12:50
 Operator : CG\JU
 Sample : N3434-03
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B2204-27.5

Quant Time: Jun 28 06:54:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 28 06:51:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.687	152	93076	20.000	ng	0.00
21) Naphthalene-d8	7.969	136	513989	20.000	ng	0.00
39) Acenaphthene-d10	9.722	164	282732	20.000	ng	0.00
64) Phenanthrene-d10	11.210	188	395128	20.000	ng	# 0.00
76) Chrysene-d12	13.851	240	325392	20.000	ng	# 0.00
86) Perylene-d12	15.280	264	290238	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.322	112	761287	126.978	ng	0.01
7) Phenol-d6	6.351	99	988638	141.065	ng	0.00
23) Nitrobenzene-d5	7.257	82	692906	64.900	ng	0.00
42) 2,4,6-Tribromophenol	10.522	330	396019	134.914	ng	0.00
45) 2-Fluorobiphenyl	9.051	172	1612028	82.613	ng	0.00
79) Terphenyl-d14	12.804	244	1424104	81.130	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062722\
Data File : BF129026.D
Acq On : 27 Jun 2022 12:50
Operator : CG\JU
Sample : N3434-03
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B2204-27.5

Quant Time: Jun 28 06:54:56 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 28 06:51:48 2022
Response via : Initial Calibration

